



14-Apr-2016

Roger Bellas
Pennsylvania DEP Bureau of Air Quality
12th Floor RCSOB
400 Market Street
Harrisburg, PA 17105

Tel: (570) 826-2511
Fax:

Re: Keystone Sanitary Landfill- 03/23/16

Work Order: **16031000**

Dear Roger,

ALS Environmental received 6 samples on 29-Mar-2016 for the analyses presented in the following report.

The analytical data provided relates directly to the samples received by ALS Environmental and for only the analyses requested.

QC sample results for this data met laboratory specifications. Any exceptions are noted in the Case Narrative, or noted with qualifiers in the report or QC batch information. Should this laboratory report need to be reproduced, it should be reproduced in full unless written approval has been obtained from ALS Laboratory Group. Samples will be disposed in 30 days unless storage arrangements are made.

The total number of pages in this report is 11.

If you have any questions regarding this report, please feel free to contact me.

Sincerely,

Rob Nieman

Electronically approved by: Rob Nieman

Rob Nieman
Project Manager

ADDRESS 4388 Glendale Milford Rd Cincinnati, Ohio 45242- | PHONE (513) 733-5336 | FAX (513) 733-5347

ALS GROUP USA, CORP. Part of the ALS Group An ALS Limited Company

Environmental 

www.alsglobal.com

RIGHT SOLUTIONS RIGHT PARTNER

Client: Pennsylvania DEP Bureau of Air Quality
Project: Keystone Sanitary Landfill- 03/23/16
Work Order: 16031000

Work Order Sample Summary

<u>Lab Samp ID</u>	<u>Client Sample ID</u>	<u>Matrix</u>	<u>Tag Number</u>	<u>Collection Date</u>	<u>Date Received</u>	<u>Hold</u>
16031000-01	KSL032316-1 / Red	Air		3/23/2016	3/29/2016	<input type="checkbox"/>
16031000-02	KSL032316-2 / Blue	Air		3/23/2016	3/29/2016	<input type="checkbox"/>
16031000-03	KSL032316-3 / Green	Air		3/23/2016	3/29/2016	<input type="checkbox"/>
16031000-04	KSL032316-4 / Orange	Air		3/23/2016	3/29/2016	<input type="checkbox"/>
16031000-05	KSL032316-5 / Yellow	Air		3/23/2016	3/29/2016	<input type="checkbox"/>
16031000-06	KSL032316-Summa	Air		3/23/2016	3/29/2016	<input type="checkbox"/>

Client: Pennsylvania DEP Bureau of Air Quality
Project: Keystone Sanitary Landfill- 03/23/16

Work Order: 16031000

Analytical Results

Lab ID: 16031000-01A
Client Sample ID: KSL032316-1 / Red

Collection Date: 3/23/2016
Matrix: AIR

Analyses

AMMONIA BY NIOSH 6015 MOD.		Method: N6015	Air Volume (L): 95.542	Analyst: ALST
Date Analyzed: 4/8/2016		Reporting Limit		
	µg/sample	µg/sample	ug/m3	ppb
Ammonia	ND	1.2	<13	<18

Lab ID: 16031000-02A
Client Sample ID: KSL032316-2 / Blue

Collection Date: 3/23/2016
Matrix: AIR

Analyses

ALDEHYDES BY HPLC		Method: ETO-11	Air Volume (L): 213.9	Analyst: JMB
Date Analyzed: 4/8/2016		Reporting Limit		
	µg/sample	µg/sample	ug/m3	ppb
Acetaldehyde	0.88	0.20	4.1	2.3
Acrolein	ND	5.0	<23	<10
Formaldehyde	0.38	0.20	1.8	1.5

Lab ID: 16031000-03A
Client Sample ID: KSL032316-3 / Green

Collection Date: 3/23/2016
Matrix: AIR

Analyses

METHANOL BY NIOSH 2000 MOD.		Method: N2000	Air Volume (L): 7.13	Analyst: MHW
Date Analyzed: 4/11/2016		Reporting Limit		
	µg/sample	µg/sample	ug/m3	ppb
Methanol	ND	10	<1,400	<1,100

Lab ID: 16031000-04A
Client Sample ID: KSL032316-4 / Orange

Collection Date: 3/23/2016
Matrix: AIR

Analyses

METHYLAMINE BY OSHA 40		Method: O40	Air Volume (L): 21.39	Analyst: MHW
Date Analyzed: 3/31/2016		Reporting Limit		
	µg/sample	µg/sample	ug/m3	ppb
Methylamine	ND	3.0	<140	<110

Note:

Client: Pennsylvania DEP Bureau of Air Quality
Project: Keystone Sanitary Landfill- 03/23/16

Work Order: 16031000

Analytical Results

Lab ID: 16031000-05A
Client Sample ID: KSL032316-5 / Yellow

Collection Date: 3/23/2016
Matrix: AIR

Analyses

AMINE(S) BY OSHA PV2060 MOD.		Method: O2060	Air Volume (L): 21.39	Analyst: MHW
Date Analyzed: 4/6/2016		Reporting Limit		
	µg/sample	µg/sample	ug/m3	ppb
Triethylamine	ND	10	<470	<110

Note:

Client: Pennsylvania DEP Bureau of Air Quality
Work Order: 16031000
Project: Keystone Sanitary Landfill- 03/23/16

QC BATCH REPORT

Batch ID: **35009** Instrument ID: **GC5** Method: **O2060**

MBLK	Sample ID: MBLK-35009-35009			Units: µg/sample		Analysis Date: 4/6/2016				
Client ID:	Run ID: GC5_160406A			SeqNo: 1255747		Prep Date: 4/4/2016		DF: 1		
Analyte	Result	PQL	SPK Val	SPK Ref Value	%REC	Control Limit	RPD Ref Value	%RPD	RPD Limit	Qual

Triethylamine ND 10

LCS	Sample ID: LCS-35009-35009			Units: µg/sample		Analysis Date: 4/6/2016				
Client ID:	Run ID: GC5_160406A			SeqNo: 1255748		Prep Date: 4/4/2016		DF: 1		
Analyte	Result	PQL	SPK Val	SPK Ref Value	%REC	Control Limit	RPD Ref Value	%RPD	RPD Limit	Qual

Triethylamine 68.77 10 90.75 0 75.8 70-130 0

LCSD	Sample ID: LCSD-35009-35009			Units: µg/sample		Analysis Date: 4/6/2016				
Client ID:	Run ID: GC5_160406A			SeqNo: 1255759		Prep Date: 4/4/2016		DF: 1		
Analyte	Result	PQL	SPK Val	SPK Ref Value	%REC	Control Limit	RPD Ref Value	%RPD	RPD Limit	Qual

Triethylamine 65.75 10 90.75 0 72.5 70-130 68.77 4.49 20

The following samples were analyzed in this batch:

Client: Pennsylvania DEP Bureau of Air Quality
 Work Order: 16031000
 Project: Keystone Sanitary Landfill- 03/23/16

QC BATCH REPORT

Batch ID: **R127744** Instrument ID: **GC1** Method: **N2000**

MBLK		Sample ID: MB-R127744-R127744			Units: µg/sample		Analysis Date: 4/11/2016			
Client ID:		Run ID: GC1_160411B			SeqNo: 1258506		Prep Date:		DF: 1	
Analyte	Result	PQL	SPK Val	SPK Ref Value	%REC	Control Limit	RPD Ref Value	%RPD	RPD Limit	Qual

Methanol ND 10

LCS		Sample ID: LCS-R127744-R127744			Units: µg/sample		Analysis Date: 4/11/2016			
Client ID:		Run ID: GC1_160411B			SeqNo: 1258507		Prep Date:		DF: 1	
Analyte	Result	PQL	SPK Val	SPK Ref Value	%REC	Control Limit	RPD Ref Value	%RPD	RPD Limit	Qual

Methanol 93.14 10 79.1 0 118 64.1-145 0

LCSD		Sample ID: LCSD-R127744-R127744			Units: µg/sample		Analysis Date: 4/11/2016			
Client ID:		Run ID: GC1_160411B			SeqNo: 1258531		Prep Date:		DF: 1	
Analyte	Result	PQL	SPK Val	SPK Ref Value	%REC	Control Limit	RPD Ref Value	%RPD	RPD Limit	Qual

Methanol 78.12 10 79.1 0 98.8 64.1-145 93.14 17.5 20

The following samples were analyzed in this batch:

16031000-03A

Client: Pennsylvania DEP Bureau of Air Quality
 Work Order: 16031000
 Project: Keystone Sanitary Landfill- 03/23/16

QC BATCH REPORT

Batch ID: 35024 Instrument ID: HPLC2 Method: ETO-11

MBLK		Sample ID: MBLK-35024-35024			Units: µg/sample		Analysis Date: 4/8/2016			
Client ID:		Run ID: HPLC2_160408B			SeqNo: 1259257		Prep Date: 4/6/2016		DF: 1	
Analyte	Result	PQL	SPK Val	SPK Ref Value	%REC	Control Limit	RPD Ref Value	%RPD	RPD Limit	Qual
Acetaldehyde	ND	0.20								
Formaldehyde	ND	0.20								

LCS		Sample ID: LCS-35024-35024			Units: µg/sample		Analysis Date: 4/8/2016			
Client ID:		Run ID: HPLC2_160408B			SeqNo: 1259258		Prep Date: 4/6/2016		DF: 1	
Analyte	Result	PQL	SPK Val	SPK Ref Value	%REC	Control Limit	RPD Ref Value	%RPD	RPD Limit	Qual
Acetaldehyde	1.714	0.20	2	0	85.7	70-130	0			
Formaldehyde	2.218	0.20	2	0	111	70-130	0			

LCSD		Sample ID: LCSD-35024-35024			Units: µg/sample		Analysis Date: 4/8/2016			
Client ID:		Run ID: HPLC2_160408B			SeqNo: 1259273		Prep Date: 4/6/2016		DF: 1	
Analyte	Result	PQL	SPK Val	SPK Ref Value	%REC	Control Limit	RPD Ref Value	%RPD	RPD Limit	Qual
Acetaldehyde	1.666	0.20	2	0	83.3	70-130	1.714	2.8	20	
Formaldehyde	2.19	0.20	2	0	110	70-130	2.218	1.25	20	

The following samples were analyzed in this batch:

16031000-02A

Client: Pennsylvania DEP Bureau of Air Quality
 Work Order: 16031000
 Project: Keystone Sanitary Landfill- 03/23/16

QC BATCH REPORT

Batch ID: R127513 Instrument ID: HPLC1 Method: O40

MBLK		Sample ID: MB-R127513-R127513			Units: µg/sample			Analysis Date: 3/31/2016		
Client ID:		Run ID: HPLC1_160331B			SeqNo: 1252868			Prep Date:		DF: 1
Analyte	Result	PQL	SPK Val	SPK Ref Value	%REC	Control Limit	RPD Ref Value	%RPD	RPD Limit	Qual

Methylamine ND 3.0

LCS		Sample ID: LCS-R127513-R127513			Units: µg/sample			Analysis Date: 3/31/2016		
Client ID:		Run ID: HPLC1_160331B			SeqNo: 1252869			Prep Date:		DF: 1
Analyte	Result	PQL	SPK Val	SPK Ref Value	%REC	Control Limit	RPD Ref Value	%RPD	RPD Limit	Qual

Methylamine 5.226 3.0 7.44 0 70.2 70-130 0

LCSD		Sample ID: LCSD-R127513-R127513			Units: µg/sample			Analysis Date: 3/31/2016		
Client ID:		Run ID: HPLC1_160331B			SeqNo: 1252890			Prep Date:		DF: 1
Analyte	Result	PQL	SPK Val	SPK Ref Value	%REC	Control Limit	RPD Ref Value	%RPD	RPD Limit	Qual

Methylamine 4.196 3.0 7.44 0 56.4 70-130 5.226 21.9 20 SR

The following samples were analyzed in this batch:

16031000-04A

Client: Pennsylvania DEP Bureau of Air Quality
 Work Order: 16031000
 Project: Keystone Sanitary Landfill- 03/23/16

QC BATCH REPORT

Batch ID: **R127786** Instrument ID: **SUB** Method: **N6015**

MBLK		Sample ID: MB-R127786-R127786			Units: µg/sample			Analysis Date: 4/8/2016		
Client ID:		Run ID: SUB_160408C			SeqNo: 1259155			Prep Date:		DF: 1
Analyte	Result	PQL	SPK Val	SPK Ref Value	%REC	Control Limit	RPD Ref Value	%RPD	RPD Limit	Qual
Ammonia	ND	1.2								

LCS		Sample ID: LCS-R127786-R127786			Units: µg/sample			Analysis Date: 4/8/2016		
Client ID:		Run ID: SUB_160408C			SeqNo: 1259156			Prep Date:		DF: 1
Analyte	Result	PQL	SPK Val	SPK Ref Value	%REC	Control Limit	RPD Ref Value	%RPD	RPD Limit	Qual
Ammonia	20.6	1.2	24.3	0	84.8	74.3-115.2	0			

LCSD		Sample ID: LCSD-R127786			Units: µg/sample			Analysis Date: 4/8/2016		
Client ID:		Run ID: SUB_160408C			SeqNo: 1259164			Prep Date:		DF: 1
Analyte	Result	PQL	SPK Val	SPK Ref Value	%REC	Control Limit	RPD Ref Value	%RPD	RPD Limit	Qual
Ammonia	21.7	1.2	24.3	0	89.3	74.3-115.2	20.6	5.2	20	

The following samples were analyzed in this batch:

Note: See Qualifiers Page for a list of Qualifiers and their explanation.

Client: Pennsylvania DEP Bureau of Air Quality
Project: Keystone Sanitary Landfill- 03/23/16
WorkOrder: 16031000

**QUALIFIERS,
ACRONYMS, UNITS**

<u>Qualifier</u>	<u>Description</u>
*	Value exceeds Regulatory Limit
a	Not accredited
B	Analyte detected in the associated Method Blank above the Reporting Limit
E	Value above quantitation range
H	Analyzed outside of Holding Time
J	Analyte detected below quantitation limit
n	Not offered for accreditation
ND	Not Detected at the Reporting Limit
O	Sample amount is > 4 times amount spiked
P	Dual Column results percent difference > 40%
R	RPD above laboratory control limit
S	Spike Recovery outside laboratory control limits
U	Analyzed but not detected above the MDL

<u>Acronym</u>	<u>Description</u>
DUP	Method Duplicate
E	EPA Method
LCS	Laboratory Control Sample
LCSD	Laboratory Control Sample Duplicate
MBLK	Method Blank
MDL	Method Detection Limit
MQL	Method Quantitation Limit
MS	Matrix Spike
MSD	Matrix Spike Duplicate
PDS	Post Digestion Spike
PQL	Practical Quantitation Limit
SDL	Sample Detection Limit
SW	SW-846 Method

<u>Units Reported</u>	<u>Description</u>
µg/sample	
ppbv	
ppm	

Sample Receipt Checklist

Client Name: PADEP-HARRISBURG

Date/Time Received: 29-Mar-16 00:00

Work Order: 16031000

Received by: MEB

Checklist completed by: Mike Bishop 29-Mar-16
eSignature Date

Reviewed by: Rob Nieman 30-Mar-16
eSignature Date

Matrices:

Carrier name: FedEx

- Shipping container/cooler in good condition? Yes [checked] No [] Not Present []
Custody seals intact on shipping container/cooler? Yes [checked] No [] Not Present []
Custody seals intact on sample bottles? Yes [] No [] Not Present [checked]
Chain of custody present? Yes [checked] No []
Chain of custody signed when relinquished and received? Yes [checked] No []
Chain of custody agrees with sample labels? Yes [checked] No []
Samples in proper container/bottle? Yes [checked] No []
Sample containers intact? Yes [checked] No []
Sufficient sample volume for indicated test? Yes [checked] No []
All samples received within holding time? Yes [checked] No []
Container/Temp Blank temperature in compliance? Yes [checked] No []

Temperature(s)/Thermometer(s): 3.4

Cooler(s)/Kit(s):


Water - VOA vials have zero headspace? Yes [] No [] No VOA vials submitted []

Water - pH acceptable upon receipt? Yes [] No [] N/A []

pH adjusted? Yes [] No [] N/A []

pH adjusted by:

Login Notes:



Client Contacted: Date Contacted: Person Contacted:

Contacted By: Regarding:

Comments:

CorrectiveAction: