



eurofins

Environment Testing America



ANALYTICAL REPORT

Eurofins Lancaster Laboratories Env, LLC
2425 New Holland Pike
Lancaster, PA 17601
Tel: (717)656-2300

Laboratory Job ID: 410-51333-1
Client Project/Site: Newberry System
Sampling Event: Newberry System

For:
SUEZ Water Environmental Services Inc
6310 Allentown Blvd
Suite 104
Harrisburg, Pennsylvania 17112

Attn: Penny Bumbarger

Elizabeth M. Zanar

Authorized for release by:
8/25/2021 8:42:43 AM

Elizabeth Zanar, Project Manager
(717)556-7290
Elizabeth.Zanar@eurofinset.com

LINKS

Review your project
results through

Total Access

Have a Question?

Ask
The
Expert

Visit us at:

www.eurofinsus.com/Env

The test results in this report meet all 2003 NELAC, 2009 TNI, and 2016 TNI requirements for accredited parameters, exceptions are noted in this report. This report may not be reproduced except in full, and with written approval from the laboratory. For questions please contact the Project Manager at the e-mail address or telephone number listed on this page.

This report has been electronically signed and authorized by the signatory. Electronic signature is intended to be the legally binding equivalent of a traditionally handwritten signature.

Results relate only to the items tested and the sample(s) as received by the laboratory.

1

2

3

4

5

6

7

8

9

10

11

12

13

14

15

Analytical test results meet all requirements of the associated regulatory program (e.g., NELAC (TNI), DoD, and ISO 17025) unless otherwise noted under the individual analysis. Data qualifiers are applied to note exceptions. Noncompliant quality control (QC) is further explained in narrative comments.

- QC results that exceed the upper limits and are associated with non-detect samples are qualified but further narration is not required since the bias is high and does not change a non-detect result. Further narration is also not required with QC blank detection when the associated sample concentration is non-detect or more than ten times the level in the blank.
- Matrix QC may not be reported if insufficient sample or site-specific QC samples were not submitted. In these situations, to demonstrate precision and accuracy at a batch level, a LCS/LCSD is performed, unless otherwise specified in the method.
- Surrogate and/or isotope dilution analyte recoveries (if applicable) which are outside of the QC window are confirmed unless attributed to a dilution or otherwise noted in the narrative.

Regulated compliance samples (e.g. SDWA, NPDES) must comply with the associated agency requirements/permits.

Measurement uncertainty values, as applicable, are available upon request.

Test results relate only to the sample tested. Clients should be aware that a critical step in a chemical or microbiological analysis is the collection of the sample. Unless the sample analyzed is truly representative of the bulk of material involved, the test results will be meaningless. If you have questions regarding the proper techniques of collecting samples, please contact us. We cannot be held responsible for sample integrity, however, unless sampling has been performed by a member of our staff. Times are local to the area of activity. Parameters listed in the 40 CFR Part 136 Table II as "analyze immediately" and tested in the laboratory are not performed within 15 minutes of collection.

This report shall not be reproduced except in full, without the written approval of the laboratory.

WARRANTY AND LIMITS OF LIABILITY - In accepting analytical work, we warrant the accuracy of test results for the sample as submitted. The foregoing express warranty is exclusive and is given in lieu of all other warranties, expressed or implied, except as otherwise agreed. We disclaim any other warranties, expressed or implied, including a warranty of fitness for particular purpose and warranty of merchantability. In no event shall Eurofins Lancaster Laboratories Environmental, LLC be liable for indirect, special, consequential, or incidental damages including, but not limited to, damages for loss of profit or goodwill regardless of (A) the negligence (either sole or concurrent) of Eurofins Lancaster Laboratories Environmental and (B) whether Eurofins Lancaster Laboratories Environmental has been informed of the possibility of such damages. We accept no legal responsibility for the purposes for which the client uses the test results. Except as otherwise agreed, no purchase order or other order for work shall be accepted by Eurofins Lancaster Laboratories Environmental which includes any conditions that vary from the Standard Terms and Conditions, and Eurofins Lancaster Laboratories Environmental hereby objects to any conflicting terms contained in any acceptance or order submitted by client.



Elizabeth Zanar
Project Manager
8/25/2021 8:42:43 AM

Table of Contents

Cover Page	1
Table of Contents	3
Definitions/Glossary	4
Case Narrative	5
Detection Summary	6
Client Sample Results	9
Surrogate Summary	20
QC Sample Results	22
QC Association Summary	26
Lab Chronicle	28
Certification Summary	32
Method Summary	33
Sample Summary	34
Chain of Custody	35
Receipt Checklists	37

Definitions/Glossary

Client: SUEZ Water Environmental Services Inc
Project/Site: Newberry System

Job ID: 410-51333-1

Qualifiers

LCMS

Qualifier	Qualifier Description
J	Result is less than the RL but greater than or equal to the MDL and the concentration is an approximate value.

Glossary

Abbreviation	These commonly used abbreviations may or may not be present in this report.
□	Listed under the "D" column to designate that the result is reported on a dry weight basis
%R	Percent Recovery
1C	Result is from the primary column on a dual-column method.
2C	Result is from the confirmation column on a dual-column method.
CFL	Contains Free Liquid
CFU	Colony Forming Unit
CNF	Contains No Free Liquid
DER	Duplicate Error Ratio (normalized absolute difference)
Dil Fac	Dilution Factor
DL	Detection Limit (DoD/DOE)
DL, RA, RE, IN	Indicates a Dilution, Re-analysis, Re-extraction, or additional Initial metals/anion analysis of the sample
DLC	Decision Level Concentration (Radiochemistry)
EDL	Estimated Detection Limit (Dioxin)
LOD	Limit of Detection (DoD/DOE)
LOQ	Limit of Quantitation (DoD/DOE)
MCL	EPA recommended "Maximum Contaminant Level"
MDA	Minimum Detectable Activity (Radiochemistry)
MDC	Minimum Detectable Concentration (Radiochemistry)
MDL	Method Detection Limit
ML	Minimum Level (Dioxin)
MPN	Most Probable Number
MQL	Method Quantitation Limit
NC	Not Calculated
ND	Not Detected at the reporting limit (or MDL or EDL if shown)
NEG	Negative / Absent
POS	Positive / Present
PQL	Practical Quantitation Limit
PRES	Presumptive
QC	Quality Control
RER	Relative Error Ratio (Radiochemistry)
RL	Reporting Limit or Requested Limit (Radiochemistry)
RPD	Relative Percent Difference, a measure of the relative difference between two points
TEF	Toxicity Equivalent Factor (Dioxin)
TEQ	Toxicity Equivalent Quotient (Dioxin)
TNTC	Too Numerous To Count

Case Narrative

Client: SUEZ Water Environmental Services Inc
Project/Site: Newberry System

Job ID: 410-51333-1

Job ID: 410-51333-1

Laboratory: Eurofins Lancaster Laboratories Env, LLC

Narrative

Job Narrative
410-51333-1

Receipt

The samples were received on 8/13/2021 5:28 PM. Unless otherwise noted below, the samples arrived in good condition, and, where required, properly preserved and on ice. The temperatures of the 2 coolers at receipt time were 0.1°C and 1.8°C

PFAS

No additional analytical or quality issues were noted, other than those described above or in the Definitions/ Glossary page.

Detection Summary

Client: SUEZ Water Environmental Services Inc
 Project/Site: Newberry System

Job ID: 410-51333-1

Client Sample ID: 7670061 001 Playground Well

Lab Sample ID: 410-51333-1

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Perfluorohexanoic acid	14		1.7	0.42	ng/L	1		EPA 537 Ver 1.1	Total/NA
Perfluoroheptanoic acid	4.2		1.7	0.42	ng/L	1		EPA 537 Ver 1.1	Total/NA
Perfluoroctanoic acid	9.3		1.7	0.42	ng/L	1		EPA 537 Ver 1.1	Total/NA
Perfluorobutanesulfonic acid	4.8		1.7	0.42	ng/L	1		EPA 537 Ver 1.1	Total/NA
Perfluorohexanesulfonic acid	6.1		1.7	0.42	ng/L	1		EPA 537 Ver 1.1	Total/NA
Perfluoroctanesulfonic acid	14		1.7	0.42	ng/L	1		EPA 537 Ver 1.1	Total/NA

Client Sample ID: 7670061 001 Playground Well FB

Lab Sample ID: 410-51333-2

No Detections.

Client Sample ID: 7670061 005 Conley Well

Lab Sample ID: 410-51333-3

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Perfluorohexanoic acid	11		1.7	0.43	ng/L	1		EPA 537 Ver 1.1	Total/NA
Perfluoroheptanoic acid	2.8		1.7	0.43	ng/L	1		EPA 537 Ver 1.1	Total/NA
Perfluoroctanoic acid	5.4		1.7	0.43	ng/L	1		EPA 537 Ver 1.1	Total/NA
Perfluorobutanesulfonic acid	3.3		1.7	0.43	ng/L	1		EPA 537 Ver 1.1	Total/NA
Perfluorohexanesulfonic acid	4.4		1.7	0.43	ng/L	1		EPA 537 Ver 1.1	Total/NA
Perfluoroctanesulfonic acid	8.0		1.7	0.43	ng/L	1		EPA 537 Ver 1.1	Total/NA

Client Sample ID: 7670061 005 Conley Well FB

Lab Sample ID: 410-51333-4

No Detections.

Client Sample ID: 7670061 301 Conley Between Lead & Lag

Lab Sample ID: 410-51333-5

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Perfluorohexanoic acid	13		1.7	0.44	ng/L	1		EPA 537 Ver 1.1	Total/NA
Perfluoroheptanoic acid	4.1		1.7	0.44	ng/L	1		EPA 537 Ver 1.1	Total/NA
Perfluoroctanoic acid	8.7		1.7	0.44	ng/L	1		EPA 537 Ver 1.1	Total/NA
Perfluorobutanesulfonic acid	4.8		1.7	0.44	ng/L	1		EPA 537 Ver 1.1	Total/NA
Perfluorohexanesulfonic acid	6.8		1.7	0.44	ng/L	1		EPA 537 Ver 1.1	Total/NA
Perfluoroctanesulfonic acid	11		1.7	0.44	ng/L	1		EPA 537 Ver 1.1	Total/NA

Client Sample ID: 7670061 301 Conley Between Lead & Lag

Lab Sample ID: 410-51333-6

FB

No Detections.

Client Sample ID: 7670061 301 Conley After Lag Vessel

Lab Sample ID: 410-51333-7

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Perfluorohexanoic acid	13		1.7	0.43	ng/L	1		EPA 537 Ver 1.1	Total/NA
Perfluoroheptanoic acid	3.9		1.7	0.43	ng/L	1		EPA 537 Ver 1.1	Total/NA
Perfluoroctanoic acid	7.2		1.7	0.43	ng/L	1		EPA 537 Ver 1.1	Total/NA
Perfluorobutanesulfonic acid	5.1		1.7	0.43	ng/L	1		EPA 537 Ver 1.1	Total/NA
Perfluorohexanesulfonic acid	5.9		1.7	0.43	ng/L	1		EPA 537 Ver 1.1	Total/NA
Perfluoroctanesulfonic acid	6.9		1.7	0.43	ng/L	1		EPA 537 Ver 1.1	Total/NA

Client Sample ID: 7670061 301 Conley After Lag Vessel FB

Lab Sample ID: 410-51333-8

No Detections.

This Detection Summary does not include radiochemical test results.

Eurofins Lancaster Laboratories Env, LLC

Detection Summary

Client: SUEZ Water Environmental Services Inc

Job ID: 410-51333-1

Project/Site: Newberry System

Client Sample ID: 7670061 101 Conley EP Grab Water

Lab Sample ID: 410-51333-9

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Perfluorohexanoic acid	13		1.7	0.43	ng/L	1	EPA 537 Ver 1.1	Total/NA	
Perfluoroheptanoic acid	3.9		1.7	0.43	ng/L	1	EPA 537 Ver 1.1	Total/NA	
Perfluoroctanoic acid	7.4		1.7	0.43	ng/L	1	EPA 537 Ver 1.1	Total/NA	
Perfluorobutanesulfonic acid	5.1		1.7	0.43	ng/L	1	EPA 537 Ver 1.1	Total/NA	
Perfluorohexanesulfonic acid	5.6		1.7	0.43	ng/L	1	EPA 537 Ver 1.1	Total/NA	
Perfluoroctanesulfonic acid	5.8		1.7	0.43	ng/L	1	EPA 537 Ver 1.1	Total/NA	

Client Sample ID: 7670061 101 Conley Field Blank Grab Water

Lab Sample ID: 410-51333-10

No Detections.

Client Sample ID: 7670061 002 Coppersmith Well

Lab Sample ID: 410-51333-11

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Perfluorohexanoic acid	17		1.7	0.44	ng/L	1	EPA 537 Ver 1.1	Total/NA	
Perfluoroheptanoic acid	8.3		1.7	0.44	ng/L	1	EPA 537 Ver 1.1	Total/NA	
Perfluoroctanoic acid	10		1.7	0.44	ng/L	1	EPA 537 Ver 1.1	Total/NA	
Perfluorononanoic acid	6.3		1.7	0.44	ng/L	1	EPA 537 Ver 1.1	Total/NA	
Perfluorobutanesulfonic acid	8.0		1.7	0.44	ng/L	1	EPA 537 Ver 1.1	Total/NA	
Perfluorohexanesulfonic acid - DL	62		17	4.4	ng/L	10	EPA 537 Ver 1.1	Total/NA	
Perfluoroctanesulfonic acid - DL	68		17	4.4	ng/L	10	EPA 537 Ver 1.1	Total/NA	

Client Sample ID: 7670061 002 Coppersmith Well FB

Lab Sample ID: 410-51333-12

No Detections.

Client Sample ID: 7670061 003 DuPont Well

Lab Sample ID: 410-51333-13

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Perfluorohexanoic acid	7.5		1.8	0.45	ng/L	1	EPA 537 Ver 1.1	Total/NA	
Perfluoroheptanoic acid	3.5		1.8	0.45	ng/L	1	EPA 537 Ver 1.1	Total/NA	
Perfluoroctanoic acid	6.4		1.8	0.45	ng/L	1	EPA 537 Ver 1.1	Total/NA	
Perfluorobutanesulfonic acid	5.4		1.8	0.45	ng/L	1	EPA 537 Ver 1.1	Total/NA	
Perfluorohexanesulfonic acid	66		1.8	0.45	ng/L	1	EPA 537 Ver 1.1	Total/NA	
Perfluoroctanesulfonic acid	58		1.8	0.45	ng/L	1	EPA 537 Ver 1.1	Total/NA	

Client Sample ID: 7670061 003 DuPont Well FB

Lab Sample ID: 410-51333-14

No Detections.

Client Sample ID: 7670061 302 DuPont Between Lead & Lag

Lab Sample ID: 410-51333-15

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Perfluorohexanoic acid	15		1.8	0.46	ng/L	1	EPA 537 Ver 1.1	Total/NA	
Perfluoroheptanoic acid	2.4		1.8	0.46	ng/L	1	EPA 537 Ver 1.1	Total/NA	
Perfluorobutanesulfonic acid	5.0		1.8	0.46	ng/L	1	EPA 537 Ver 1.1	Total/NA	
Perfluorohexanesulfonic acid	6.7		1.8	0.46	ng/L	1	EPA 537 Ver 1.1	Total/NA	
Perfluoroctanesulfonic acid	2.0		1.8	0.46	ng/L	1	EPA 537 Ver 1.1	Total/NA	

Client Sample ID: 7670061 302 DuPont Between Lead & Lag

Lab Sample ID: 410-51333-16

FB

No Detections.

This Detection Summary does not include radiochemical test results.

Eurofins Lancaster Laboratories Env, LLC

Detection Summary

Client: SUEZ Water Environmental Services Inc
Project/Site: Newberry System

Job ID: 410-51333-1

Client Sample ID: 7670061 302 DuPont After Lag Vessel

Lab Sample ID: 410-51333-17

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Perfluorohexanoic acid	2.8		1.8	0.45	ng/L	1		EPA 537 Ver 1.1	Total/NA

Client Sample ID: 7670061 302 DuPont After Lag Vessel FB

Lab Sample ID: 410-51333-18

No Detections.

Client Sample ID: 7670061 102 DuPont EP

Lab Sample ID: 410-51333-19

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Perfluorohexanoic acid	2.5		1.8	0.45	ng/L	1		EPA 537 Ver 1.1	Total/NA

Client Sample ID: 7670061 102 DuPont EP FB

Lab Sample ID: 410-51333-20

No Detections.

This Detection Summary does not include radiochemical test results.

Eurofins Lancaster Laboratories Env, LLC

Client Sample Results

Client: SUEZ Water Environmental Services Inc
 Project/Site: Newberry System

Job ID: 410-51333-1

Client Sample ID: 7670061 001 Playground Well

Lab Sample ID: 410-51333-1

Matrix: Drinking Water

Date Collected: 08/12/21 09:55
 Date Received: 08/13/21 17:28

Method: EPA 537 Ver 1.1 - EPA 537 Version 1.1

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorohexanoic acid	14		1.7	0.42	ng/L		08/17/21 07:04	08/21/21 18:38	1
Perfluoroheptanoic acid	4.2		1.7	0.42	ng/L		08/17/21 07:04	08/21/21 18:38	1
Perfluoroctanoic acid	9.3		1.7	0.42	ng/L		08/17/21 07:04	08/21/21 18:38	1
Perflurononanoic acid	<1.7		1.7	0.42	ng/L		08/17/21 07:04	08/21/21 18:38	1
Perfluorodecanoic acid	<1.7		1.7	0.42	ng/L		08/17/21 07:04	08/21/21 18:38	1
Perfluorotridecanoic acid	<1.7		1.7	0.42	ng/L		08/17/21 07:04	08/21/21 18:38	1
Perfluorotetradecanoic acid	<1.7		1.7	0.42	ng/L		08/17/21 07:04	08/21/21 18:38	1
Perfluorobutanesulfonic acid	4.8		1.7	0.42	ng/L		08/17/21 07:04	08/21/21 18:38	1
Perfluorohexanesulfonic acid	6.1		1.7	0.42	ng/L		08/17/21 07:04	08/21/21 18:38	1
Perfluooctanesulfonic acid	14		1.7	0.42	ng/L		08/17/21 07:04	08/21/21 18:38	1
NEtFOSAA	<1.7		1.7	0.42	ng/L		08/17/21 07:04	08/21/21 18:38	1
NMeFOSAA	<1.7		1.7	0.42	ng/L		08/17/21 07:04	08/21/21 18:38	1
Perfluoroundecanoic acid	<1.7		1.7	0.42	ng/L		08/17/21 07:04	08/21/21 18:38	1
Perfluorododecanoic acid	<1.7		1.7	0.42	ng/L		08/17/21 07:04	08/21/21 18:38	1
Surrogate		%Recovery	Qualifier	Limits			Prepared	Analyzed	Dil Fac
d5-NEtFOSAA		109		70 - 130			08/17/21 07:04	08/21/21 18:38	1
13C2 PFDA		112		70 - 130			08/17/21 07:04	08/21/21 18:38	1
13C2 PFHxA		102		70 - 130			08/17/21 07:04	08/21/21 18:38	1

Client Sample ID: 7670061 001 Playground Well FB

Lab Sample ID: 410-51333-2

Matrix: Potable Water

Date Collected: 08/12/21 09:55
 Date Received: 08/13/21 17:28

Method: EPA 537 Ver 1.1 - EPA 537 Version 1.1

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorohexanoic acid	<1.7		1.7	0.43	ng/L		08/17/21 07:04	08/21/21 18:50	1
Perfluoroheptanoic acid	<1.7		1.7	0.43	ng/L		08/17/21 07:04	08/21/21 18:50	1
Perfluoroctanoic acid	<1.7		1.7	0.43	ng/L		08/17/21 07:04	08/21/21 18:50	1
Perflurononanoic acid	<1.7		1.7	0.43	ng/L		08/17/21 07:04	08/21/21 18:50	1
Perfluorodecanoic acid	<1.7		1.7	0.43	ng/L		08/17/21 07:04	08/21/21 18:50	1
Perfluorotridecanoic acid	<1.7		1.7	0.43	ng/L		08/17/21 07:04	08/21/21 18:50	1
Perfluorotetradecanoic acid	<1.7		1.7	0.43	ng/L		08/17/21 07:04	08/21/21 18:50	1
Perfluorobutanesulfonic acid	<1.7		1.7	0.43	ng/L		08/17/21 07:04	08/21/21 18:50	1
Perfluorohexanesulfonic acid	<1.7		1.7	0.43	ng/L		08/17/21 07:04	08/21/21 18:50	1
Perfluooctanesulfonic acid	<1.7		1.7	0.43	ng/L		08/17/21 07:04	08/21/21 18:50	1
NEtFOSAA	<1.7		1.7	0.43	ng/L		08/17/21 07:04	08/21/21 18:50	1
NMeFOSAA	<1.7		1.7	0.43	ng/L		08/17/21 07:04	08/21/21 18:50	1
Perfluoroundecanoic acid	<1.7		1.7	0.43	ng/L		08/17/21 07:04	08/21/21 18:50	1
Perfluorododecanoic acid	<1.7		1.7	0.43	ng/L		08/17/21 07:04	08/21/21 18:50	1
Surrogate		%Recovery	Qualifier	Limits			Prepared	Analyzed	Dil Fac
d5-NEtFOSAA		109		70 - 130			08/17/21 07:04	08/21/21 18:50	1
13C2 PFDA		112		70 - 130			08/17/21 07:04	08/21/21 18:50	1
13C2 PFHxA		110		70 - 130			08/17/21 07:04	08/21/21 18:50	1

Client Sample Results

Client: SUEZ Water Environmental Services Inc
Project/Site: Newberry System

Job ID: 410-51333-1

Client Sample ID: 7670061 005 Conley Well

Lab Sample ID: 410-51333-3

Date Collected: 08/12/21 10:00

Matrix: Drinking Water

Date Received: 08/13/21 17:28

Method: EPA 537 Ver 1.1 - EPA 537 Version 1.1

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorohexanoic acid	11		1.7	0.43	ng/L		08/17/21 07:04	08/21/21 19:02	1
Perfluoroheptanoic acid	2.8		1.7	0.43	ng/L		08/17/21 07:04	08/21/21 19:02	1
Perfluoroctanoic acid	5.4		1.7	0.43	ng/L		08/17/21 07:04	08/21/21 19:02	1
Perflurooronanoic acid	<1.7		1.7	0.43	ng/L		08/17/21 07:04	08/21/21 19:02	1
Perfluorodecanoic acid	<1.7		1.7	0.43	ng/L		08/17/21 07:04	08/21/21 19:02	1
Perfluorotridecanoic acid	<1.7		1.7	0.43	ng/L		08/17/21 07:04	08/21/21 19:02	1
Perfluorotetradecanoic acid	<1.7		1.7	0.43	ng/L		08/17/21 07:04	08/21/21 19:02	1
Perfluorobutanesulfonic acid	3.3		1.7	0.43	ng/L		08/17/21 07:04	08/21/21 19:02	1
Perfluorohexanesulfonic acid	4.4		1.7	0.43	ng/L		08/17/21 07:04	08/21/21 19:02	1
Perfluoroctanesulfonic acid	8.0		1.7	0.43	ng/L		08/17/21 07:04	08/21/21 19:02	1
NEtFOSAA	<1.7		1.7	0.43	ng/L		08/17/21 07:04	08/21/21 19:02	1
NMeFOSAA	<1.7		1.7	0.43	ng/L		08/17/21 07:04	08/21/21 19:02	1
Perfluoroundecanoic acid	<1.7		1.7	0.43	ng/L		08/17/21 07:04	08/21/21 19:02	1
Perfluorododecanoic acid	<1.7		1.7	0.43	ng/L		08/17/21 07:04	08/21/21 19:02	1
Surrogate		%Recovery	Qualifier	Limits			Prepared	Analyzed	Dil Fac
d5-NEtFOSAA		108		70 - 130			08/17/21 07:04	08/21/21 19:02	1
13C2 PFDA		113		70 - 130			08/17/21 07:04	08/21/21 19:02	1
13C2 PFHxA		107		70 - 130			08/17/21 07:04	08/21/21 19:02	1

Client Sample ID: 7670061 005 Conley Well FB

Lab Sample ID: 410-51333-4

Date Collected: 08/12/21 10:00

Matrix: Potable Water

Date Received: 08/13/21 17:28

Method: EPA 537 Ver 1.1 - EPA 537 Version 1.1

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorohexanoic acid	<1.7		1.7	0.43	ng/L		08/17/21 07:04	08/21/21 19:13	1
Perfluoroheptanoic acid	<1.7		1.7	0.43	ng/L		08/17/21 07:04	08/21/21 19:13	1
Perfluoroctanoic acid	<1.7		1.7	0.43	ng/L		08/17/21 07:04	08/21/21 19:13	1
Perflurooronanoic acid	<1.7		1.7	0.43	ng/L		08/17/21 07:04	08/21/21 19:13	1
Perfluorodecanoic acid	<1.7		1.7	0.43	ng/L		08/17/21 07:04	08/21/21 19:13	1
Perfluorotridecanoic acid	<1.7		1.7	0.43	ng/L		08/17/21 07:04	08/21/21 19:13	1
Perfluorotetradecanoic acid	<1.7		1.7	0.43	ng/L		08/17/21 07:04	08/21/21 19:13	1
Perfluorobutanesulfonic acid	<1.7		1.7	0.43	ng/L		08/17/21 07:04	08/21/21 19:13	1
Perfluorohexanesulfonic acid	<1.7		1.7	0.43	ng/L		08/17/21 07:04	08/21/21 19:13	1
Perfluoroctanesulfonic acid	<1.7		1.7	0.43	ng/L		08/17/21 07:04	08/21/21 19:13	1
NEtFOSAA	<1.7		1.7	0.43	ng/L		08/17/21 07:04	08/21/21 19:13	1
NMeFOSAA	<1.7		1.7	0.43	ng/L		08/17/21 07:04	08/21/21 19:13	1
Perfluoroundecanoic acid	<1.7		1.7	0.43	ng/L		08/17/21 07:04	08/21/21 19:13	1
Perfluorododecanoic acid	<1.7		1.7	0.43	ng/L		08/17/21 07:04	08/21/21 19:13	1
Surrogate		%Recovery	Qualifier	Limits			Prepared	Analyzed	Dil Fac
d5-NEtFOSAA		106		70 - 130			08/17/21 07:04	08/21/21 19:13	1
13C2 PFDA		114		70 - 130			08/17/21 07:04	08/21/21 19:13	1
13C2 PFHxA		113		70 - 130			08/17/21 07:04	08/21/21 19:13	1

Client Sample Results

Client: SUEZ Water Environmental Services Inc
Project/Site: Newberry System

Job ID: 410-51333-1

Client Sample ID: 7670061 301 Conley Between Lead & Lag

Lab Sample ID: 410-51333-5

Matrix: Drinking Water

Date Collected: 08/12/21 09:50

Date Received: 08/13/21 17:28

Method: EPA 537 Ver 1.1 - EPA 537 Version 1.1

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorohexanoic acid	13		1.7	0.44	ng/L		08/17/21 07:04	08/21/21 19:25	1
Perfluoroheptanoic acid	4.1		1.7	0.44	ng/L		08/17/21 07:04	08/21/21 19:25	1
Perfluoroctanoic acid	8.7		1.7	0.44	ng/L		08/17/21 07:04	08/21/21 19:25	1
Perfluorononanoic acid	<1.7		1.7	0.44	ng/L		08/17/21 07:04	08/21/21 19:25	1
Perfluorodecanoic acid	<1.7		1.7	0.44	ng/L		08/17/21 07:04	08/21/21 19:25	1
Perfluorotridecanoic acid	<1.7		1.7	0.44	ng/L		08/17/21 07:04	08/21/21 19:25	1
Perfluorotetradecanoic acid	<1.7		1.7	0.44	ng/L		08/17/21 07:04	08/21/21 19:25	1
Perfluorobutanesulfonic acid	4.8		1.7	0.44	ng/L		08/17/21 07:04	08/21/21 19:25	1
Perfluorohexanesulfonic acid	6.8		1.7	0.44	ng/L		08/17/21 07:04	08/21/21 19:25	1
Perfluoroctanesulfonic acid	11		1.7	0.44	ng/L		08/17/21 07:04	08/21/21 19:25	1
NEtFOSAA	<1.7		1.7	0.44	ng/L		08/17/21 07:04	08/21/21 19:25	1
NMeFOSAA	<1.7		1.7	0.44	ng/L		08/17/21 07:04	08/21/21 19:25	1
Perfluoroundecanoic acid	<1.7		1.7	0.44	ng/L		08/17/21 07:04	08/21/21 19:25	1
Perfluorododecanoic acid	<1.7		1.7	0.44	ng/L		08/17/21 07:04	08/21/21 19:25	1
Surrogate		%Recovery	Qualifier	Limits			Prepared	Analyzed	Dil Fac
d5-NEtFOSAA		104		70 - 130			08/17/21 07:04	08/21/21 19:25	1
13C2 PFDA		105		70 - 130			08/17/21 07:04	08/21/21 19:25	1
13C2 PFHxA		104		70 - 130			08/17/21 07:04	08/21/21 19:25	1

Client Sample ID: 7670061 301 Conley Between Lead & Lag

Lab Sample ID: 410-51333-6

FB

Date Collected: 08/12/21 09:50

Date Received: 08/13/21 17:28

Matrix: Potable Water

Method: EPA 537 Ver 1.1 - EPA 537 Version 1.1

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorohexanoic acid	<1.7		1.7	0.43	ng/L		08/17/21 07:04	08/21/21 19:36	1
Perfluoroheptanoic acid	<1.7		1.7	0.43	ng/L		08/17/21 07:04	08/21/21 19:36	1
Perfluoroctanoic acid	<1.7		1.7	0.43	ng/L		08/17/21 07:04	08/21/21 19:36	1
Perfluorononanoic acid	<1.7		1.7	0.43	ng/L		08/17/21 07:04	08/21/21 19:36	1
Perfluorodecanoic acid	<1.7		1.7	0.43	ng/L		08/17/21 07:04	08/21/21 19:36	1
Perfluorotridecanoic acid	<1.7		1.7	0.43	ng/L		08/17/21 07:04	08/21/21 19:36	1
Perfluorotetradecanoic acid	<1.7		1.7	0.43	ng/L		08/17/21 07:04	08/21/21 19:36	1
Perfluorobutanesulfonic acid	<1.7		1.7	0.43	ng/L		08/17/21 07:04	08/21/21 19:36	1
Perfluorohexanesulfonic acid	<1.7		1.7	0.43	ng/L		08/17/21 07:04	08/21/21 19:36	1
Perfluoroctanesulfonic acid	<1.7		1.7	0.43	ng/L		08/17/21 07:04	08/21/21 19:36	1
NEtFOSAA	<1.7		1.7	0.43	ng/L		08/17/21 07:04	08/21/21 19:36	1
NMeFOSAA	<1.7		1.7	0.43	ng/L		08/17/21 07:04	08/21/21 19:36	1
Perfluoroundecanoic acid	<1.7		1.7	0.43	ng/L		08/17/21 07:04	08/21/21 19:36	1
Perfluorododecanoic acid	<1.7		1.7	0.43	ng/L		08/17/21 07:04	08/21/21 19:36	1
Surrogate		%Recovery	Qualifier	Limits			Prepared	Analyzed	Dil Fac
d5-NEtFOSAA		107		70 - 130			08/17/21 07:04	08/21/21 19:36	1
13C2 PFDA		112		70 - 130			08/17/21 07:04	08/21/21 19:36	1
13C2 PFHxA		108		70 - 130			08/17/21 07:04	08/21/21 19:36	1

Client Sample Results

Client: SUEZ Water Environmental Services Inc
 Project/Site: Newberry System

Job ID: 410-51333-1

Client Sample ID: 7670061 301 Conley After Lag Vessel

Lab Sample ID: 410-51333-7

Matrix: Drinking Water

Date Collected: 08/12/21 09:45
 Date Received: 08/13/21 17:28

Method: EPA 537 Ver 1.1 - EPA 537 Version 1.1

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorohexanoic acid	13		1.7	0.43	ng/L		08/17/21 07:04	08/21/21 19:48	1
Perfluoroheptanoic acid	3.9		1.7	0.43	ng/L		08/17/21 07:04	08/21/21 19:48	1
Perfluoroctanoic acid	7.2		1.7	0.43	ng/L		08/17/21 07:04	08/21/21 19:48	1
Perflurooronanoic acid	<1.7		1.7	0.43	ng/L		08/17/21 07:04	08/21/21 19:48	1
Perfluorodecanoic acid	<1.7		1.7	0.43	ng/L		08/17/21 07:04	08/21/21 19:48	1
Perfluorotridecanoic acid	<1.7		1.7	0.43	ng/L		08/17/21 07:04	08/21/21 19:48	1
Perfluorotetradecanoic acid	<1.7		1.7	0.43	ng/L		08/17/21 07:04	08/21/21 19:48	1
Perfluorobutanesulfonic acid	5.1		1.7	0.43	ng/L		08/17/21 07:04	08/21/21 19:48	1
Perfluorohexanesulfonic acid	5.9		1.7	0.43	ng/L		08/17/21 07:04	08/21/21 19:48	1
Perfluoroctanesulfonic acid	6.9		1.7	0.43	ng/L		08/17/21 07:04	08/21/21 19:48	1
NEtFOSAA	<1.7		1.7	0.43	ng/L		08/17/21 07:04	08/21/21 19:48	1
NMeFOSAA	<1.7		1.7	0.43	ng/L		08/17/21 07:04	08/21/21 19:48	1
Perfluoroundecanoic acid	<1.7		1.7	0.43	ng/L		08/17/21 07:04	08/21/21 19:48	1
Perfluorododecanoic acid	<1.7		1.7	0.43	ng/L		08/17/21 07:04	08/21/21 19:48	1
Surrogate		%Recovery	Qualifier	Limits			Prepared	Analyzed	Dil Fac
d5-NEtFOSAA		108		70 - 130			08/17/21 07:04	08/21/21 19:48	1
13C2 PFDA		106		70 - 130			08/17/21 07:04	08/21/21 19:48	1
13C2 PFHxA		100		70 - 130			08/17/21 07:04	08/21/21 19:48	1

Client Sample ID: 7670061 301 Conley After Lag Vessel FB

Lab Sample ID: 410-51333-8

Matrix: Potable Water

Date Collected: 08/12/21 09:45
 Date Received: 08/13/21 17:28

Method: EPA 537 Ver 1.1 - EPA 537 Version 1.1

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorohexanoic acid	<1.8		1.8	0.44	ng/L		08/17/21 07:04	08/21/21 19:59	1
Perfluoroheptanoic acid	<1.8		1.8	0.44	ng/L		08/17/21 07:04	08/21/21 19:59	1
Perfluoroctanoic acid	<1.8		1.8	0.44	ng/L		08/17/21 07:04	08/21/21 19:59	1
Perflurooronanoic acid	<1.8		1.8	0.44	ng/L		08/17/21 07:04	08/21/21 19:59	1
Perfluorodecanoic acid	<1.8		1.8	0.44	ng/L		08/17/21 07:04	08/21/21 19:59	1
Perfluorotridecanoic acid	<1.8		1.8	0.44	ng/L		08/17/21 07:04	08/21/21 19:59	1
Perfluorotetradecanoic acid	<1.8		1.8	0.44	ng/L		08/17/21 07:04	08/21/21 19:59	1
Perfluorobutanesulfonic acid	<1.8		1.8	0.44	ng/L		08/17/21 07:04	08/21/21 19:59	1
Perfluorohexanesulfonic acid	<1.8		1.8	0.44	ng/L		08/17/21 07:04	08/21/21 19:59	1
Perfluoroctanesulfonic acid	<1.8		1.8	0.44	ng/L		08/17/21 07:04	08/21/21 19:59	1
NEtFOSAA	<1.8		1.8	0.44	ng/L		08/17/21 07:04	08/21/21 19:59	1
NMeFOSAA	<1.8		1.8	0.44	ng/L		08/17/21 07:04	08/21/21 19:59	1
Perfluoroundecanoic acid	<1.8		1.8	0.44	ng/L		08/17/21 07:04	08/21/21 19:59	1
Perfluorododecanoic acid	<1.8		1.8	0.44	ng/L		08/17/21 07:04	08/21/21 19:59	1
Surrogate		%Recovery	Qualifier	Limits			Prepared	Analyzed	Dil Fac
d5-NEtFOSAA		108		70 - 130			08/17/21 07:04	08/21/21 19:59	1
13C2 PFDA		110		70 - 130			08/17/21 07:04	08/21/21 19:59	1
13C2 PFHxA		107		70 - 130			08/17/21 07:04	08/21/21 19:59	1

Client Sample Results

Client: SUEZ Water Environmental Services Inc
Project/Site: Newberry System

Job ID: 410-51333-1

Client Sample ID: 7670061 101 Conley EP Grab Water

Lab Sample ID: 410-51333-9

Matrix: Drinking Water

Date Collected: 08/12/21 09:40
Date Received: 08/13/21 17:28

Method: EPA 537 Ver 1.1 - EPA 537 Version 1.1

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorohexanoic acid	13		1.7	0.43	ng/L		08/17/21 07:04	08/21/21 20:11	1
Perfluoroheptanoic acid	3.9		1.7	0.43	ng/L		08/17/21 07:04	08/21/21 20:11	1
Perfluorooctanoic acid	7.4		1.7	0.43	ng/L		08/17/21 07:04	08/21/21 20:11	1
Perfluorononanoic acid	<1.7		1.7	0.43	ng/L		08/17/21 07:04	08/21/21 20:11	1
Perfluorodecanoic acid	<1.7		1.7	0.43	ng/L		08/17/21 07:04	08/21/21 20:11	1
Perfluorotridecanoic acid	<1.7		1.7	0.43	ng/L		08/17/21 07:04	08/21/21 20:11	1
Perfluorotetradecanoic acid	<1.7		1.7	0.43	ng/L		08/17/21 07:04	08/21/21 20:11	1
Perfluorobutanesulfonic acid	5.1		1.7	0.43	ng/L		08/17/21 07:04	08/21/21 20:11	1
Perfluorohexanesulfonic acid	5.6		1.7	0.43	ng/L		08/17/21 07:04	08/21/21 20:11	1
Perfluorooctanesulfonic acid	5.8		1.7	0.43	ng/L		08/17/21 07:04	08/21/21 20:11	1
NEtFOSAA	<1.7		1.7	0.43	ng/L		08/17/21 07:04	08/21/21 20:11	1
NMeFOSAA	<1.7		1.7	0.43	ng/L		08/17/21 07:04	08/21/21 20:11	1
Perfluoroundecanoic acid	<1.7		1.7	0.43	ng/L		08/17/21 07:04	08/21/21 20:11	1
Perfluorododecanoic acid	<1.7		1.7	0.43	ng/L		08/17/21 07:04	08/21/21 20:11	1
Surrogate		%Recovery	Qualifier	Limits			Prepared	Analyzed	Dil Fac
d5-NEtFOSAA		118		70 - 130			08/17/21 07:04	08/21/21 20:11	1
13C2 PFDA		109		70 - 130			08/17/21 07:04	08/21/21 20:11	1
13C2 PFHxA		105		70 - 130			08/17/21 07:04	08/21/21 20:11	1

Client Sample ID: 7670061 101 Conley Field Blank Grab Water

Lab Sample ID: 410-51333-10

Matrix: Potable Water

Date Collected: 08/12/21 09:40
Date Received: 08/13/21 17:28

Method: EPA 537 Ver 1.1 - EPA 537 Version 1.1

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorohexanoic acid	<1.8		1.8	0.45	ng/L		08/17/21 07:04	08/21/21 20:23	1
Perfluoroheptanoic acid	<1.8		1.8	0.45	ng/L		08/17/21 07:04	08/21/21 20:23	1
Perfluorooctanoic acid	<1.8		1.8	0.45	ng/L		08/17/21 07:04	08/21/21 20:23	1
Perfluorononanoic acid	<1.8		1.8	0.45	ng/L		08/17/21 07:04	08/21/21 20:23	1
Perfluorodecanoic acid	<1.8		1.8	0.45	ng/L		08/17/21 07:04	08/21/21 20:23	1
Perfluorotridecanoic acid	<1.8		1.8	0.45	ng/L		08/17/21 07:04	08/21/21 20:23	1
Perfluorotetradecanoic acid	<1.8		1.8	0.45	ng/L		08/17/21 07:04	08/21/21 20:23	1
Perfluorobutanesulfonic acid	<1.8		1.8	0.45	ng/L		08/17/21 07:04	08/21/21 20:23	1
Perfluorohexanesulfonic acid	<1.8		1.8	0.45	ng/L		08/17/21 07:04	08/21/21 20:23	1
Perfluorooctanesulfonic acid	<1.8		1.8	0.45	ng/L		08/17/21 07:04	08/21/21 20:23	1
NEtFOSAA	<1.8		1.8	0.45	ng/L		08/17/21 07:04	08/21/21 20:23	1
NMeFOSAA	<1.8		1.8	0.45	ng/L		08/17/21 07:04	08/21/21 20:23	1
Perfluoroundecanoic acid	<1.8		1.8	0.45	ng/L		08/17/21 07:04	08/21/21 20:23	1
Perfluorododecanoic acid	<1.8		1.8	0.45	ng/L		08/17/21 07:04	08/21/21 20:23	1
Surrogate		%Recovery	Qualifier	Limits			Prepared	Analyzed	Dil Fac
d5-NEtFOSAA		104		70 - 130			08/17/21 07:04	08/21/21 20:23	1
13C2 PFDA		109		70 - 130			08/17/21 07:04	08/21/21 20:23	1
13C2 PFHxA		107		70 - 130			08/17/21 07:04	08/21/21 20:23	1

Client Sample Results

Client: SUEZ Water Environmental Services Inc
Project/Site: Newberry System

Job ID: 410-51333-1

Client Sample ID: 7670061 002 Coppersmith Well

Lab Sample ID: 410-51333-11

Date Collected: 08/12/21 08:50
Date Received: 08/13/21 17:28

Matrix: Drinking Water

Method: EPA 537 Ver 1.1 - EPA 537 Version 1.1

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorohexanoic acid	17		1.7	0.44	ng/L		08/19/21 05:42	08/21/21 21:32	1
Perfluoroheptanoic acid	8.3		1.7	0.44	ng/L		08/19/21 05:42	08/21/21 21:32	1
Perfluoroctanoic acid	10		1.7	0.44	ng/L		08/19/21 05:42	08/21/21 21:32	1
Perfluorononanoic acid	6.3		1.7	0.44	ng/L		08/19/21 05:42	08/21/21 21:32	1
Perfluorodecanoic acid	<1.7		1.7	0.44	ng/L		08/19/21 05:42	08/21/21 21:32	1
Perfluorotridecanoic acid	<1.7		1.7	0.44	ng/L		08/19/21 05:42	08/21/21 21:32	1
Perfluorotetradecanoic acid	<1.7		1.7	0.44	ng/L		08/19/21 05:42	08/21/21 21:32	1
Perfluorobutanesulfonic acid	8.0		1.7	0.44	ng/L		08/19/21 05:42	08/21/21 21:32	1
NEtFOSAA	<1.7		1.7	0.44	ng/L		08/19/21 05:42	08/21/21 21:32	1
NMeFOSAA	<1.7		1.7	0.44	ng/L		08/19/21 05:42	08/21/21 21:32	1
Perfluoroundecanoic acid	<1.7		1.7	0.44	ng/L		08/19/21 05:42	08/21/21 21:32	1
Perfluorododecanoic acid	<1.7		1.7	0.44	ng/L		08/19/21 05:42	08/21/21 21:32	1
Surrogate	%Recovery	Qualifier	Limits				Prepared	Analyzed	Dil Fac
d5-NEtFOSAA	91		70 - 130				08/19/21 05:42	08/21/21 21:32	1
13C2 PFDA	108		70 - 130				08/19/21 05:42	08/21/21 21:32	1
13C2 PFHxA	106		70 - 130				08/19/21 05:42	08/21/21 21:32	1

Method: EPA 537 Ver 1.1 - EPA 537 Version 1.1 - DL

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorohexanesulfonic acid	62		17	4.4	ng/L		08/19/21 05:42	08/24/21 15:17	10
Perfluoroctanesulfonic acid	68		17	4.4	ng/L		08/19/21 05:42	08/24/21 15:17	10
Surrogate	%Recovery	Qualifier	Limits				Prepared	Analyzed	Dil Fac
d5-NEtFOSAA	85		70 - 130				08/19/21 05:42	08/24/21 15:17	10
13C2 PFDA	93		70 - 130				08/19/21 05:42	08/24/21 15:17	10
13C2 PFHxA	104		70 - 130				08/19/21 05:42	08/24/21 15:17	10

Client Sample ID: 7670061 002 Coppersmith Well FB

Lab Sample ID: 410-51333-12

Date Collected: 08/12/21 08:50
Date Received: 08/13/21 17:28

Matrix: Potable Water

Method: EPA 537 Ver 1.1 - EPA 537 Version 1.1

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorohexanoic acid	<1.7		1.7	0.43	ng/L		08/19/21 05:42	08/21/21 21:44	1
Perfluoroheptanoic acid	<1.7		1.7	0.43	ng/L		08/19/21 05:42	08/21/21 21:44	1
Perfluoroctanoic acid	<1.7		1.7	0.43	ng/L		08/19/21 05:42	08/21/21 21:44	1
Perfluorononanoic acid	<1.7		1.7	0.43	ng/L		08/19/21 05:42	08/21/21 21:44	1
Perfluorodecanoic acid	<1.7		1.7	0.43	ng/L		08/19/21 05:42	08/21/21 21:44	1
Perfluorotridecanoic acid	<1.7		1.7	0.43	ng/L		08/19/21 05:42	08/21/21 21:44	1
Perfluorotetradecanoic acid	<1.7		1.7	0.43	ng/L		08/19/21 05:42	08/21/21 21:44	1
Perfluorobutanesulfonic acid	<1.7		1.7	0.43	ng/L		08/19/21 05:42	08/21/21 21:44	1
Perfluorohexanesulfonic acid	<1.7		1.7	0.43	ng/L		08/19/21 05:42	08/21/21 21:44	1
Perfluoroctanesulfonic acid	<1.7		1.7	0.43	ng/L		08/19/21 05:42	08/21/21 21:44	1
NEtFOSAA	<1.7		1.7	0.43	ng/L		08/19/21 05:42	08/21/21 21:44	1
NMeFOSAA	<1.7		1.7	0.43	ng/L		08/19/21 05:42	08/21/21 21:44	1
Perfluoroundecanoic acid	<1.7		1.7	0.43	ng/L		08/19/21 05:42	08/21/21 21:44	1
Perfluorododecanoic acid	<1.7		1.7	0.43	ng/L		08/19/21 05:42	08/21/21 21:44	1
Surrogate	%Recovery	Qualifier	Limits				Prepared	Analyzed	Dil Fac
d5-NEtFOSAA	98		70 - 130				08/19/21 05:42	08/21/21 21:44	1

Eurofins Lancaster Laboratories Env, LLC

Client Sample Results

Client: SUEZ Water Environmental Services Inc
 Project/Site: Newberry System

Job ID: 410-51333-1

Client Sample ID: 7670061 002 Coppersmith Well FB

Lab Sample ID: 410-51333-12

Matrix: Potable Water

Date Collected: 08/12/21 08:50
 Date Received: 08/13/21 17:28

Method: EPA 537 Ver 1.1 - EPA 537 Version 1.1 (Continued)

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C2 PFDA	114		70 - 130	08/19/21 05:42	08/21/21 21:44	1
13C2 PFHxA	115		70 - 130	08/19/21 05:42	08/21/21 21:44	1

Client Sample ID: 7670061 003 DuPont Well

Lab Sample ID: 410-51333-13

Matrix: Drinking Water

Date Collected: 08/12/21 09:20
 Date Received: 08/13/21 17:28

Method: EPA 537 Ver 1.1 - EPA 537 Version 1.1

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorohexanoic acid	7.5		1.8	0.45	ng/L	08/19/21 05:42	08/21/21 21:55	1	10
Perfluoroheptanoic acid	3.5		1.8	0.45	ng/L	08/19/21 05:42	08/21/21 21:55	1	11
Perfluoroctanoic acid	6.4		1.8	0.45	ng/L	08/19/21 05:42	08/21/21 21:55	1	12
Perfluorononanoic acid	<1.8		1.8	0.45	ng/L	08/19/21 05:42	08/21/21 21:55	1	13
Perfluorodecanoic acid	<1.8		1.8	0.45	ng/L	08/19/21 05:42	08/21/21 21:55	1	14
Perfluorotridecanoic acid	<1.8		1.8	0.45	ng/L	08/19/21 05:42	08/21/21 21:55	1	15
Perfluorotetradecanoic acid	<1.8		1.8	0.45	ng/L	08/19/21 05:42	08/21/21 21:55	1	
Perfluorobutanesulfonic acid	5.4		1.8	0.45	ng/L	08/19/21 05:42	08/21/21 21:55	1	
Perfluorohexanesulfonic acid	66		1.8	0.45	ng/L	08/19/21 05:42	08/21/21 21:55	1	
Perfluoroctanesulfonic acid	58		1.8	0.45	ng/L	08/19/21 05:42	08/21/21 21:55	1	
NEtFOSAA	<1.8		1.8	0.45	ng/L	08/19/21 05:42	08/21/21 21:55	1	
NMeFOSAA	<1.8		1.8	0.45	ng/L	08/19/21 05:42	08/21/21 21:55	1	
Perfluoroundecanoic acid	<1.8		1.8	0.45	ng/L	08/19/21 05:42	08/21/21 21:55	1	
Perfluorododecanoic acid	<1.8		1.8	0.45	ng/L	08/19/21 05:42	08/21/21 21:55	1	
Surrogate	%Recovery	Qualifier	Limits			D	Prepared	Analyzed	Dil Fac
d5-NEtFOSAA	88		70 - 130			08/19/21 05:42	08/21/21 21:55	1	
13C2 PFDA	108		70 - 130			08/19/21 05:42	08/21/21 21:55	1	
13C2 PFHxA	111		70 - 130			08/19/21 05:42	08/21/21 21:55	1	

Client Sample ID: 7670061 003 DuPont Well FB

Lab Sample ID: 410-51333-14

Matrix: Potable Water

Date Collected: 08/12/21 09:20
 Date Received: 08/13/21 17:28

Method: EPA 537 Ver 1.1 - EPA 537 Version 1.1

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorohexanoic acid	<1.8		1.8	0.44	ng/L	08/19/21 05:42	08/21/21 22:07	1	
Perfluoroheptanoic acid	<1.8		1.8	0.44	ng/L	08/19/21 05:42	08/21/21 22:07	1	
Perfluoroctanoic acid	<1.8		1.8	0.44	ng/L	08/19/21 05:42	08/21/21 22:07	1	
Perfluorononanoic acid	<1.8		1.8	0.44	ng/L	08/19/21 05:42	08/21/21 22:07	1	
Perfluorodecanoic acid	<1.8		1.8	0.44	ng/L	08/19/21 05:42	08/21/21 22:07	1	
Perfluorotridecanoic acid	<1.8		1.8	0.44	ng/L	08/19/21 05:42	08/21/21 22:07	1	
Perfluorotetradecanoic acid	<1.8		1.8	0.44	ng/L	08/19/21 05:42	08/21/21 22:07	1	
Perfluorobutanesulfonic acid	<1.8		1.8	0.44	ng/L	08/19/21 05:42	08/21/21 22:07	1	
Perfluorohexanesulfonic acid	<1.8		1.8	0.44	ng/L	08/19/21 05:42	08/21/21 22:07	1	
Perfluoroctanesulfonic acid	<1.8		1.8	0.44	ng/L	08/19/21 05:42	08/21/21 22:07	1	
NEtFOSAA	<1.8		1.8	0.44	ng/L	08/19/21 05:42	08/21/21 22:07	1	
NMeFOSAA	<1.8		1.8	0.44	ng/L	08/19/21 05:42	08/21/21 22:07	1	
Perfluoroundecanoic acid	<1.8		1.8	0.44	ng/L	08/19/21 05:42	08/21/21 22:07	1	
Perfluorododecanoic acid	<1.8		1.8	0.44	ng/L	08/19/21 05:42	08/21/21 22:07	1	

Client Sample Results

Client: SUEZ Water Environmental Services Inc
Project/Site: Newberry System

Job ID: 410-51333-1

Client Sample ID: 7670061 003 DuPont Well FB

Date Collected: 08/12/21 09:20
Date Received: 08/13/21 17:28

Lab Sample ID: 410-51333-14

Matrix: Potable Water

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
d5-NEtFOSAA	98		70 - 130	08/19/21 05:42	08/21/21 22:07	1
13C2 PFDA	117		70 - 130	08/19/21 05:42	08/21/21 22:07	1
13C2 PFHxA	116		70 - 130	08/19/21 05:42	08/21/21 22:07	1

Client Sample ID: 7670061 302 DuPont Between Lead & Lag

Date Collected: 08/12/21 09:15
Date Received: 08/13/21 17:28

Lab Sample ID: 410-51333-15

Matrix: Drinking Water

Method: EPA 537 Ver 1.1 - EPA 537 Version 1.1						
Analyte	Result	Qualifier	RL	MDL	Unit	D
Perfluorohexanoic acid	15		1.8	0.46	ng/L	08/19/21 05:42
Perfluoroheptanoic acid	2.4		1.8	0.46	ng/L	08/19/21 05:42
Perfluoroctanoic acid	<1.8		1.8	0.46	ng/L	08/19/21 05:42
Perfluorononanoic acid	<1.8		1.8	0.46	ng/L	08/19/21 05:42
Perfluorodecanoic acid	<1.8		1.8	0.46	ng/L	08/19/21 05:42
Perfluorotridecanoic acid	<1.8		1.8	0.46	ng/L	08/19/21 05:42
Perfluorotetradecanoic acid	<1.8		1.8	0.46	ng/L	08/19/21 05:42
Perfluorobutanesulfonic acid	5.0		1.8	0.46	ng/L	08/19/21 05:42
Perfluorohexanesulfonic acid	6.7		1.8	0.46	ng/L	08/19/21 05:42
Perfluoroctanesulfonic acid	2.0		1.8	0.46	ng/L	08/19/21 05:42
NEtFOSAA	<1.8		1.8	0.46	ng/L	08/19/21 05:42
NMeFOSAA	<1.8		1.8	0.46	ng/L	08/19/21 05:42
Perfluoroundecanoic acid	<1.8		1.8	0.46	ng/L	08/19/21 05:42
Perfluorododecanoic acid	<1.8		1.8	0.46	ng/L	08/19/21 05:42
Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
d5-NEtFOSAA	100		70 - 130	08/19/21 05:42	08/21/21 22:18	1
13C2 PFDA	105		70 - 130	08/19/21 05:42	08/21/21 22:18	1
13C2 PFHxA	111		70 - 130	08/19/21 05:42	08/21/21 22:18	1

Client Sample ID: 7670061 302 DuPont Between Lead & Lag

FB

Date Collected: 08/12/21 09:15
Date Received: 08/13/21 17:28

Lab Sample ID: 410-51333-16

Matrix: Potable Water

Method: EPA 537 Ver 1.1 - EPA 537 Version 1.1						
Analyte	Result	Qualifier	RL	MDL	Unit	D
Perfluorohexanoic acid	<1.8		1.8	0.44	ng/L	08/19/21 05:42
Perfluoroheptanoic acid	<1.8		1.8	0.44	ng/L	08/19/21 05:42
Perfluoroctanoic acid	<1.8		1.8	0.44	ng/L	08/19/21 05:42
Perfluorononanoic acid	<1.8		1.8	0.44	ng/L	08/19/21 05:42
Perfluorodecanoic acid	<1.8		1.8	0.44	ng/L	08/19/21 05:42
Perfluorotridecanoic acid	<1.8		1.8	0.44	ng/L	08/19/21 05:42
Perfluorotetradecanoic acid	<1.8		1.8	0.44	ng/L	08/19/21 05:42
Perfluorobutanesulfonic acid	<1.8		1.8	0.44	ng/L	08/19/21 05:42
Perfluorohexanesulfonic acid	<1.8		1.8	0.44	ng/L	08/19/21 05:42
Perfluoroctanesulfonic acid	<1.8		1.8	0.44	ng/L	08/19/21 05:42
NEtFOSAA	<1.8		1.8	0.44	ng/L	08/19/21 05:42
NMeFOSAA	<1.8		1.8	0.44	ng/L	08/19/21 05:42
Perfluoroundecanoic acid	<1.8		1.8	0.44	ng/L	08/19/21 05:42
Perfluorododecanoic acid	<1.8		1.8	0.44	ng/L	08/19/21 05:42

Eurofins Lancaster Laboratories Env, LLC

Client Sample Results

Client: SUEZ Water Environmental Services Inc
Project/Site: Newberry System

Job ID: 410-51333-1

Client Sample ID: 7670061 302 DuPont Between Lead & Lag

FB

Date Collected: 08/12/21 09:15

Date Received: 08/13/21 17:28

Lab Sample ID: 410-51333-16

Matrix: Potable Water

Surrogate	%Recovery	Qualifier	Limits
d5-NEtFOSAA	96		70 - 130
13C2 PFDA	111		70 - 130
13C2 PFHxA	116		70 - 130

Lab Sample ID: 410-51333-17

Matrix: Drinking Water

Client Sample ID: 7670061 302 DuPont After Lag Vessel

Date Collected: 08/12/21 09:10

Date Received: 08/13/21 17:28

Method: EPA 537 Ver 1.1 - EPA 537 Version 1.1

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorohexanoic acid	2.8		1.8	0.45	ng/L	08/19/21 05:42	08/21/21 22:42	1	
Perfluoroheptanoic acid	<1.8		1.8	0.45	ng/L	08/19/21 05:42	08/21/21 22:42	1	
Perfluoroctanoic acid	<1.8		1.8	0.45	ng/L	08/19/21 05:42	08/21/21 22:42	1	
Perfluorononanoic acid	<1.8		1.8	0.45	ng/L	08/19/21 05:42	08/21/21 22:42	1	
Perfluorodecanoic acid	<1.8		1.8	0.45	ng/L	08/19/21 05:42	08/21/21 22:42	1	
Perfluorotridecanoic acid	<1.8		1.8	0.45	ng/L	08/19/21 05:42	08/21/21 22:42	1	
Perfluorotetradecanoic acid	<1.8		1.8	0.45	ng/L	08/19/21 05:42	08/21/21 22:42	1	
Perfluorobutanesulfonic acid	<1.8		1.8	0.45	ng/L	08/19/21 05:42	08/21/21 22:42	1	
Perfluorohexanesulfonic acid	<1.8		1.8	0.45	ng/L	08/19/21 05:42	08/21/21 22:42	1	
Perfluoroctanesulfonic acid	<1.8		1.8	0.45	ng/L	08/19/21 05:42	08/21/21 22:42	1	
NEtFOSAA	<1.8		1.8	0.45	ng/L	08/19/21 05:42	08/21/21 22:42	1	
NMeFOSAA	<1.8		1.8	0.45	ng/L	08/19/21 05:42	08/21/21 22:42	1	
Perfluoroundecanoic acid	<1.8		1.8	0.45	ng/L	08/19/21 05:42	08/21/21 22:42	1	
Perfluorododecanoic acid	<1.8		1.8	0.45	ng/L	08/19/21 05:42	08/21/21 22:42	1	
Surrogate	%Recovery	Qualifier	Limits				Prepared	Analyzed	Dil Fac
d5-NEtFOSAA	107		70 - 130				08/19/21 05:42	08/21/21 22:42	1
13C2 PFDA	103		70 - 130				08/19/21 05:42	08/21/21 22:42	1
13C2 PFHxA	107		70 - 130				08/19/21 05:42	08/21/21 22:42	1

Client Sample ID: 7670061 302 DuPont After Lag Vessel FB

Date Collected: 08/12/21 09:10

Date Received: 08/13/21 17:28

Lab Sample ID: 410-51333-18

Matrix: Potable Water

Method: EPA 537 Ver 1.1 - EPA 537 Version 1.1

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorohexanoic acid	<1.8		1.8	0.44	ng/L	08/19/21 05:42	08/21/21 22:53	1	
Perfluoroheptanoic acid	<1.8		1.8	0.44	ng/L	08/19/21 05:42	08/21/21 22:53	1	
Perfluoroctanoic acid	<1.8		1.8	0.44	ng/L	08/19/21 05:42	08/21/21 22:53	1	
Perfluorononanoic acid	<1.8		1.8	0.44	ng/L	08/19/21 05:42	08/21/21 22:53	1	
Perfluorodecanoic acid	<1.8		1.8	0.44	ng/L	08/19/21 05:42	08/21/21 22:53	1	
Perfluorotridecanoic acid	<1.8		1.8	0.44	ng/L	08/19/21 05:42	08/21/21 22:53	1	
Perfluorotetradecanoic acid	<1.8		1.8	0.44	ng/L	08/19/21 05:42	08/21/21 22:53	1	
Perfluorobutanesulfonic acid	<1.8		1.8	0.44	ng/L	08/19/21 05:42	08/21/21 22:53	1	
Perfluorohexanesulfonic acid	<1.8		1.8	0.44	ng/L	08/19/21 05:42	08/21/21 22:53	1	
Perfluoroctanesulfonic acid	<1.8		1.8	0.44	ng/L	08/19/21 05:42	08/21/21 22:53	1	
NEtFOSAA	<1.8		1.8	0.44	ng/L	08/19/21 05:42	08/21/21 22:53	1	
NMeFOSAA	<1.8		1.8	0.44	ng/L	08/19/21 05:42	08/21/21 22:53	1	
Perfluoroundecanoic acid	<1.8		1.8	0.44	ng/L	08/19/21 05:42	08/21/21 22:53	1	
Perfluorododecanoic acid	<1.8		1.8	0.44	ng/L	08/19/21 05:42	08/21/21 22:53	1	

Eurofins Lancaster Laboratories Env, LLC

Client Sample Results

Client: SUEZ Water Environmental Services Inc
 Project/Site: Newberry System

Job ID: 410-51333-1

Client Sample ID: 7670061 302 DuPont After Lag Vessel FB

Lab Sample ID: 410-51333-18

Date Collected: 08/12/21 09:10
 Date Received: 08/13/21 17:28

Matrix: Potable Water

Surrogate	%Recovery	Qualifier	Limits
d5-NEtFOSAA	100		70 - 130
13C2 PFDA	107		70 - 130
13C2 PFHxA	110		70 - 130

Prepared Analyzed Dil Fac

08/19/21 05:42	08/21/21 22:53	1
08/19/21 05:42	08/21/21 22:53	1
08/19/21 05:42	08/21/21 22:53	1

Client Sample ID: 7670061 102 DuPont EP

Lab Sample ID: 410-51333-19

Date Collected: 08/12/21 09:05
 Date Received: 08/13/21 17:28

Matrix: Drinking Water

Method: EPA 537 Ver 1.1 - EPA 537 Version 1.1

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorohexanoic acid	2.5		1.8	0.45	ng/L		08/19/21 05:42	08/21/21 23:05	1
Perfluoroheptanoic acid	<1.8		1.8	0.45	ng/L		08/19/21 05:42	08/21/21 23:05	1
Perfluoroctanoic acid	<1.8		1.8	0.45	ng/L		08/19/21 05:42	08/21/21 23:05	1
Perfluorononanoic acid	<1.8		1.8	0.45	ng/L		08/19/21 05:42	08/21/21 23:05	1
Perfluorodecanoic acid	<1.8		1.8	0.45	ng/L		08/19/21 05:42	08/21/21 23:05	1
Perfluorotridecanoic acid	<1.8		1.8	0.45	ng/L		08/19/21 05:42	08/21/21 23:05	1
Perfluorotetradecanoic acid	<1.8		1.8	0.45	ng/L		08/19/21 05:42	08/21/21 23:05	1
Perfluorobutanesulfonic acid	<1.8		1.8	0.45	ng/L		08/19/21 05:42	08/21/21 23:05	1
Perfluorohexanesulfonic acid	<1.8		1.8	0.45	ng/L		08/19/21 05:42	08/21/21 23:05	1
Perfluoroctanesulfonic acid	<1.8		1.8	0.45	ng/L		08/19/21 05:42	08/21/21 23:05	1
NEtFOSAA	<1.8		1.8	0.45	ng/L		08/19/21 05:42	08/21/21 23:05	1
NMeFOSAA	<1.8		1.8	0.45	ng/L		08/19/21 05:42	08/21/21 23:05	1
Perfluoroundecanoic acid	<1.8		1.8	0.45	ng/L		08/19/21 05:42	08/21/21 23:05	1
Perfluorododecanoic acid	<1.8		1.8	0.45	ng/L		08/19/21 05:42	08/21/21 23:05	1

Surrogate	%Recovery	Qualifier	Limits
d5-NEtFOSAA	105		70 - 130
13C2 PFDA	100		70 - 130
13C2 PFHxA	101		70 - 130

Prepared Analyzed Dil Fac

08/19/21 05:42	08/21/21 23:05	1
08/19/21 05:42	08/21/21 23:05	1
08/19/21 05:42	08/21/21 23:05	1

Client Sample ID: 7670061 102 DuPont EP FB

Lab Sample ID: 410-51333-20

Date Collected: 08/12/21 09:05
 Date Received: 08/13/21 17:28

Matrix: Potable Water

Method: EPA 537 Ver 1.1 - EPA 537 Version 1.1

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorohexanoic acid	<1.8		1.8	0.44	ng/L		08/19/21 05:42	08/21/21 23:16	1
Perfluoroheptanoic acid	<1.8		1.8	0.44	ng/L		08/19/21 05:42	08/21/21 23:16	1
Perfluoroctanoic acid	<1.8		1.8	0.44	ng/L		08/19/21 05:42	08/21/21 23:16	1
Perfluorononanoic acid	<1.8		1.8	0.44	ng/L		08/19/21 05:42	08/21/21 23:16	1
Perfluorodecanoic acid	<1.8		1.8	0.44	ng/L		08/19/21 05:42	08/21/21 23:16	1
Perfluorotridecanoic acid	<1.8		1.8	0.44	ng/L		08/19/21 05:42	08/21/21 23:16	1
Perfluorotetradecanoic acid	<1.8		1.8	0.44	ng/L		08/19/21 05:42	08/21/21 23:16	1
Perfluorobutanesulfonic acid	<1.8		1.8	0.44	ng/L		08/19/21 05:42	08/21/21 23:16	1
Perfluorohexanesulfonic acid	<1.8		1.8	0.44	ng/L		08/19/21 05:42	08/21/21 23:16	1
Perfluoroctanesulfonic acid	<1.8		1.8	0.44	ng/L		08/19/21 05:42	08/21/21 23:16	1
NEtFOSAA	<1.8		1.8	0.44	ng/L		08/19/21 05:42	08/21/21 23:16	1
NMeFOSAA	<1.8		1.8	0.44	ng/L		08/19/21 05:42	08/21/21 23:16	1
Perfluoroundecanoic acid	<1.8		1.8	0.44	ng/L		08/19/21 05:42	08/21/21 23:16	1
Perfluorododecanoic acid	<1.8		1.8	0.44	ng/L		08/19/21 05:42	08/21/21 23:16	1

Client Sample Results

Client: SUEZ Water Environmental Services Inc
Project/Site: Newberry System

Job ID: 410-51333-1

Client Sample ID: 7670061 102 DuPont EP FB
Date Collected: 08/12/21 09:05
Date Received: 08/13/21 17:28

Lab Sample ID: 410-51333-20
Matrix: Potable Water

Surrogate	%Recovery	Qualifier	Limits
d5-NETFOSAA	105		70 - 130
13C2 PFDA	105		70 - 130
13C2 PFHxA	103		70 - 130

Prepared	Analyzed	Dil Fac
08/19/21 05:42	08/21/21 23:16	1
08/19/21 05:42	08/21/21 23:16	1
08/19/21 05:42	08/21/21 23:16	1

Surrogate Summary

Client: SUEZ Water Environmental Services Inc
 Project/Site: Newberry System

Job ID: 410-51333-1

Method: EPA 537 Ver 1.1 - EPA 537 Version 1.1

Matrix: Drinking Water

Prep Type: Total/NA

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)		
		d5NEFOS (70-130)	PFDA (70-130)	PFHxA (70-130)
410-51333-1	7670061 001 Playground Well	109	112	102
410-51333-3	7670061 005 Conley Well	108	113	107
410-51333-5	7670061 301 Conley Between	104	105	104
	Lead & Lag			
410-51333-7	7670061 301 Conley After Lag Vessel	108	106	100
410-51333-9	7670061 101 Conley EP Grab Water	118	109	105
410-51333-11	7670061 002 Coppersmith Well	91	108	106
410-51333-11 - DL	7670061 002 Coppersmith Well	85	93	104
410-51333-13	7670061 003 DuPont Well	88	108	111
410-51333-15	7670061 302 DuPont Between Lead & Lag	100	105	111
410-51333-17	7670061 302 DuPont After Lag Vessel	107	103	107
410-51333-19	7670061 102 DuPont EP	105	100	101
LCS 410-161150/2-A	Lab Control Sample	116	108	105
LCS 410-162148/2-A	Lab Control Sample	97	118	120
LCSD 410-161150/3-A	Lab Control Sample Dup	119	108	104
LCSD 410-162148/3-A	Lab Control Sample Dup	104	104	102
LLCS 410-161150/4-A	Lab Control Sample	117	113	107
LLCS 410-162148/4-A	Lab Control Sample	92	107	114
MB 410-161150/1-A	Method Blank	112	107	105
MB 410-162148/1-A	Method Blank	111	122	123

Surrogate Legend

d5NEFOS = d5-NETFOSAA

PFDA = 13C2 PFDA

PFHxA = 13C2 PFHxA

Method: EPA 537 Ver 1.1 - EPA 537 Version 1.1

Matrix: Potable Water

Prep Type: Total/NA

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)		
		d5NEFOS (70-130)	PFDA (70-130)	PFHxA (70-130)
410-51333-2	7670061 001 Playground Well FB	109	112	110
410-51333-4	7670061 005 Conley Well FB	106	114	113
410-51333-6	7670061 301 Conley Between Lead & Lag FB	107	112	108
410-51333-8	7670061 301 Conley After Lag Vessel FB	108	110	107
410-51333-10	7670061 101 Conley Field Blank Grab Water	104	109	107
410-51333-12	7670061 002 Coppersmith Well FB	98	114	115
410-51333-14	7670061 003 DuPont Well FB	98	117	116
410-51333-16	7670061 302 DuPont Between Lead & Lag FB	96	111	116
410-51333-18	7670061 302 DuPont After Lag Vessel FB	100	107	110
410-51333-20	7670061 102 DuPont EP FB	105	105	103

Surrogate Legend

Eurofins Lancaster Laboratories Env, LLC

Surrogate Summary

Client: SUEZ Water Environmental Services Inc

Project/Site: Newberry System

d5NEFOS = d5-NEtFOSAA

PFDA = 13C2 PFDA

PFHxA = 13C2 PFHxA

Job ID: 410-51333-1

1

2

3

4

5

6

7

8

9

10

11

12

13

14

15

QC Sample Results

Client: SUEZ Water Environmental Services Inc
Project/Site: Newberry System

Job ID: 410-51333-1

Method: EPA 537 Ver 1.1 - EPA 537 Version 1.1

Lab Sample ID: MB 410-161150/1-A

Matrix: Drinking Water

Analysis Batch: 163138

Client Sample ID: Method Blank

Prep Type: Total/NA

Prep Batch: 161150

Analyte	MB Result	MB Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorohexanoic acid	<2.0		2.0	0.50	ng/L		08/17/21 07:04	08/21/21 15:45	1
Perfluoroheptanoic acid	<2.0		2.0	0.50	ng/L		08/17/21 07:04	08/21/21 15:45	1
Perfluoroctanoic acid	<2.0		2.0	0.50	ng/L		08/17/21 07:04	08/21/21 15:45	1
Perfluorononanoic acid	<2.0		2.0	0.50	ng/L		08/17/21 07:04	08/21/21 15:45	1
Perfluorodecanoic acid	<2.0		2.0	0.50	ng/L		08/17/21 07:04	08/21/21 15:45	1
Perfluorotridecanoic acid	<2.0		2.0	0.50	ng/L		08/17/21 07:04	08/21/21 15:45	1
Perfluorotetradecanoic acid	<2.0		2.0	0.50	ng/L		08/17/21 07:04	08/21/21 15:45	1
Perfluorobutanesulfonic acid	<2.0		2.0	0.50	ng/L		08/17/21 07:04	08/21/21 15:45	1
Perfluorohexanesulfonic acid	<2.0		2.0	0.50	ng/L		08/17/21 07:04	08/21/21 15:45	1
Perfluoroctanesulfonic acid	<2.0		2.0	0.50	ng/L		08/17/21 07:04	08/21/21 15:45	1
NEtFOSAA	<2.0		2.0	0.50	ng/L		08/17/21 07:04	08/21/21 15:45	1
NMeFOSAA	<2.0		2.0	0.50	ng/L		08/17/21 07:04	08/21/21 15:45	1
Perfluoroundecanoic acid	<2.0		2.0	0.50	ng/L		08/17/21 07:04	08/21/21 15:45	1
Perfluorododecanoic acid	<2.0		2.0	0.50	ng/L		08/17/21 07:04	08/21/21 15:45	1

Surrogate	MB %Recovery	MB Qualifier	MB Limits	Prepared	Analyzed	Dil Fac
d5-NEtFOSAA	112		70 - 130	08/17/21 07:04	08/21/21 15:45	1
13C2 PFDA	107		70 - 130	08/17/21 07:04	08/21/21 15:45	1
13C2 PFHxA	105		70 - 130	08/17/21 07:04	08/21/21 15:45	1

Lab Sample ID: LCS 410-161150/2-A

Matrix: Drinking Water

Analysis Batch: 163138

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 161150

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	Limits
Perfluorohexanoic acid	20.5	21.1		ng/L		103	70 - 130
Perfluoroheptanoic acid	20.5	21.2		ng/L		104	70 - 130
Perfluoroctanoic acid	20.5	21.2		ng/L		104	70 - 130
Perfluorononanoic acid	20.5	21.4		ng/L		104	70 - 130
Perfluorodecanoic acid	20.5	21.5		ng/L		105	70 - 130
Perfluorotridecanoic acid	20.5	21.8		ng/L		107	70 - 130
Perfluorotetradecanoic acid	20.5	21.6		ng/L		106	70 - 130
Perfluorobutanesulfonic acid	18.1	19.7		ng/L		108	70 - 130
Perfluorohexanesulfonic acid	18.7	20.4		ng/L		109	70 - 130
Perfluoroctanesulfonic acid	19.0	21.0		ng/L		111	70 - 130
NEtFOSAA	20.5	22.0		ng/L		108	70 - 130
NMeFOSAA	20.5	21.2		ng/L		104	70 - 130
Perfluoroundecanoic acid	20.5	20.8		ng/L		102	70 - 130
Perfluorododecanoic acid	20.5	21.1		ng/L		103	70 - 130

Surrogate	LCS %Recovery	LCS Qualifier	Limits
d5-NEtFOSAA	116		70 - 130
13C2 PFDA	108		70 - 130
13C2 PFHxA	105		70 - 130

QC Sample Results

Client: SUEZ Water Environmental Services Inc
 Project/Site: Newberry System

Job ID: 410-51333-1

Method: EPA 537 Ver 1.1 - EPA 537 Version 1.1 (Continued)

Lab Sample ID: LCSD 410-161150/3-A

Matrix: Drinking Water

Analysis Batch: 163138

Client Sample ID: Lab Control Sample Dup

Prep Type: Total/NA

Prep Batch: 161150

Analyte	Spike Added	LCSD Result	LCSD Qualifier	Unit	D	%Rec	Limits	RPD	RPD Limit
Perfluorohexanoic acid	20.5	20.5		ng/L		100	70 - 130	3	30
Perfluoroheptanoic acid	20.5	21.0		ng/L		103	70 - 130	1	30
Perfluoroctanoic acid	20.5	21.2		ng/L		104	70 - 130	0	30
Perfluorononanoic acid	20.5	21.0		ng/L		102	70 - 130	2	30
Perfluorodecanoic acid	20.5	20.9		ng/L		102	70 - 130	3	30
Perfluorotridecanoic acid	20.5	21.2		ng/L		103	70 - 130	3	30
Perfluorotetradecanoic acid	20.5	21.4		ng/L		104	70 - 130	1	30
Perfluorobutanesulfonic acid	18.1	20.0		ng/L		110	70 - 130	2	30
Perfluorohexanesulfonic acid	18.7	20.3		ng/L		109	70 - 130	0	30
Perfluoroctanesulfonic acid	19.0	20.7		ng/L		109	70 - 130	1	30
NEtFOSAA	20.5	22.6		ng/L		110	70 - 130	2	30
NMeFOSAA	20.5	21.6		ng/L		105	70 - 130	2	30
Perfluoroundecanoic acid	20.5	21.1		ng/L		103	70 - 130	1	30
Perfluorododecanoic acid	20.5	21.0		ng/L		103	70 - 130	0	30

Surrogate	LCSD	LCSD	Limits
	%Recovery	Qualifier	
d5-NEtFOSAA	119		70 - 130
13C2 PFDA	108		70 - 130
13C2 PFHxA	104		70 - 130

Lab Sample ID: LLCS 410-161150/4-A

Matrix: Drinking Water

Analysis Batch: 163138

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 161150

Analyte	Spike Added	LLCS Result	LLCS Qualifier	Unit	D	%Rec	Limits
Perfluorohexanoic acid	1.92	1.99	J	ng/L		104	50 - 150
Perfluoroheptanoic acid	1.92	1.97	J	ng/L		102	50 - 150
Perfluoroctanoic acid	1.92	2.12		ng/L		110	50 - 150
Perfluorononanoic acid	1.92	2.00		ng/L		104	50 - 150
Perfluorodecanoic acid	1.92	1.99	J	ng/L		104	50 - 150
Perfluorotridecanoic acid	1.92	1.94	J	ng/L		101	50 - 150
Perfluorotetradecanoic acid	1.92	1.85	J	ng/L		97	50 - 150
Perfluorobutanesulfonic acid	1.70	1.82	J	ng/L		107	50 - 150
Perfluorohexanesulfonic acid	1.75	1.73	J	ng/L		99	50 - 150
Perfluoroctanesulfonic acid	1.78	1.86	J	ng/L		105	50 - 150
NEtFOSAA	1.92	2.13		ng/L		111	50 - 150
NMeFOSAA	1.92	1.98	J	ng/L		103	50 - 150
Perfluoroundecanoic acid	1.92	2.00		ng/L		104	50 - 150
Perfluorododecanoic acid	1.92	1.95	J	ng/L		101	50 - 150

Surrogate	LLCS	LLCS	Limits
	%Recovery	Qualifier	
d5-NEtFOSAA	117		70 - 130
13C2 PFDA	113		70 - 130
13C2 PFHxA	107		70 - 130

QC Sample Results

Client: SUEZ Water Environmental Services Inc
Project/Site: Newberry System

Job ID: 410-51333-1

Method: EPA 537 Ver 1.1 - EPA 537 Version 1.1 (Continued)

Lab Sample ID: MB 410-162148/1-A

Matrix: Drinking Water

Analysis Batch: 163138

Client Sample ID: Method Blank

Prep Type: Total/NA

Prep Batch: 162148

Analyte	MB Result	MB Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorohexanoic acid	<2.0		2.0	0.50	ng/L		08/19/21 05:42	08/21/21 20:46	1
Perfluoroheptanoic acid	<2.0		2.0	0.50	ng/L		08/19/21 05:42	08/21/21 20:46	1
Perfluoroctanoic acid	<2.0		2.0	0.50	ng/L		08/19/21 05:42	08/21/21 20:46	1
Perfluorononanoic acid	<2.0		2.0	0.50	ng/L		08/19/21 05:42	08/21/21 20:46	1
Perfluorodecanoic acid	<2.0		2.0	0.50	ng/L		08/19/21 05:42	08/21/21 20:46	1
Perfluorotridecanoic acid	<2.0		2.0	0.50	ng/L		08/19/21 05:42	08/21/21 20:46	1
Perfluorotetradecanoic acid	<2.0		2.0	0.50	ng/L		08/19/21 05:42	08/21/21 20:46	1
Perfluorobutanesulfonic acid	<2.0		2.0	0.50	ng/L		08/19/21 05:42	08/21/21 20:46	1
Perfluorohexanesulfonic acid	<2.0		2.0	0.50	ng/L		08/19/21 05:42	08/21/21 20:46	1
Perfluoroctanesulfonic acid	<2.0		2.0	0.50	ng/L		08/19/21 05:42	08/21/21 20:46	1
NEtFOSAA	<2.0		2.0	0.50	ng/L		08/19/21 05:42	08/21/21 20:46	1
NMeFOSAA	<2.0		2.0	0.50	ng/L		08/19/21 05:42	08/21/21 20:46	1
Perfluoroundecanoic acid	<2.0		2.0	0.50	ng/L		08/19/21 05:42	08/21/21 20:46	1
Perfluorododecanoic acid	<2.0		2.0	0.50	ng/L		08/19/21 05:42	08/21/21 20:46	1

MB MB

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
d5-NEtFOSAA	111		70 - 130	08/19/21 05:42	08/21/21 20:46	1
13C2 PFDA	122		70 - 130	08/19/21 05:42	08/21/21 20:46	1
13C2 PFHxA	123		70 - 130	08/19/21 05:42	08/21/21 20:46	1

Lab Sample ID: LCS 410-162148/2-A

Matrix: Drinking Water

Analysis Batch: 163138

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 162148

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	Limits
Perfluorohexanoic acid	20.5	24.3		ng/L		118	70 - 130
Perfluoroheptanoic acid	20.5	24.6		ng/L		120	70 - 130
Perfluoroctanoic acid	20.5	24.4		ng/L		119	70 - 130
Perfluorononanoic acid	20.5	24.4		ng/L		119	70 - 130
Perfluorodecanoic acid	20.5	23.4		ng/L		114	70 - 130
Perfluorotridecanoic acid	20.5	24.6		ng/L		120	70 - 130
Perfluorotetradecanoic acid	20.5	24.2		ng/L		118	70 - 130
Perfluorobutanesulfonic acid	18.1	21.1		ng/L		117	70 - 130
Perfluorohexanesulfonic acid	18.7	21.4		ng/L		115	70 - 130
Perfluoroctanesulfonic acid	19.0	21.4		ng/L		113	70 - 130
NEtFOSAA	20.5	20.3		ng/L		99	70 - 130
NMeFOSAA	20.5	20.0		ng/L		97	70 - 130
Perfluoroundecanoic acid	20.5	24.1		ng/L		118	70 - 130
Perfluorododecanoic acid	20.5	24.7		ng/L		120	70 - 130

LCS LCS

Surrogate	%Recovery	Qualifier	Limits
d5-NEtFOSAA	97		70 - 130
13C2 PFDA	118		70 - 130
13C2 PFHxA	120		70 - 130

Eurofins Lancaster Laboratories Env, LLC

QC Sample Results

Client: SUEZ Water Environmental Services Inc
 Project/Site: Newberry System

Job ID: 410-51333-1

Method: EPA 537 Ver 1.1 - EPA 537 Version 1.1 (Continued)

Lab Sample ID: LCSD 410-162148/3-A

Matrix: Drinking Water

Analysis Batch: 163138

Client Sample ID: Lab Control Sample Dup

Prep Type: Total/NA

Prep Batch: 162148

Analyte	Spike Added	LCSD Result	LCSD Qualifier	Unit	D	%Rec	Limits	RPD	RPD Limit
Perfluorohexanoic acid	20.5	20.7		ng/L		101	70 - 130	16	30
Perfluoroheptanoic acid	20.5	20.9		ng/L		102	70 - 130	16	30
Perfluoroctanoic acid	20.5	21.2		ng/L		103	70 - 130	14	30
Perfluorononanoic acid	20.5	21.5		ng/L		105	70 - 130	13	30
Perfluorodecanoic acid	20.5	21.0		ng/L		103	70 - 130	11	30
Perfluorotridecanoic acid	20.5	20.8		ng/L		102	70 - 130	17	30
Perfluorotetradecanoic acid	20.5	20.8		ng/L		101	70 - 130	16	30
Perfluorobutanesulfonic acid	18.1	20.1		ng/L		111	70 - 130	5	30
Perfluorohexanesulfonic acid	18.7	20.6		ng/L		110	70 - 130	4	30
Perfluoroctanesulfonic acid	19.0	20.6		ng/L		109	70 - 130	4	30
NEtFOSAA	20.5	21.0		ng/L		103	70 - 130	4	30
NMeFOSAA	20.5	20.5		ng/L		100	70 - 130	3	30
Perfluoroundecanoic acid	20.5	21.2		ng/L		103	70 - 130	13	30
Perfluorododecanoic acid	20.5	21.3		ng/L		104	70 - 130	15	30

Surrogate	LCSD	LCSD	Limits
	%Recovery	Qualifier	
d5-NEtFOSAA	104		70 - 130
13C2 PFDA	104		70 - 130
13C2 PFHxA	102		70 - 130

Lab Sample ID: LLCS 410-162148/4-A

Matrix: Drinking Water

Analysis Batch: 163138

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 162148

Analyte	Spike Added	LLCS Result	LLCS Qualifier	Unit	D	%Rec	Limits
Perfluorohexanoic acid	1.92	2.19		ng/L		114	50 - 150
Perfluoroheptanoic acid	1.92	2.23		ng/L		116	50 - 150
Perfluoroctanoic acid	1.92	2.27		ng/L		118	50 - 150
Perfluorononanoic acid	1.92	2.19		ng/L		114	50 - 150
Perfluorodecanoic acid	1.92	2.02		ng/L		105	50 - 150
Perfluorotridecanoic acid	1.92	1.78 J		ng/L		92	50 - 150
Perfluorotetradecanoic acid	1.92	1.87 J		ng/L		97	50 - 150
Perfluorobutanesulfonic acid	1.70	1.94 J		ng/L		114	50 - 150
Perfluorohexanesulfonic acid	1.75	2.00		ng/L		114	50 - 150
Perfluoroctanesulfonic acid	1.78	1.95 J		ng/L		110	50 - 150
NEtFOSAA	1.92	1.70 J		ng/L		89	50 - 150
NMeFOSAA	1.92	1.84 J		ng/L		96	50 - 150
Perfluoroundecanoic acid	1.92	1.92 J		ng/L		100	50 - 150
Perfluorododecanoic acid	1.92	1.87 J		ng/L		97	50 - 150

Surrogate	LLCS	LLCS	Limits
	%Recovery	Qualifier	
d5-NEtFOSAA	92		70 - 130
13C2 PFDA	107		70 - 130
13C2 PFHxA	114		70 - 130

QC Association Summary

Client: SUEZ Water Environmental Services Inc
Project/Site: Newberry System

Job ID: 410-51333-1

LCMS

Prep Batch: 161150

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
410-51333-1	7670061 001 Playground Well	Total/NA	Drinking Water	EPA 537 Ver 1.1	
410-51333-2	7670061 001 Playground Well FB	Total/NA	Potable Water	EPA 537 Ver 1.1	
410-51333-3	7670061 005 Conley Well	Total/NA	Drinking Water	EPA 537 Ver 1.1	
410-51333-4	7670061 005 Conley Well FB	Total/NA	Potable Water	EPA 537 Ver 1.1	
410-51333-5	7670061 301 Conley Between Lead & Lag	Total/NA	Drinking Water	EPA 537 Ver 1.1	
410-51333-6	7670061 301 Conley Between Lead & Lag FB	Total/NA	Potable Water	EPA 537 Ver 1.1	
410-51333-7	7670061 301 Conley After Lag Vessel	Total/NA	Drinking Water	EPA 537 Ver 1.1	
410-51333-8	7670061 301 Conley After Lag Vessel FB	Total/NA	Potable Water	EPA 537 Ver 1.1	
410-51333-9	7670061 101 Conley EP Grab Water	Total/NA	Drinking Water	EPA 537 Ver 1.1	
410-51333-10	7670061 101 Conley Field Blank Grab Water	Total/NA	Potable Water	EPA 537 Ver 1.1	
MB 410-161150/1-A	Method Blank	Total/NA	Drinking Water	EPA 537 Ver 1.1	
LCS 410-161150/2-A	Lab Control Sample	Total/NA	Drinking Water	EPA 537 Ver 1.1	
LCSD 410-161150/3-A	Lab Control Sample Dup	Total/NA	Drinking Water	EPA 537 Ver 1.1	
LLCS 410-161150/4-A	Lab Control Sample	Total/NA	Drinking Water	EPA 537 Ver 1.1	

Prep Batch: 162148

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
410-51333-11	7670061 002 Coppersmith Well	Total/NA	Drinking Water	EPA 537 Ver 1.1	
410-51333-11 - DL	7670061 002 Coppersmith Well	Total/NA	Drinking Water	EPA 537 Ver 1.1	
410-51333-12	7670061 002 Coppersmith Well FB	Total/NA	Potable Water	EPA 537 Ver 1.1	
410-51333-13	7670061 003 DuPont Well	Total/NA	Drinking Water	EPA 537 Ver 1.1	
410-51333-14	7670061 003 DuPont Well FB	Total/NA	Potable Water	EPA 537 Ver 1.1	
410-51333-15	7670061 302 DuPont Between Lead & Lag	Total/NA	Drinking Water	EPA 537 Ver 1.1	
410-51333-16	7670061 302 DuPont Between Lead & Lag FB	Total/NA	Potable Water	EPA 537 Ver 1.1	
410-51333-17	7670061 302 DuPont After Lag Vessel	Total/NA	Drinking Water	EPA 537 Ver 1.1	
410-51333-18	7670061 302 DuPont After Lag Vessel FB	Total/NA	Potable Water	EPA 537 Ver 1.1	
410-51333-19	7670061 102 DuPont EP	Total/NA	Drinking Water	EPA 537 Ver 1.1	
410-51333-20	7670061 102 DuPont EP FB	Total/NA	Potable Water	EPA 537 Ver 1.1	
MB 410-162148/1-A	Method Blank	Total/NA	Drinking Water	EPA 537 Ver 1.1	
LCS 410-162148/2-A	Lab Control Sample	Total/NA	Drinking Water	EPA 537 Ver 1.1	
LCSD 410-162148/3-A	Lab Control Sample Dup	Total/NA	Drinking Water	EPA 537 Ver 1.1	
LLCS 410-162148/4-A	Lab Control Sample	Total/NA	Drinking Water	EPA 537 Ver 1.1	

Analysis Batch: 163138

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
410-51333-1	7670061 001 Playground Well	Total/NA	Drinking Water	EPA 537 Ver 1.1	161150
410-51333-2	7670061 001 Playground Well FB	Total/NA	Potable Water	EPA 537 Ver 1.1	161150
410-51333-3	7670061 005 Conley Well	Total/NA	Drinking Water	EPA 537 Ver 1.1	161150
410-51333-4	7670061 005 Conley Well FB	Total/NA	Potable Water	EPA 537 Ver 1.1	161150
410-51333-5	7670061 301 Conley Between Lead & Lag	Total/NA	Drinking Water	EPA 537 Ver 1.1	161150
410-51333-6	7670061 301 Conley Between Lead & Lag FB	Total/NA	Potable Water	EPA 537 Ver 1.1	161150
410-51333-7	7670061 301 Conley After Lag Vessel	Total/NA	Drinking Water	EPA 537 Ver 1.1	161150
410-51333-8	7670061 301 Conley After Lag Vessel FB	Total/NA	Potable Water	EPA 537 Ver 1.1	161150
410-51333-9	7670061 101 Conley EP Grab Water	Total/NA	Drinking Water	EPA 537 Ver 1.1	161150
410-51333-10	7670061 101 Conley Field Blank Grab Water	Total/NA	Potable Water	EPA 537 Ver 1.1	161150
410-51333-11	7670061 002 Coppersmith Well	Total/NA	Drinking Water	EPA 537 Ver 1.1	162148
410-51333-12	7670061 002 Coppersmith Well FB	Total/NA	Potable Water	EPA 537 Ver 1.1	162148
410-51333-13	7670061 003 DuPont Well	Total/NA	Drinking Water	EPA 537 Ver 1.1	162148
410-51333-14	7670061 003 DuPont Well FB	Total/NA	Potable Water	EPA 537 Ver 1.1	162148
410-51333-15	7670061 302 DuPont Between Lead & Lag	Total/NA	Drinking Water	EPA 537 Ver 1.1	162148
410-51333-16	7670061 302 DuPont Between Lead & Lag FB	Total/NA	Potable Water	EPA 537 Ver 1.1	162148

QC Association Summary

Client: SUEZ Water Environmental Services Inc
 Project/Site: Newberry System

Job ID: 410-51333-1

LCMS (Continued)

Analysis Batch: 163138 (Continued)

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
410-51333-17	7670061 302 DuPont After Lag Vessel	Total/NA	Drinking Water	EPA 537 Ver 1.1	162148
410-51333-18	7670061 302 DuPont After Lag Vessel FB	Total/NA	Potable Water	EPA 537 Ver 1.1	162148
410-51333-19	7670061 102 DuPont EP	Total/NA	Drinking Water	EPA 537 Ver 1.1	162148
410-51333-20	7670061 102 DuPont EP FB	Total/NA	Potable Water	EPA 537 Ver 1.1	162148
MB 410-161150/1-A	Method Blank	Total/NA	Drinking Water	EPA 537 Ver 1.1	161150
MB 410-162148/1-A	Method Blank	Total/NA	Drinking Water	EPA 537 Ver 1.1	162148
LCS 410-161150/2-A	Lab Control Sample	Total/NA	Drinking Water	EPA 537 Ver 1.1	161150
LCS 410-162148/2-A	Lab Control Sample	Total/NA	Drinking Water	EPA 537 Ver 1.1	162148
LCSD 410-161150/3-A	Lab Control Sample Dup	Total/NA	Drinking Water	EPA 537 Ver 1.1	161150
LCSD 410-162148/3-A	Lab Control Sample Dup	Total/NA	Drinking Water	EPA 537 Ver 1.1	162148
LLCS 410-161150/4-A	Lab Control Sample	Total/NA	Drinking Water	EPA 537 Ver 1.1	161150
LLCS 410-162148/4-A	Lab Control Sample	Total/NA	Drinking Water	EPA 537 Ver 1.1	162148

Analysis Batch: 164020

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
410-51333-11 - DL	7670061 002 Coppersmith Well	Total/NA	Drinking Water	EPA 537 Ver 1.1	162148

Lab Chronicle

Client: SUEZ Water Environmental Services Inc
 Project/Site: Newberry System

Job ID: 410-51333-1

Client Sample ID: 7670061 001 Playground Well

Lab Sample ID: 410-51333-1

Date Collected: 08/12/21 09:55

Matrix: Drinking Water

Date Received: 08/13/21 17:28

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	EPA 537 Ver 1.1			161150	08/17/21 07:04	RDL8	ELLE
Total/NA	Analysis	EPA 537 Ver 1.1		1	163138	08/21/21 18:38	DCS9	ELLE

Client Sample ID: 7670061 001 Playground Well FB

Lab Sample ID: 410-51333-2

Date Collected: 08/12/21 09:55

Matrix: Potable Water

Date Received: 08/13/21 17:28

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	EPA 537 Ver 1.1			161150	08/17/21 07:04	RDL8	ELLE
Total/NA	Analysis	EPA 537 Ver 1.1		1	163138	08/21/21 18:50	DCS9	ELLE

Client Sample ID: 7670061 005 Conley Well

Lab Sample ID: 410-51333-3

Date Collected: 08/12/21 10:00

Matrix: Drinking Water

Date Received: 08/13/21 17:28

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	EPA 537 Ver 1.1			161150	08/17/21 07:04	RDL8	ELLE
Total/NA	Analysis	EPA 537 Ver 1.1		1	163138	08/21/21 19:02	DCS9	ELLE

Client Sample ID: 7670061 005 Conley Well FB

Lab Sample ID: 410-51333-4

Date Collected: 08/12/21 10:00

Matrix: Potable Water

Date Received: 08/13/21 17:28

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	EPA 537 Ver 1.1			161150	08/17/21 07:04	RDL8	ELLE
Total/NA	Analysis	EPA 537 Ver 1.1		1	163138	08/21/21 19:13	DCS9	ELLE

Client Sample ID: 7670061 301 Conley Between Lead & Lag

Lab Sample ID: 410-51333-5

Date Collected: 08/12/21 09:50

Matrix: Drinking Water

Date Received: 08/13/21 17:28

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	EPA 537 Ver 1.1			161150	08/17/21 07:04	RDL8	ELLE
Total/NA	Analysis	EPA 537 Ver 1.1		1	163138	08/21/21 19:25	DCS9	ELLE

Client Sample ID: 7670061 301 Conley Between Lead & Lag

Lab Sample ID: 410-51333-6

FB

Date Collected: 08/12/21 09:50

Matrix: Potable Water

Date Received: 08/13/21 17:28

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	EPA 537 Ver 1.1			161150	08/17/21 07:04	RDL8	ELLE
Total/NA	Analysis	EPA 537 Ver 1.1		1	163138	08/21/21 19:36	DCS9	ELLE

Lab Chronicle

Client: SUEZ Water Environmental Services Inc
 Project/Site: Newberry System

Job ID: 410-51333-1

Client Sample ID: 7670061 301 Conley After Lag Vessel

Lab Sample ID: 410-51333-7

Matrix: Drinking Water

Date Collected: 08/12/21 09:45
 Date Received: 08/13/21 17:28

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	EPA 537 Ver 1.1			161150	08/17/21 07:04	RDL8	ELLE
Total/NA	Analysis	EPA 537 Ver 1.1		1	163138	08/21/21 19:48	DCS9	ELLE

Client Sample ID: 7670061 301 Conley After Lag Vessel FB

Lab Sample ID: 410-51333-8

Matrix: Potable Water

Date Collected: 08/12/21 09:45
 Date Received: 08/13/21 17:28

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	EPA 537 Ver 1.1			161150	08/17/21 07:04	RDL8	ELLE
Total/NA	Analysis	EPA 537 Ver 1.1		1	163138	08/21/21 19:59	DCS9	ELLE

Client Sample ID: 7670061 101 Conley EP Grab Water

Lab Sample ID: 410-51333-9

Matrix: Drinking Water

Date Collected: 08/12/21 09:40
 Date Received: 08/13/21 17:28

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	EPA 537 Ver 1.1			161150	08/17/21 07:04	RDL8	ELLE
Total/NA	Analysis	EPA 537 Ver 1.1		1	163138	08/21/21 20:11	DCS9	ELLE

Client Sample ID: 7670061 101 Conley Field Blank Grab Water

Lab Sample ID: 410-51333-10

Matrix: Potable Water

Date Collected: 08/12/21 09:40
 Date Received: 08/13/21 17:28

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	EPA 537 Ver 1.1			161150	08/17/21 07:04	RDL8	ELLE
Total/NA	Analysis	EPA 537 Ver 1.1		1	163138	08/21/21 20:23	DCS9	ELLE

Client Sample ID: 7670061 002 Coppersmith Well

Lab Sample ID: 410-51333-11

Matrix: Drinking Water

Date Collected: 08/12/21 08:50
 Date Received: 08/13/21 17:28

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	EPA 537 Ver 1.1			162148	08/19/21 05:42	GK2L	ELLE
Total/NA	Analysis	EPA 537 Ver 1.1		1	163138	08/21/21 21:32	DCS9	ELLE
Total/NA	Prep	EPA 537 Ver 1.1	DL		162148	08/19/21 05:42	GK2L	ELLE
Total/NA	Analysis	EPA 537 Ver 1.1	DL	10	164020	08/24/21 15:17	Y6ZN	ELLE

Client Sample ID: 7670061 002 Coppersmith Well FB

Lab Sample ID: 410-51333-12

Matrix: Potable Water

Date Collected: 08/12/21 08:50
 Date Received: 08/13/21 17:28

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	EPA 537 Ver 1.1			162148	08/19/21 05:42	GK2L	ELLE
Total/NA	Analysis	EPA 537 Ver 1.1		1	163138	08/21/21 21:44	DCS9	ELLE

Lab Chronicle

Client: SUEZ Water Environmental Services Inc
 Project/Site: Newberry System

Job ID: 410-51333-1

Client Sample ID: 7670061 003 DuPont Well

Date Collected: 08/12/21 09:20

Date Received: 08/13/21 17:28

Lab Sample ID: 410-51333-13

Matrix: Drinking Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	EPA 537 Ver 1.1			162148	08/19/21 05:42	GK2L	ELLE
Total/NA	Analysis	EPA 537 Ver 1.1		1	163138	08/21/21 21:55	DCS9	ELLE

Client Sample ID: 7670061 003 DuPont Well FB

Date Collected: 08/12/21 09:20

Date Received: 08/13/21 17:28

Lab Sample ID: 410-51333-14

Matrix: Potable Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	EPA 537 Ver 1.1			162148	08/19/21 05:42	GK2L	ELLE
Total/NA	Analysis	EPA 537 Ver 1.1		1	163138	08/21/21 22:07	DCS9	ELLE

Client Sample ID: 7670061 302 DuPont Between Lead & Lag

Date Collected: 08/12/21 09:15

Date Received: 08/13/21 17:28

Lab Sample ID: 410-51333-15

Matrix: Drinking Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	EPA 537 Ver 1.1			162148	08/19/21 05:42	GK2L	ELLE
Total/NA	Analysis	EPA 537 Ver 1.1		1	163138	08/21/21 22:18	DCS9	ELLE

Client Sample ID: 7670061 302 DuPont Between Lead & Lag

FB

Date Collected: 08/12/21 09:15

Date Received: 08/13/21 17:28

Lab Sample ID: 410-51333-16

Matrix: Potable Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	EPA 537 Ver 1.1			162148	08/19/21 05:42	GK2L	ELLE
Total/NA	Analysis	EPA 537 Ver 1.1		1	163138	08/21/21 22:30	DCS9	ELLE

Client Sample ID: 7670061 302 DuPont After Lag Vessel

Date Collected: 08/12/21 09:10

Date Received: 08/13/21 17:28

Lab Sample ID: 410-51333-17

Matrix: Drinking Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	EPA 537 Ver 1.1			162148	08/19/21 05:42	GK2L	ELLE
Total/NA	Analysis	EPA 537 Ver 1.1		1	163138	08/21/21 22:42	DCS9	ELLE

Client Sample ID: 7670061 302 DuPont After Lag Vessel FB

Date Collected: 08/12/21 09:10

Date Received: 08/13/21 17:28

Lab Sample ID: 410-51333-18

Matrix: Potable Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	EPA 537 Ver 1.1			162148	08/19/21 05:42	GK2L	ELLE
Total/NA	Analysis	EPA 537 Ver 1.1		1	163138	08/21/21 22:53	DCS9	ELLE

Lab Chronicle

Client: SUEZ Water Environmental Services Inc
Project/Site: Newberry System

Job ID: 410-51333-1

Client Sample ID: 7670061 102 DuPont EP

Lab Sample ID: 410-51333-19

Matrix: Drinking Water

Date Collected: 08/12/21 09:05
Date Received: 08/13/21 17:28

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	EPA 537 Ver 1.1			162148	08/19/21 05:42	GK2L	ELLE
Total/NA	Analysis	EPA 537 Ver 1.1		1	163138	08/21/21 23:05	DCS9	ELLE

Client Sample ID: 7670061 102 DuPont EP FB

Lab Sample ID: 410-51333-20

Matrix: Potable Water

Date Collected: 08/12/21 09:05
Date Received: 08/13/21 17:28

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	EPA 537 Ver 1.1			162148	08/19/21 05:42	GK2L	ELLE
Total/NA	Analysis	EPA 537 Ver 1.1		1	163138	08/21/21 23:16	DCS9	ELLE

Laboratory References:

ELLE = Eurofins Lancaster Laboratories Env, LLC, 2425 New Holland Pike, Lancaster, PA 17601, TEL (717)656-2300

Accreditation/Certification Summary

Client: SUEZ Water Environmental Services Inc
Project/Site: Newberry System

Job ID: 410-51333-1

Laboratory: Eurofins Lancaster Laboratories Env, LLC

The accreditations/certifications listed below are applicable to this report.

Authority	Program	Identification Number	Expiration Date
Pennsylvania	NEILAP	36-00037	01-31-22

1

2

3

4

5

6

7

8

9

10

11

12

13

14

15

Method Summary

Client: SUEZ Water Environmental Services Inc
Project/Site: Newberry System

Job ID: 410-51333-1

Method	Method Description	Protocol	Laboratory
EPA 537 Ver 1.1	EPA 537 Version 1.1	EPA	ELLE
EPA 537 Ver 1.1	EPA 537 Version 1.1	EPA	ELLE

Protocol References:

EPA = US Environmental Protection Agency

Laboratory References:

ELLE = Eurofins Lancaster Laboratories Env, LLC, 2425 New Holland Pike, Lancaster, PA 17601, TEL (717)656-2300

1

2

3

4

5

6

7

8

9

10

11

12

13

14

15

Sample Summary

Client: SUEZ Water Environmental Services Inc
 Project/Site: Newberry System

Job ID: 410-51333-1

Lab Sample ID	Client Sample ID	Matrix	Collected	Received	
410-51333-1	7670061 001 Playground Well	Drinking Water	08/12/21 09:55	08/13/21 17:28	1
410-51333-2	7670061 001 Playground Well FB	Potable Water	08/12/21 09:55	08/13/21 17:28	2
410-51333-3	7670061 005 Conley Well	Drinking Water	08/12/21 10:00	08/13/21 17:28	3
410-51333-4	7670061 005 Conley Well FB	Potable Water	08/12/21 10:00	08/13/21 17:28	4
410-51333-5	7670061 301 Conley Between Lead & Lag	Drinking Water	08/12/21 09:50	08/13/21 17:28	5
410-51333-6	7670061 301 Conley Between Lead & Lag FB	Potable Water	08/12/21 09:50	08/13/21 17:28	6
410-51333-7	7670061 301 Conley After Lag Vessel	Drinking Water	08/12/21 09:45	08/13/21 17:28	7
410-51333-8	7670061 301 Conley After Lag Vessel FB	Potable Water	08/12/21 09:45	08/13/21 17:28	8
410-51333-9	7670061 101 Conley EP Grab Water	Drinking Water	08/12/21 09:40	08/13/21 17:28	9
410-51333-10	7670061 101 Conley Field Blank Grab Water	Potable Water	08/12/21 09:40	08/13/21 17:28	10
410-51333-11	7670061 002 Coppersmith Well	Drinking Water	08/12/21 08:50	08/13/21 17:28	11
410-51333-12	7670061 002 Coppersmith Well FB	Potable Water	08/12/21 08:50	08/13/21 17:28	12
410-51333-13	7670061 003 DuPont Well	Drinking Water	08/12/21 09:20	08/13/21 17:28	13
410-51333-14	7670061 003 DuPont Well FB	Potable Water	08/12/21 09:20	08/13/21 17:28	14
410-51333-15	7670061 302 DuPont Between Lead & Lag	Drinking Water	08/12/21 09:15	08/13/21 17:28	15
410-51333-16	7670061 302 DuPont Between Lead & Lag FB	Potable Water	08/12/21 09:15	08/13/21 17:28	
410-51333-17	7670061 302 DuPont After Lag Vessel	Drinking Water	08/12/21 09:10	08/13/21 17:28	
410-51333-18	7670061 302 DuPont After Lag Vessel FB	Potable Water	08/12/21 09:10	08/13/21 17:28	
410-51333-19	7670061 102 DuPont EP	Drinking Water	08/12/21 09:05	08/13/21 17:28	
410-51333-20	7670061 102 DuPont EP FB	Potable Water	08/12/21 09:05	08/13/21 17:28	

En



410-51333 Chain of Custody

eurofins

Lancaster Laboratories
Environmental

'sis Request/Chain of Custody

Client: SUEZ WATER PA		Site ID #:		Matrix		Analyses Requested						For Lab Use Only	
Project Name: Newberry System				<input type="checkbox"/> Soil	<input type="checkbox"/> Sediment	<input checked="" type="checkbox"/> Tissue	<input type="checkbox"/> Ground	<input type="checkbox"/> Surface	<input type="checkbox"/> O	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	SF #:
Project Manager: Elizabeth Zanar		P.O. #:		<input type="checkbox"/> Water	<input type="checkbox"/> NPDES	<input type="checkbox"/> Other: GAC Filtered Water	<input type="checkbox"/> Potable	<input type="checkbox"/> NPDES	<input type="checkbox"/> Other: GAC Filtered Water	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	SCR #:
Sampler: Penny Bumbarger		PWSID #: 7670061		<input type="checkbox"/> Total # of Containers									Preservation Codes
Phone #: 717-773-0185		Quote #: 219948A											H = HCl T = Thiosulfate
State where samples were collected: PA		For Compliance: Yes <input checked="" type="checkbox"/> No <input type="checkbox"/>											N = HNO ₃ B = NaOH
Sample Identification		Collection											S = H ₂ SO ₄ P = H ₃ PO ₄
		Date	Time	Grab	Composite								
001 Playground Well	8/14/21	0955	X			X							Remarks
FB - Playground Well	8/12/21	0955											Monthly Compliance
005 Conley Well	8/14/21	1000	X			X							
FB - Conley Well	8/12/21	1000											
301s Conley Between Lead and Lag	8/12/21	0950	X				X	2	X				
FB - Conley Between Lead and Lag	8/12/21	0950						2	X				
301s Conley After Lag	8/12/21	0945	X				X	2	X				
FB - Conely After Lag	8/12/21	0945						2	X				
EP 101 Conley	8/12/21	0940	X				X	2	X				
FB - EP 101 Conley	8/12/21	0940						2	X				
Turnaround Time Requested (TAT) (please check): Standard <input checked="" type="checkbox"/> Rush <input type="checkbox"/>				Relinquished by:		Date	Time	Received by:	Date	Time			
(Rush TAT is subject to laboratory approval and surcharges.)				<i>Penny Bumbarger</i>		8/13/21	1129	<i>Elizabeth Zanar</i>	8/13/21	1129			
Date results are needed:				Relinquished by:		Date	Time	Received by:	Date	Time			
Rush results requested by (please check): E-Mail <input checked="" type="checkbox"/> Phone <input type="checkbox"/>				<i>Jane Doe</i>		8/13/21	1719	<i>Elizabeth Zanar</i>					
E-mail Address: penny.bumbarger@suez.com				Relinquished by:		Date	Time	Received by:	Date	Time			
Phone: 717-773-0185				Relinquished by:		Date	Time	Received by:	Date	Time			
Data Package Options (please check if required)				Relinquished by:		Date	Time	Received by:	Date	Time			
Type I (Validation/non-CLP)	<input type="checkbox"/>	MA MCP	<input type="checkbox"/>	Relinquished by:		Date	Time	Received by:	Date	Time			
Type III (Reduced non-CLP)	<input type="checkbox"/>	CT RCP	<input type="checkbox"/>	Relinquished by:		Date	Time	Received by:	Date	Time			
Type VI (Raw Data Only)	<input type="checkbox"/>	TX TRRP-13	<input type="checkbox"/>	Relinquished by:		Date	Time	Received by:	Date	Time			
NJ DKQP	<input type="checkbox"/>	NYSDEC Category	<input type="checkbox"/> A or <input type="checkbox"/> B	Relinquished by Commercial Carrier:		Temperature upon receipt <i>0.1-1.8</i> °C							
EDD Required?	Yes <input type="checkbox"/>	No <input type="checkbox"/>	If yes, format:	UPS	FedEx	Other							

Environmental Analysis Request/Chain of Custody



Lancaster Laboratories
Environmental

Acct. # 44297 Group #

Sample #

Client: SUEZ WATER PA				Analyses Requested								For Lab Use Only					
Project Name: Newberry System		Site ID #:		Matrix				Preservation and Filtration Codes								SF #: _____	
Project Manager: Elizabeth Zanar		P.O. #:						<input type="checkbox"/>	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	O					
Sampler: Penny Bumbarger		PWSID #: 7670061														Preservation Codes	
Phone #: 717-773-0185		Quote #: 219948A														H = HCl	T = Thiosulfate
State where samples were collected: PA		For Compliance: Yes <input checked="" type="checkbox"/> No <input type="checkbox"/>														N = HNO ₃	B = NaOH
																S = H ₂ SO ₄	P = H ₃ PO ₄
																F = Field Filtered	O = Other
																Remarks	
																Monthly Compliance	
Sample Identification		Collection		Grab	Composite	Total # of Containers	PFAS (14) 537 v 1.1										
		Date	Time	Soil	Sediment	Tissue	Water	Potable	Ground	NPDES	Surface	GAC Filtered Water					
002 Coppersmith Well		8/12/21	0850	X			X										
FB - Coppersmith Well		8/12/21	0850														
003 DuPont Well		8/12/21	0920	X			X										
FB - DuPont Well		8/12/21	0920														
302s DuPont Between Lead and Lag		8/12/21	0915	X				X									
FB - DuPont Between Lead and Lag		8/12/21	0915														
302s DuPont After Lag		8/12/21	0910	X				X									
FB - DuPont After Lag		8/12/21	0910														
EP 102 DuPont		8/12/21	0905	X				X									
FB - EP 102 DuPont		8/12/21	0905														
Turnaround Time Requested (TAT) (please check): Standard <input checked="" type="checkbox"/> Rush <input type="checkbox"/> (Rush TAT is subject to laboratory approval and surcharges.)				Relinquished by:		Date	Time	Received by:	Date		Time						
				<i>Penny Bumbarger</i>		8/13/21	1129	<i>Carla Deek</i>	8/13/21		1129						
Date results are needed:				Relinquished by:		Date	Time	Received by:	Date		Time						
				<i>Carla Deek</i>		8/13/21	1719										
Rush results requested by (please check): E-Mail <input checked="" type="checkbox"/> Phone <input type="checkbox"/>				Relinquished by:		Date	Time	Received by:	Date		Time						
E-mail Address: penny.bumbarger@suez.com																	
Phone: 717-773-0185				Relinquished by:		Date	Time	Received by:	Date		Time						
Data Package Options (please check if required)				Relinquished by:		Date	Time	Received by:	Date		Time						
Type I (Validation/non-CLP) <input type="checkbox"/> MA MCP <input type="checkbox"/>																	
Type III (Reduced non-CLP) <input type="checkbox"/> CT RCP <input type="checkbox"/>				Relinquished by:		Date	Time	Received by:	Date		Time						
Type VI (Raw Data Only) <input type="checkbox"/> TX TRRP-13 <input type="checkbox"/>																	
NJ DKQP <input type="checkbox"/> NYSDEC Category <input type="checkbox"/> A or <input type="checkbox"/> B				Relinquished by Commercial Carrier:													
EDD Required? Yes <input type="checkbox"/> No <input type="checkbox"/> If yes, format:				UPS	FedEx	Other			Temperature upon receipt		0.1-18 °C						

Login Sample Receipt Checklist

Client: SUEZ Water Environmental Services Inc

Job Number: 410-51333-1

Login Number: 51333

List Source: Eurofins Lancaster Laboratories Env, LLC

List Number: 1

Creator: Cyms, Carolyn M

Question	Answer	Comment
Radioactivity wasn't checked or is </= background as measured by a survey meter.	N/A	
The cooler's custody seal is intact.	True	
The cooler or samples do not appear to have been compromised or tampered with.	True	
Samples were received on ice.	True	
Cooler Temperature is acceptable (</=6C, not frozen).	True	
Cooler Temperature is recorded.	True	
WV: Container Temperature is acceptable (</=6C, not frozen).	N/A	
WV: Container Temperature is recorded.	N/A	
COC is present.	True	
COC is filled out in ink and legible.	True	
COC is filled out with all pertinent information.	True	
There are no discrepancies between the containers received and the COC.	True	
Samples are received within Holding Time (excluding tests with immediate HTs)	True	
Sample containers have legible labels.	True	
Containers are not broken or leaking.	True	
Sample collection date/times are provided.	True	
Appropriate sample containers are used.	True	
Sample bottles are completely filled.	True	
There is sufficient vol. for all requested analyses.	True	
Multiphasic samples are not present.	True	
Samples do not require splitting or compositing.	N/A	
Is the Field Sampler's name present on COC?	True	
Sample Preservation Verified.	N/A	
Residual Chlorine Checked.	N/A	
Sample custody seals are intact.	N/A	