


ANALYTICAL REPORT

Eurofins Lancaster Laboratories Env, LLC
2425 New Holland Pike
Lancaster, PA 17601
Tel: (717)656-2300

Laboratory Job ID: 410-67522-1
Client Project/Site: Newberry System
Sampling Event: Newberry System

For:
SUEZ Water Environmental Services Inc
6310 Allentown Blvd
Suite 104
Harrisburg, Pennsylvania 17112

Attn: Penny Bumbarger



Authorized for release by:
1/4/2022 8:43:00 AM

Elizabeth Zanar, Project Manager
(717)556-7290
Elizabeth.Zanar@eurofinset.com

LINKS

Review your project
results through
TotalAccess

Have a Question?



Visit us at:

www.eurofinsus.com/Env

The test results in this report meet all 2003 NELAC, 2009 TNI, and 2016 TNI requirements for accredited parameters, exceptions are noted in this report. This report may not be reproduced except in full, and with written approval from the laboratory. For questions please contact the Project Manager at the e-mail address or telephone number listed on this page.

This report has been electronically signed and authorized by the signatory. Electronic signature is intended to be the legally binding equivalent of a traditionally handwritten signature.

Results relate only to the items tested and the sample(s) as received by the laboratory.



Analytical test results meet all requirements of the associated regulatory program (e.g., NELAC (TNI), DoD, and ISO 17025) unless otherwise noted under the individual analysis. Data qualifiers are applied to note exceptions. Noncompliant quality control (QC) is further explained in narrative comments.

- QC results that exceed the upper limits and are associated with non-detect samples are qualified but further narration is not required since the bias is high and does not change a non-detect result. Further narration is also not required with QC blank detection when the associated sample concentration is non-detect or more than ten times the level in the blank.
 - Matrix QC may not be reported if insufficient sample or site-specific QC samples were not submitted. In these situations, to demonstrate precision and accuracy at a batch level, a LCS/LCSD is performed, unless otherwise specified in the method.
 - Surrogate and/or isotope dilution analyte recoveries (if applicable) which are outside of the QC window are confirmed unless attributed to a dilution or otherwise noted in the narrative.
- Regulated compliance samples (e.g. SDWA, NPDES) must comply with the associated agency requirements/permits.

Measurement uncertainty values, as applicable, are available upon request.

Test results relate only to the sample tested. Clients should be aware that a critical step in a chemical or microbiological analysis is the collection of the sample. Unless the sample analyzed is truly representative of the bulk of material involved, the test results will be meaningless. If you have questions regarding the proper techniques of collecting samples, please contact us. We cannot be held responsible for sample integrity, however, unless sampling has been performed by a member of our staff. Times are local to the area of activity. Parameters listed in the 40 CFR Part 136 Table II as "analyze immediately" and tested in the laboratory are not performed within 15 minutes of collection.

This report shall not be reproduced except in full, without the written approval of the laboratory.

WARRANTY AND LIMITS OF LIABILITY - In accepting analytical work, we warrant the accuracy of test results for the sample as submitted. The foregoing express warranty is exclusive and is given in lieu of all other warranties, expressed or implied, except as otherwise agreed. We disclaim any other warranties, expressed or implied, including a warranty of fitness for particular purpose and warranty of merchantability. In no event shall Eurofins Lancaster Laboratories Environmental, LLC be liable for indirect, special, consequential, or incidental damages including, but not limited to, damages for loss of profit or goodwill regardless of (A) the negligence (either sole or concurrent) of Eurofins Lancaster Laboratories Environmental and (B) whether Eurofins Lancaster Laboratories Environmental has been informed of the possibility of such damages. We accept no legal responsibility for the purposes for which the client uses the test results. Except as otherwise agreed, no purchase order or other order for work shall be accepted by Eurofins Lancaster Laboratories Environmental which includes any conditions that vary from the Standard Terms and Conditions, and Eurofins Lancaster Laboratories Environmental hereby objects to any conflicting terms contained in any acceptance or order submitted by client.

Elizabeth Zanar
Project Manager
1/4/2022 8:43:00 AM



Table of Contents

Cover Page	1
Table of Contents	3
Definitions/Glossary	4
Case Narrative	5
Detection Summary	6
Client Sample Results	9
Surrogate Summary	18
QC Sample Results	19
QC Association Summary	21
Lab Chronicle	22
Certification Summary	25
Method Summary	26
Sample Summary	27
Chain of Custody	28
Receipt Checklists	30

Definitions/Glossary

Client: SUEZ Water Environmental Services Inc
Project/Site: Newberry System

Job ID: 410-67522-1

Qualifiers

LCMS

Qualifier	Qualifier Description
J	Result is less than the RL but greater than or equal to the MDL and the concentration is an approximate value.

Glossary

Abbreviation	These commonly used abbreviations may or may not be present in this report.
α	Listed under the "D" column to designate that the result is reported on a dry weight basis
%R	Percent Recovery
1C	Result is from the primary column on a dual-column method.
2C	Result is from the confirmation column on a dual-column method.
CFL	Contains Free Liquid
CFU	Colony Forming Unit
CNF	Contains No Free Liquid
DER	Duplicate Error Ratio (normalized absolute difference)
Dil Fac	Dilution Factor
DL	Detection Limit (DoD/DOE)
DL, RA, RE, IN	Indicates a Dilution, Re-analysis, Re-extraction, or additional Initial metals/anion analysis of the sample
DLC	Decision Level Concentration (Radiochemistry)
EDL	Estimated Detection Limit (Dioxin)
LOD	Limit of Detection (DoD/DOE)
LOQ	Limit of Quantitation (DoD/DOE)
MCL	EPA recommended "Maximum Contaminant Level"
MDA	Minimum Detectable Activity (Radiochemistry)
MDC	Minimum Detectable Concentration (Radiochemistry)
MDL	Method Detection Limit
ML	Minimum Level (Dioxin)
MPN	Most Probable Number
MQL	Method Quantitation Limit
NC	Not Calculated
ND	Not Detected at the reporting limit (or MDL or EDL if shown)
NEG	Negative / Absent
POS	Positive / Present
PQL	Practical Quantitation Limit
PRES	Presumptive
QC	Quality Control
RER	Relative Error Ratio (Radiochemistry)
RL	Reporting Limit or Requested Limit (Radiochemistry)
RPD	Relative Percent Difference, a measure of the relative difference between two points
TEF	Toxicity Equivalent Factor (Dioxin)
TEQ	Toxicity Equivalent Quotient (Dioxin)
TNTC	Too Numerous To Count

Case Narrative

Client: SUEZ Water Environmental Services Inc
Project/Site: Newberry System

Job ID: 410-67522-1

Job ID: 410-67522-1

Laboratory: Eurofins Lancaster Laboratories Env, LLC

Narrative

Job Narrative
410-67522-1

Receipt

The samples were received on 12/17/2021 3:20 PM. Unless otherwise noted below, the samples arrived in good condition, and, where required, properly preserved and on ice. The temperature of the cooler at receipt time was 1.3°C

PFAS

No additional analytical or quality issues were noted, other than those described above or in the Definitions/ Glossary page.

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15

Detection Summary

Client: SUEZ Water Environmental Services Inc
Project/Site: Newberry System

Job ID: 410-67522-1

Client Sample ID: 7670061 001 Playground Well

Lab Sample ID: 410-67522-1

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Perfluorohexanoic acid	8.9		1.7	0.43	ng/L	1		EPA 537 Ver 1.1	Total/NA
Perfluoroheptanoic acid	2.6		1.7	0.43	ng/L	1		EPA 537 Ver 1.1	Total/NA
Perfluorooctanoic acid	6.4		1.7	0.43	ng/L	1		EPA 537 Ver 1.1	Total/NA
Perfluorobutanesulfonic acid	3.8		1.7	0.43	ng/L	1		EPA 537 Ver 1.1	Total/NA
Perfluorohexanesulfonic acid	4.8		1.7	0.43	ng/L	1		EPA 537 Ver 1.1	Total/NA
Perfluorooctanesulfonic acid	9.8		1.7	0.43	ng/L	1		EPA 537 Ver 1.1	Total/NA

Client Sample ID: 7670061 001 Playground Well FB

Lab Sample ID: 410-67522-2

No Detections.

Client Sample ID: 7670061 005 Conley Well

Lab Sample ID: 410-67522-3

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Perfluorohexanoic acid	5.5		1.7	0.43	ng/L	1		EPA 537 Ver 1.1	Total/NA
Perfluorooctanoic acid	3.5		1.7	0.43	ng/L	1		EPA 537 Ver 1.1	Total/NA
Perfluorobutanesulfonic acid	2.7		1.7	0.43	ng/L	1		EPA 537 Ver 1.1	Total/NA
Perfluorohexanesulfonic acid	3.3		1.7	0.43	ng/L	1		EPA 537 Ver 1.1	Total/NA
Perfluorooctanesulfonic acid	5.5		1.7	0.43	ng/L	1		EPA 537 Ver 1.1	Total/NA

Client Sample ID: 7670061 005 Conley Well FB

Lab Sample ID: 410-67522-4

No Detections.

Client Sample ID: 7670061 301 Conley Between Lead & Lag

Lab Sample ID: 410-67522-5

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Perfluorohexanoic acid	7.6		1.7	0.42	ng/L	1		EPA 537 Ver 1.1	Total/NA
Perfluoroheptanoic acid	2.7		1.7	0.42	ng/L	1		EPA 537 Ver 1.1	Total/NA
Perfluorooctanoic acid	6.8		1.7	0.42	ng/L	1		EPA 537 Ver 1.1	Total/NA
Perfluorobutanesulfonic acid	3.8		1.7	0.42	ng/L	1		EPA 537 Ver 1.1	Total/NA
Perfluorohexanesulfonic acid	5.5		1.7	0.42	ng/L	1		EPA 537 Ver 1.1	Total/NA
Perfluorooctanesulfonic acid	9.6		1.7	0.42	ng/L	1		EPA 537 Ver 1.1	Total/NA

Client Sample ID: 7670061 301 Conley Between Lead & Lag FB

Lab Sample ID: 410-67522-6

No Detections.

Client Sample ID: 7670061 301 Conley After Lag Vessel

Lab Sample ID: 410-67522-7

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Perfluorohexanoic acid	8.4		1.7	0.42	ng/L	1		EPA 537 Ver 1.1	Total/NA
Perfluoroheptanoic acid	2.8		1.7	0.42	ng/L	1		EPA 537 Ver 1.1	Total/NA
Perfluorooctanoic acid	6.5		1.7	0.42	ng/L	1		EPA 537 Ver 1.1	Total/NA
Perfluorobutanesulfonic acid	4.3		1.7	0.42	ng/L	1		EPA 537 Ver 1.1	Total/NA
Perfluorohexanesulfonic acid	5.2		1.7	0.42	ng/L	1		EPA 537 Ver 1.1	Total/NA
Perfluorooctanesulfonic acid	6.2		1.7	0.42	ng/L	1		EPA 537 Ver 1.1	Total/NA

Client Sample ID: 7670061 301 Conley After Lag Vessel FB

Lab Sample ID: 410-67522-8

No Detections.

This Detection Summary does not include radiochemical test results.

Eurofins Lancaster Laboratories Env, LLC

Detection Summary

Client: SUEZ Water Environmental Services Inc
Project/Site: Newberry System

Job ID: 410-67522-1

Client Sample ID: 7670061 101 Conley EP Grab Water

Lab Sample ID: 410-67522-9

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Perfluorohexanoic acid	8.4		1.8	0.44	ng/L	1		EPA 537 Ver 1.1	Total/NA
Perfluoroheptanoic acid	2.8		1.8	0.44	ng/L	1		EPA 537 Ver 1.1	Total/NA
Perfluorooctanoic acid	6.5		1.8	0.44	ng/L	1		EPA 537 Ver 1.1	Total/NA
Perfluorobutanesulfonic acid	4.2		1.8	0.44	ng/L	1		EPA 537 Ver 1.1	Total/NA
Perfluorohexanesulfonic acid	5.0		1.8	0.44	ng/L	1		EPA 537 Ver 1.1	Total/NA
Perfluorooctanesulfonic acid	6.2		1.8	0.44	ng/L	1		EPA 537 Ver 1.1	Total/NA

Client Sample ID: 7670061 101 Conley Field Blank Grab Water

Lab Sample ID: 410-67522-10

No Detections.

Client Sample ID: 7670061 003 DuPont Well

Lab Sample ID: 410-67522-11

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Perfluorohexanoic acid	8.3		1.8	0.44	ng/L	1		EPA 537 Ver 1.1	Total/NA
Perfluoroheptanoic acid	3.9		1.8	0.44	ng/L	1		EPA 537 Ver 1.1	Total/NA
Perfluorooctanoic acid	6.5		1.8	0.44	ng/L	1		EPA 537 Ver 1.1	Total/NA
Perfluorononanoic acid	3.2		1.8	0.44	ng/L	1		EPA 537 Ver 1.1	Total/NA
Perfluorobutanesulfonic acid	7.9		1.8	0.44	ng/L	1		EPA 537 Ver 1.1	Total/NA
Perfluorohexanesulfonic acid	72		18	4.4	ng/L	10		EPA 537 Ver 1.1	Total/NA
Perfluorooctanesulfonic acid	74		18	4.4	ng/L	10		EPA 537 Ver 1.1	Total/NA

Client Sample ID: 7670061 003 DuPont Well FB

Lab Sample ID: 410-67522-12

No Detections.

Client Sample ID: 7670061 302 DuPont Between Lead & Lag

Lab Sample ID: 410-67522-13

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Perfluorohexanoic acid	18		1.8	0.44	ng/L	1		EPA 537 Ver 1.1	Total/NA
Perfluoroheptanoic acid	3.4		1.8	0.44	ng/L	1		EPA 537 Ver 1.1	Total/NA
Perfluorooctanoic acid	2.6		1.8	0.44	ng/L	1		EPA 537 Ver 1.1	Total/NA
Perfluorobutanesulfonic acid	8.2		1.8	0.44	ng/L	1		EPA 537 Ver 1.1	Total/NA
Perfluorohexanesulfonic acid	16		1.8	0.44	ng/L	1		EPA 537 Ver 1.1	Total/NA
Perfluorooctanesulfonic acid	4.4		1.8	0.44	ng/L	1		EPA 537 Ver 1.1	Total/NA

Client Sample ID: 7670061 302 DuPont Between Lead & Lag FB

Lab Sample ID: 410-67522-14

No Detections.

Client Sample ID: 7670061 302 DuPont After Lag Vessel

Lab Sample ID: 410-67522-15

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Perfluorohexanoic acid	12		1.8	0.44	ng/L	1		EPA 537 Ver 1.1	Total/NA
Perfluorobutanesulfonic acid	6.1		1.8	0.44	ng/L	1		EPA 537 Ver 1.1	Total/NA
Perfluorohexanesulfonic acid	1.9		1.8	0.44	ng/L	1		EPA 537 Ver 1.1	Total/NA

Client Sample ID: 7670061 302 DuPont After Lag Vessel FB

Lab Sample ID: 410-67522-16

No Detections.

Client Sample ID: 7670061 102 DuPont EP

Lab Sample ID: 410-67522-17

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Perfluorohexanoic acid	13		1.8	0.45	ng/L	1		EPA 537 Ver 1.1	Total/NA

This Detection Summary does not include radiochemical test results.

Eurofins Lancaster Laboratories Env, LLC

Detection Summary

Client: SUEZ Water Environmental Services Inc
Project/Site: Newberry System

Job ID: 410-67522-1

Client Sample ID: 7670061 102 DuPont EP (Continued)

Lab Sample ID: 410-67522-17

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Perfluorobutanesulfonic acid	6.5		1.8	0.45	ng/L	1		EPA 537 Ver 1.1	Total/NA
Perfluorohexanesulfonic acid	2.1		1.8	0.45	ng/L	1		EPA 537 Ver 1.1	Total/NA

Client Sample ID: 7670061 102 DuPont EP FB

Lab Sample ID: 410-67522-18

No Detections.

This Detection Summary does not include radiochemical test results.

Eurofins Lancaster Laboratories Env, LLC



Client Sample Results

Client: SUEZ Water Environmental Services Inc
Project/Site: Newberry System

Job ID: 410-67522-1

Client Sample ID: 7670061 001 Playground Well

Lab Sample ID: 410-67522-1

Date Collected: 12/14/21 10:40

Matrix: Drinking Water

Date Received: 12/17/21 15:20

Method: EPA 537 Ver 1.1 - EPA 537 Version 1.1

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorohexanoic acid	8.9		1.7	0.43	ng/L		12/22/21 07:45	12/28/21 14:10	1
Perfluoroheptanoic acid	2.6		1.7	0.43	ng/L		12/22/21 07:45	12/28/21 14:10	1
Perfluorooctanoic acid	6.4		1.7	0.43	ng/L		12/22/21 07:45	12/28/21 14:10	1
Perfluorononanoic acid	<1.7		1.7	0.43	ng/L		12/22/21 07:45	12/28/21 14:10	1
Perfluorodecanoic acid	<1.7		1.7	0.43	ng/L		12/22/21 07:45	12/28/21 14:10	1
Perfluorotridecanoic acid	<1.7		1.7	0.43	ng/L		12/22/21 07:45	12/28/21 14:10	1
Perfluorotetradecanoic acid	<1.7		1.7	0.43	ng/L		12/22/21 07:45	12/28/21 14:10	1
Perfluorobutanesulfonic acid	3.8		1.7	0.43	ng/L		12/22/21 07:45	12/28/21 14:10	1
Perfluorohexanesulfonic acid	4.8		1.7	0.43	ng/L		12/22/21 07:45	12/28/21 14:10	1
Perfluorooctanesulfonic acid	9.8		1.7	0.43	ng/L		12/22/21 07:45	12/28/21 14:10	1
NEtFOSAA	<1.7		1.7	0.43	ng/L		12/22/21 07:45	12/28/21 14:10	1
NMeFOSAA	<1.7		1.7	0.43	ng/L		12/22/21 07:45	12/28/21 14:10	1
Perfluoroundecanoic acid	<1.7		1.7	0.43	ng/L		12/22/21 07:45	12/28/21 14:10	1
Perfluorododecanoic acid	<1.7		1.7	0.43	ng/L		12/22/21 07:45	12/28/21 14:10	1
Surrogate	%Recovery	Qualifier	Limits				Prepared	Analyzed	Dil Fac
d5-NEtFOSAA	90		70 - 130				12/22/21 07:45	12/28/21 14:10	1
13C2 PFDA	88		70 - 130				12/22/21 07:45	12/28/21 14:10	1
13C2 PFHxA	89		70 - 130				12/22/21 07:45	12/28/21 14:10	1

Client Sample ID: 7670061 001 Playground Well FB

Lab Sample ID: 410-67522-2

Date Collected: 12/14/21 10:40

Matrix: Potable Water

Date Received: 12/17/21 15:20

Method: EPA 537 Ver 1.1 - EPA 537 Version 1.1

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorohexanoic acid	<1.8		1.8	0.46	ng/L		12/22/21 07:45	12/28/21 14:22	1
Perfluoroheptanoic acid	<1.8		1.8	0.46	ng/L		12/22/21 07:45	12/28/21 14:22	1
Perfluorooctanoic acid	<1.8		1.8	0.46	ng/L		12/22/21 07:45	12/28/21 14:22	1
Perfluorononanoic acid	<1.8		1.8	0.46	ng/L		12/22/21 07:45	12/28/21 14:22	1
Perfluorodecanoic acid	<1.8		1.8	0.46	ng/L		12/22/21 07:45	12/28/21 14:22	1
Perfluorotridecanoic acid	<1.8		1.8	0.46	ng/L		12/22/21 07:45	12/28/21 14:22	1
Perfluorotetradecanoic acid	<1.8		1.8	0.46	ng/L		12/22/21 07:45	12/28/21 14:22	1
Perfluorobutanesulfonic acid	<1.8		1.8	0.46	ng/L		12/22/21 07:45	12/28/21 14:22	1
Perfluorohexanesulfonic acid	<1.8		1.8	0.46	ng/L		12/22/21 07:45	12/28/21 14:22	1
Perfluorooctanesulfonic acid	<1.8		1.8	0.46	ng/L		12/22/21 07:45	12/28/21 14:22	1
NEtFOSAA	<1.8		1.8	0.46	ng/L		12/22/21 07:45	12/28/21 14:22	1
NMeFOSAA	<1.8		1.8	0.46	ng/L		12/22/21 07:45	12/28/21 14:22	1
Perfluoroundecanoic acid	<1.8		1.8	0.46	ng/L		12/22/21 07:45	12/28/21 14:22	1
Perfluorododecanoic acid	<1.8		1.8	0.46	ng/L		12/22/21 07:45	12/28/21 14:22	1
Surrogate	%Recovery	Qualifier	Limits				Prepared	Analyzed	Dil Fac
d5-NEtFOSAA	92		70 - 130				12/22/21 07:45	12/28/21 14:22	1
13C2 PFDA	91		70 - 130				12/22/21 07:45	12/28/21 14:22	1
13C2 PFHxA	93		70 - 130				12/22/21 07:45	12/28/21 14:22	1

Client Sample Results

Client: SUEZ Water Environmental Services Inc
Project/Site: Newberry System

Job ID: 410-67522-1

Client Sample ID: 7670061 005 Conley Well

Lab Sample ID: 410-67522-3

Date Collected: 12/14/21 10:35

Matrix: Drinking Water

Date Received: 12/17/21 15:20

Method: EPA 537 Ver 1.1 - EPA 537 Version 1.1

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorohexanoic acid	5.5		1.7	0.43	ng/L		12/22/21 07:45	12/28/21 14:34	1
Perfluoroheptanoic acid	<1.7		1.7	0.43	ng/L		12/22/21 07:45	12/28/21 14:34	1
Perfluorooctanoic acid	3.5		1.7	0.43	ng/L		12/22/21 07:45	12/28/21 14:34	1
Perfluorononanoic acid	<1.7		1.7	0.43	ng/L		12/22/21 07:45	12/28/21 14:34	1
Perfluorodecanoic acid	<1.7		1.7	0.43	ng/L		12/22/21 07:45	12/28/21 14:34	1
Perfluorotridecanoic acid	<1.7		1.7	0.43	ng/L		12/22/21 07:45	12/28/21 14:34	1
Perfluorotetradecanoic acid	<1.7		1.7	0.43	ng/L		12/22/21 07:45	12/28/21 14:34	1
Perfluorobutanesulfonic acid	2.7		1.7	0.43	ng/L		12/22/21 07:45	12/28/21 14:34	1
Perfluorohexanesulfonic acid	3.3		1.7	0.43	ng/L		12/22/21 07:45	12/28/21 14:34	1
Perfluorooctanesulfonic acid	5.5		1.7	0.43	ng/L		12/22/21 07:45	12/28/21 14:34	1
NEtFOSAA	<1.7		1.7	0.43	ng/L		12/22/21 07:45	12/28/21 14:34	1
NMeFOSAA	<1.7		1.7	0.43	ng/L		12/22/21 07:45	12/28/21 14:34	1
Perfluoroundecanoic acid	<1.7		1.7	0.43	ng/L		12/22/21 07:45	12/28/21 14:34	1
Perfluorododecanoic acid	<1.7		1.7	0.43	ng/L		12/22/21 07:45	12/28/21 14:34	1
Surrogate	%Recovery	Qualifier	Limits				Prepared	Analyzed	Dil Fac
d5-NEtFOSAA	91		70 - 130				12/22/21 07:45	12/28/21 14:34	1
13C2 PFDA	87		70 - 130				12/22/21 07:45	12/28/21 14:34	1
13C2 PFHxA	90		70 - 130				12/22/21 07:45	12/28/21 14:34	1

Client Sample ID: 7670061 005 Conley Well FB

Lab Sample ID: 410-67522-4

Date Collected: 12/14/21 10:35

Matrix: Potable Water

Date Received: 12/17/21 15:20

Method: EPA 537 Ver 1.1 - EPA 537 Version 1.1

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorohexanoic acid	<1.8		1.8	0.45	ng/L		12/22/21 07:45	12/28/21 14:46	1
Perfluoroheptanoic acid	<1.8		1.8	0.45	ng/L		12/22/21 07:45	12/28/21 14:46	1
Perfluorooctanoic acid	<1.8		1.8	0.45	ng/L		12/22/21 07:45	12/28/21 14:46	1
Perfluorononanoic acid	<1.8		1.8	0.45	ng/L		12/22/21 07:45	12/28/21 14:46	1
Perfluorodecanoic acid	<1.8		1.8	0.45	ng/L		12/22/21 07:45	12/28/21 14:46	1
Perfluorotridecanoic acid	<1.8		1.8	0.45	ng/L		12/22/21 07:45	12/28/21 14:46	1
Perfluorotetradecanoic acid	<1.8		1.8	0.45	ng/L		12/22/21 07:45	12/28/21 14:46	1
Perfluorobutanesulfonic acid	<1.8		1.8	0.45	ng/L		12/22/21 07:45	12/28/21 14:46	1
Perfluorohexanesulfonic acid	<1.8		1.8	0.45	ng/L		12/22/21 07:45	12/28/21 14:46	1
Perfluorooctanesulfonic acid	<1.8		1.8	0.45	ng/L		12/22/21 07:45	12/28/21 14:46	1
NEtFOSAA	<1.8		1.8	0.45	ng/L		12/22/21 07:45	12/28/21 14:46	1
NMeFOSAA	<1.8		1.8	0.45	ng/L		12/22/21 07:45	12/28/21 14:46	1
Perfluoroundecanoic acid	<1.8		1.8	0.45	ng/L		12/22/21 07:45	12/28/21 14:46	1
Perfluorododecanoic acid	<1.8		1.8	0.45	ng/L		12/22/21 07:45	12/28/21 14:46	1
Surrogate	%Recovery	Qualifier	Limits				Prepared	Analyzed	Dil Fac
d5-NEtFOSAA	94		70 - 130				12/22/21 07:45	12/28/21 14:46	1
13C2 PFDA	91		70 - 130				12/22/21 07:45	12/28/21 14:46	1
13C2 PFHxA	96		70 - 130				12/22/21 07:45	12/28/21 14:46	1

Client Sample Results

Client: SUEZ Water Environmental Services Inc
Project/Site: Newberry System

Job ID: 410-67522-1

Client Sample ID: 7670061 301 Conley Between Lead & Lag

Lab Sample ID: 410-67522-5

Date Collected: 12/14/21 10:30

Matrix: Drinking Water

Date Received: 12/17/21 15:20

Method: EPA 537 Ver 1.1 - EPA 537 Version 1.1

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorohexanoic acid	7.6		1.7	0.42	ng/L		12/22/21 07:45	12/28/21 14:57	1
Perfluoroheptanoic acid	2.7		1.7	0.42	ng/L		12/22/21 07:45	12/28/21 14:57	1
Perfluorooctanoic acid	6.8		1.7	0.42	ng/L		12/22/21 07:45	12/28/21 14:57	1
Perfluorononanoic acid	<1.7		1.7	0.42	ng/L		12/22/21 07:45	12/28/21 14:57	1
Perfluorodecanoic acid	<1.7		1.7	0.42	ng/L		12/22/21 07:45	12/28/21 14:57	1
Perfluorotridecanoic acid	<1.7		1.7	0.42	ng/L		12/22/21 07:45	12/28/21 14:57	1
Perfluorotetradecanoic acid	<1.7		1.7	0.42	ng/L		12/22/21 07:45	12/28/21 14:57	1
Perfluorobutanesulfonic acid	3.8		1.7	0.42	ng/L		12/22/21 07:45	12/28/21 14:57	1
Perfluorohexanesulfonic acid	5.5		1.7	0.42	ng/L		12/22/21 07:45	12/28/21 14:57	1
Perfluorooctanesulfonic acid	9.6		1.7	0.42	ng/L		12/22/21 07:45	12/28/21 14:57	1
NEtFOSAA	<1.7		1.7	0.42	ng/L		12/22/21 07:45	12/28/21 14:57	1
NMeFOSAA	<1.7		1.7	0.42	ng/L		12/22/21 07:45	12/28/21 14:57	1
Perfluoroundecanoic acid	<1.7		1.7	0.42	ng/L		12/22/21 07:45	12/28/21 14:57	1
Perfluorododecanoic acid	<1.7		1.7	0.42	ng/L		12/22/21 07:45	12/28/21 14:57	1
Surrogate	%Recovery	Qualifier	Limits				Prepared	Analyzed	Dil Fac
d5-NEtFOSAA	91		70 - 130				12/22/21 07:45	12/28/21 14:57	1
13C2 PFDA	88		70 - 130				12/22/21 07:45	12/28/21 14:57	1
13C2 PFHxA	89		70 - 130				12/22/21 07:45	12/28/21 14:57	1

Client Sample ID: 7670061 301 Conley Between Lead & Lag

Lab Sample ID: 410-67522-6

FB

Date Collected: 12/14/21 10:30

Matrix: Potable Water

Date Received: 12/17/21 15:20

Method: EPA 537 Ver 1.1 - EPA 537 Version 1.1

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorohexanoic acid	<1.8		1.8	0.44	ng/L		12/22/21 07:45	12/28/21 15:09	1
Perfluoroheptanoic acid	<1.8		1.8	0.44	ng/L		12/22/21 07:45	12/28/21 15:09	1
Perfluorooctanoic acid	<1.8		1.8	0.44	ng/L		12/22/21 07:45	12/28/21 15:09	1
Perfluorononanoic acid	<1.8		1.8	0.44	ng/L		12/22/21 07:45	12/28/21 15:09	1
Perfluorodecanoic acid	<1.8		1.8	0.44	ng/L		12/22/21 07:45	12/28/21 15:09	1
Perfluorotridecanoic acid	<1.8		1.8	0.44	ng/L		12/22/21 07:45	12/28/21 15:09	1
Perfluorotetradecanoic acid	<1.8		1.8	0.44	ng/L		12/22/21 07:45	12/28/21 15:09	1
Perfluorobutanesulfonic acid	<1.8		1.8	0.44	ng/L		12/22/21 07:45	12/28/21 15:09	1
Perfluorohexanesulfonic acid	<1.8		1.8	0.44	ng/L		12/22/21 07:45	12/28/21 15:09	1
Perfluorooctanesulfonic acid	<1.8		1.8	0.44	ng/L		12/22/21 07:45	12/28/21 15:09	1
NEtFOSAA	<1.8		1.8	0.44	ng/L		12/22/21 07:45	12/28/21 15:09	1
NMeFOSAA	<1.8		1.8	0.44	ng/L		12/22/21 07:45	12/28/21 15:09	1
Perfluoroundecanoic acid	<1.8		1.8	0.44	ng/L		12/22/21 07:45	12/28/21 15:09	1
Perfluorododecanoic acid	<1.8		1.8	0.44	ng/L		12/22/21 07:45	12/28/21 15:09	1
Surrogate	%Recovery	Qualifier	Limits				Prepared	Analyzed	Dil Fac
d5-NEtFOSAA	93		70 - 130				12/22/21 07:45	12/28/21 15:09	1
13C2 PFDA	89		70 - 130				12/22/21 07:45	12/28/21 15:09	1
13C2 PFHxA	96		70 - 130				12/22/21 07:45	12/28/21 15:09	1

Client Sample Results

Client: SUEZ Water Environmental Services Inc
Project/Site: Newberry System

Job ID: 410-67522-1

Client Sample ID: 7670061 301 Conley After Lag Vessel

Lab Sample ID: 410-67522-7

Date Collected: 12/14/21 10:25

Matrix: Drinking Water

Date Received: 12/17/21 15:20

Method: EPA 537 Ver 1.1 - EPA 537 Version 1.1

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorohexanoic acid	8.4		1.7	0.42	ng/L		12/22/21 07:45	12/28/21 15:20	1
Perfluoroheptanoic acid	2.8		1.7	0.42	ng/L		12/22/21 07:45	12/28/21 15:20	1
Perfluorooctanoic acid	6.5		1.7	0.42	ng/L		12/22/21 07:45	12/28/21 15:20	1
Perfluorononanoic acid	<1.7		1.7	0.42	ng/L		12/22/21 07:45	12/28/21 15:20	1
Perfluorodecanoic acid	<1.7		1.7	0.42	ng/L		12/22/21 07:45	12/28/21 15:20	1
Perfluorotridecanoic acid	<1.7		1.7	0.42	ng/L		12/22/21 07:45	12/28/21 15:20	1
Perfluorotetradecanoic acid	<1.7		1.7	0.42	ng/L		12/22/21 07:45	12/28/21 15:20	1
Perfluorobutanesulfonic acid	4.3		1.7	0.42	ng/L		12/22/21 07:45	12/28/21 15:20	1
Perfluorohexanesulfonic acid	5.2		1.7	0.42	ng/L		12/22/21 07:45	12/28/21 15:20	1
Perfluorooctanesulfonic acid	6.2		1.7	0.42	ng/L		12/22/21 07:45	12/28/21 15:20	1
NEtFOSAA	<1.7		1.7	0.42	ng/L		12/22/21 07:45	12/28/21 15:20	1
NMeFOSAA	<1.7		1.7	0.42	ng/L		12/22/21 07:45	12/28/21 15:20	1
Perfluoroundecanoic acid	<1.7		1.7	0.42	ng/L		12/22/21 07:45	12/28/21 15:20	1
Perfluorododecanoic acid	<1.7		1.7	0.42	ng/L		12/22/21 07:45	12/28/21 15:20	1
Surrogate	%Recovery	Qualifier	Limits				Prepared	Analyzed	Dil Fac
d5-NEtFOSAA	93		70 - 130				12/22/21 07:45	12/28/21 15:20	1
13C2 PFDA	85		70 - 130				12/22/21 07:45	12/28/21 15:20	1
13C2 PFHxA	89		70 - 130				12/22/21 07:45	12/28/21 15:20	1

Client Sample ID: 7670061 301 Conley After Lag Vessel FB

Lab Sample ID: 410-67522-8

Date Collected: 12/14/21 10:25

Matrix: Potable Water

Date Received: 12/17/21 15:20

Method: EPA 537 Ver 1.1 - EPA 537 Version 1.1

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorohexanoic acid	<1.8		1.8	0.45	ng/L		12/22/21 07:45	12/28/21 15:32	1
Perfluoroheptanoic acid	<1.8		1.8	0.45	ng/L		12/22/21 07:45	12/28/21 15:32	1
Perfluorooctanoic acid	<1.8		1.8	0.45	ng/L		12/22/21 07:45	12/28/21 15:32	1
Perfluorononanoic acid	<1.8		1.8	0.45	ng/L		12/22/21 07:45	12/28/21 15:32	1
Perfluorodecanoic acid	<1.8		1.8	0.45	ng/L		12/22/21 07:45	12/28/21 15:32	1
Perfluorotridecanoic acid	<1.8		1.8	0.45	ng/L		12/22/21 07:45	12/28/21 15:32	1
Perfluorotetradecanoic acid	<1.8		1.8	0.45	ng/L		12/22/21 07:45	12/28/21 15:32	1
Perfluorobutanesulfonic acid	<1.8		1.8	0.45	ng/L		12/22/21 07:45	12/28/21 15:32	1
Perfluorohexanesulfonic acid	<1.8		1.8	0.45	ng/L		12/22/21 07:45	12/28/21 15:32	1
Perfluorooctanesulfonic acid	<1.8		1.8	0.45	ng/L		12/22/21 07:45	12/28/21 15:32	1
NEtFOSAA	<1.8		1.8	0.45	ng/L		12/22/21 07:45	12/28/21 15:32	1
NMeFOSAA	<1.8		1.8	0.45	ng/L		12/22/21 07:45	12/28/21 15:32	1
Perfluoroundecanoic acid	<1.8		1.8	0.45	ng/L		12/22/21 07:45	12/28/21 15:32	1
Perfluorododecanoic acid	<1.8		1.8	0.45	ng/L		12/22/21 07:45	12/28/21 15:32	1
Surrogate	%Recovery	Qualifier	Limits				Prepared	Analyzed	Dil Fac
d5-NEtFOSAA	92		70 - 130				12/22/21 07:45	12/28/21 15:32	1
13C2 PFDA	87		70 - 130				12/22/21 07:45	12/28/21 15:32	1
13C2 PFHxA	91		70 - 130				12/22/21 07:45	12/28/21 15:32	1

Client Sample Results

Client: SUEZ Water Environmental Services Inc
Project/Site: Newberry System

Job ID: 410-67522-1

Client Sample ID: 7670061 101 Conley EP Grab Water

Lab Sample ID: 410-67522-9

Date Collected: 12/14/21 10:20

Matrix: Drinking Water

Date Received: 12/17/21 15:20

Method: EPA 537 Ver 1.1 - EPA 537 Version 1.1

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorohexanoic acid	8.4		1.8	0.44	ng/L		12/22/21 07:45	12/28/21 15:43	1
Perfluoroheptanoic acid	2.8		1.8	0.44	ng/L		12/22/21 07:45	12/28/21 15:43	1
Perfluorooctanoic acid	6.5		1.8	0.44	ng/L		12/22/21 07:45	12/28/21 15:43	1
Perfluorononanoic acid	<1.8		1.8	0.44	ng/L		12/22/21 07:45	12/28/21 15:43	1
Perfluorodecanoic acid	<1.8		1.8	0.44	ng/L		12/22/21 07:45	12/28/21 15:43	1
Perfluorotridecanoic acid	<1.8		1.8	0.44	ng/L		12/22/21 07:45	12/28/21 15:43	1
Perfluorotetradecanoic acid	<1.8		1.8	0.44	ng/L		12/22/21 07:45	12/28/21 15:43	1
Perfluorobutanesulfonic acid	4.2		1.8	0.44	ng/L		12/22/21 07:45	12/28/21 15:43	1
Perfluorohexanesulfonic acid	5.0		1.8	0.44	ng/L		12/22/21 07:45	12/28/21 15:43	1
Perfluorooctanesulfonic acid	6.2		1.8	0.44	ng/L		12/22/21 07:45	12/28/21 15:43	1
NEtFOSAA	<1.8		1.8	0.44	ng/L		12/22/21 07:45	12/28/21 15:43	1
NMeFOSAA	<1.8		1.8	0.44	ng/L		12/22/21 07:45	12/28/21 15:43	1
Perfluoroundecanoic acid	<1.8		1.8	0.44	ng/L		12/22/21 07:45	12/28/21 15:43	1
Perfluorododecanoic acid	<1.8		1.8	0.44	ng/L		12/22/21 07:45	12/28/21 15:43	1
Surrogate	%Recovery	Qualifier	Limits				Prepared	Analyzed	Dil Fac
d5-NEtFOSAA	91		70 - 130				12/22/21 07:45	12/28/21 15:43	1
13C2 PFDA	89		70 - 130				12/22/21 07:45	12/28/21 15:43	1
13C2 PFHxA	89		70 - 130				12/22/21 07:45	12/28/21 15:43	1

Client Sample ID: 7670061 101 Conley Field Blank Grab Water

Lab Sample ID: 410-67522-10

Date Collected: 12/14/21 10:20

Matrix: Potable Water

Date Received: 12/17/21 15:20

Method: EPA 537 Ver 1.1 - EPA 537 Version 1.1

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorohexanoic acid	<1.9		1.9	0.46	ng/L		12/22/21 07:45	12/28/21 15:55	1
Perfluoroheptanoic acid	<1.9		1.9	0.46	ng/L		12/22/21 07:45	12/28/21 15:55	1
Perfluorooctanoic acid	<1.9		1.9	0.46	ng/L		12/22/21 07:45	12/28/21 15:55	1
Perfluorononanoic acid	<1.9		1.9	0.46	ng/L		12/22/21 07:45	12/28/21 15:55	1
Perfluorodecanoic acid	<1.9		1.9	0.46	ng/L		12/22/21 07:45	12/28/21 15:55	1
Perfluorotridecanoic acid	<1.9		1.9	0.46	ng/L		12/22/21 07:45	12/28/21 15:55	1
Perfluorotetradecanoic acid	<1.9		1.9	0.46	ng/L		12/22/21 07:45	12/28/21 15:55	1
Perfluorobutanesulfonic acid	<1.9		1.9	0.46	ng/L		12/22/21 07:45	12/28/21 15:55	1
Perfluorohexanesulfonic acid	<1.9		1.9	0.46	ng/L		12/22/21 07:45	12/28/21 15:55	1
Perfluorooctanesulfonic acid	<1.9		1.9	0.46	ng/L		12/22/21 07:45	12/28/21 15:55	1
NEtFOSAA	<1.9		1.9	0.46	ng/L		12/22/21 07:45	12/28/21 15:55	1
NMeFOSAA	<1.9		1.9	0.46	ng/L		12/22/21 07:45	12/28/21 15:55	1
Perfluoroundecanoic acid	<1.9		1.9	0.46	ng/L		12/22/21 07:45	12/28/21 15:55	1
Perfluorododecanoic acid	<1.9		1.9	0.46	ng/L		12/22/21 07:45	12/28/21 15:55	1
Surrogate	%Recovery	Qualifier	Limits				Prepared	Analyzed	Dil Fac
d5-NEtFOSAA	90		70 - 130				12/22/21 07:45	12/28/21 15:55	1
13C2 PFDA	84		70 - 130				12/22/21 07:45	12/28/21 15:55	1
13C2 PFHxA	89		70 - 130				12/22/21 07:45	12/28/21 15:55	1

Client Sample Results

Client: SUEZ Water Environmental Services Inc
Project/Site: Newberry System

Job ID: 410-67522-1

Client Sample ID: 7670061 003 DuPont Well

Lab Sample ID: 410-67522-11

Date Collected: 12/14/21 09:55

Matrix: Drinking Water

Date Received: 12/17/21 15:20

Method: EPA 537 Ver 1.1 - EPA 537 Version 1.1

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorohexanoic acid	8.3		1.8	0.44	ng/L		12/22/21 07:45	12/28/21 16:18	1
Perfluoroheptanoic acid	3.9		1.8	0.44	ng/L		12/22/21 07:45	12/28/21 16:18	1
Perfluorooctanoic acid	6.5		1.8	0.44	ng/L		12/22/21 07:45	12/28/21 16:18	1
Perfluorononanoic acid	3.2		1.8	0.44	ng/L		12/22/21 07:45	12/28/21 16:18	1
Perfluorodecanoic acid	<1.8		1.8	0.44	ng/L		12/22/21 07:45	12/28/21 16:18	1
Perfluorotridecanoic acid	<1.8		1.8	0.44	ng/L		12/22/21 07:45	12/28/21 16:18	1
Perfluorotetradecanoic acid	<1.8		1.8	0.44	ng/L		12/22/21 07:45	12/28/21 16:18	1
Perfluorobutanesulfonic acid	7.9		1.8	0.44	ng/L		12/22/21 07:45	12/28/21 16:18	1
Perfluorohexanesulfonic acid	72		18	4.4	ng/L		12/22/21 07:45	01/03/22 14:10	10
Perfluorooctanesulfonic acid	74		18	4.4	ng/L		12/22/21 07:45	01/03/22 14:10	10
NEtFOSAA	<1.8		1.8	0.44	ng/L		12/22/21 07:45	12/28/21 16:18	1
NMeFOSAA	<1.8		1.8	0.44	ng/L		12/22/21 07:45	12/28/21 16:18	1
Perfluoroundecanoic acid	<1.8		1.8	0.44	ng/L		12/22/21 07:45	12/28/21 16:18	1
Perfluorododecanoic acid	<1.8		1.8	0.44	ng/L		12/22/21 07:45	12/28/21 16:18	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
d5-NEtFOSAA	95		70 - 130	12/22/21 07:45	12/28/21 16:18	1
d5-NMeFOSAA	102		70 - 130	12/22/21 07:45	01/03/22 14:10	10
13C2 PFDA	94		70 - 130	12/22/21 07:45	12/28/21 16:18	1
13C2 PFDA	102		70 - 130	12/22/21 07:45	01/03/22 14:10	10
13C2 PFHxA	90		70 - 130	12/22/21 07:45	12/28/21 16:18	1
13C2 PFHxA	107		70 - 130	12/22/21 07:45	01/03/22 14:10	10

Client Sample ID: 7670061 003 DuPont Well FB

Lab Sample ID: 410-67522-12

Date Collected: 12/14/21 09:55

Matrix: Potable Water

Date Received: 12/17/21 15:20

Method: EPA 537 Ver 1.1 - EPA 537 Version 1.1

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorohexanoic acid	<1.9		1.9	0.48	ng/L		12/22/21 07:45	12/28/21 16:30	1
Perfluoroheptanoic acid	<1.9		1.9	0.48	ng/L		12/22/21 07:45	12/28/21 16:30	1
Perfluorooctanoic acid	<1.9		1.9	0.48	ng/L		12/22/21 07:45	12/28/21 16:30	1
Perfluorononanoic acid	<1.9		1.9	0.48	ng/L		12/22/21 07:45	12/28/21 16:30	1
Perfluorodecanoic acid	<1.9		1.9	0.48	ng/L		12/22/21 07:45	12/28/21 16:30	1
Perfluorotridecanoic acid	<1.9		1.9	0.48	ng/L		12/22/21 07:45	12/28/21 16:30	1
Perfluorotetradecanoic acid	<1.9		1.9	0.48	ng/L		12/22/21 07:45	12/28/21 16:30	1
Perfluorobutanesulfonic acid	<1.9		1.9	0.48	ng/L		12/22/21 07:45	12/28/21 16:30	1
Perfluorohexanesulfonic acid	<1.9		1.9	0.48	ng/L		12/22/21 07:45	12/28/21 16:30	1
Perfluorooctanesulfonic acid	<1.9		1.9	0.48	ng/L		12/22/21 07:45	12/28/21 16:30	1
NEtFOSAA	<1.9		1.9	0.48	ng/L		12/22/21 07:45	12/28/21 16:30	1
NMeFOSAA	<1.9		1.9	0.48	ng/L		12/22/21 07:45	12/28/21 16:30	1
Perfluoroundecanoic acid	<1.9		1.9	0.48	ng/L		12/22/21 07:45	12/28/21 16:30	1
Perfluorododecanoic acid	<1.9		1.9	0.48	ng/L		12/22/21 07:45	12/28/21 16:30	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
d5-NEtFOSAA	105		70 - 130	12/22/21 07:45	12/28/21 16:30	1
13C2 PFDA	97		70 - 130	12/22/21 07:45	12/28/21 16:30	1
13C2 PFHxA	103		70 - 130	12/22/21 07:45	12/28/21 16:30	1

Client Sample Results

Client: SUEZ Water Environmental Services Inc
Project/Site: Newberry System

Job ID: 410-67522-1

Client Sample ID: 7670061 302 DuPont Between Lead & Lag

Lab Sample ID: 410-67522-13

Date Collected: 12/14/21 09:50

Matrix: Drinking Water

Date Received: 12/17/21 15:20

Method: EPA 537 Ver 1.1 - EPA 537 Version 1.1

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorohexanoic acid	18		1.8	0.44	ng/L		12/22/21 07:45	12/28/21 16:41	1
Perfluoroheptanoic acid	3.4		1.8	0.44	ng/L		12/22/21 07:45	12/28/21 16:41	1
Perfluorooctanoic acid	2.6		1.8	0.44	ng/L		12/22/21 07:45	12/28/21 16:41	1
Perfluorononanoic acid	<1.8		1.8	0.44	ng/L		12/22/21 07:45	12/28/21 16:41	1
Perfluorodecanoic acid	<1.8		1.8	0.44	ng/L		12/22/21 07:45	12/28/21 16:41	1
Perfluorotridecanoic acid	<1.8		1.8	0.44	ng/L		12/22/21 07:45	12/28/21 16:41	1
Perfluorotetradecanoic acid	<1.8		1.8	0.44	ng/L		12/22/21 07:45	12/28/21 16:41	1
Perfluorobutanesulfonic acid	8.2		1.8	0.44	ng/L		12/22/21 07:45	12/28/21 16:41	1
Perfluorohexanesulfonic acid	16		1.8	0.44	ng/L		12/22/21 07:45	12/28/21 16:41	1
Perfluorooctanesulfonic acid	4.4		1.8	0.44	ng/L		12/22/21 07:45	12/28/21 16:41	1
NETFOSAA	<1.8		1.8	0.44	ng/L		12/22/21 07:45	12/28/21 16:41	1
NMeFOSAA	<1.8		1.8	0.44	ng/L		12/22/21 07:45	12/28/21 16:41	1
Perfluoroundecanoic acid	<1.8		1.8	0.44	ng/L		12/22/21 07:45	12/28/21 16:41	1
Perfluorododecanoic acid	<1.8		1.8	0.44	ng/L		12/22/21 07:45	12/28/21 16:41	1
Surrogate	%Recovery	Qualifier	Limits				Prepared	Analyzed	Dil Fac
d5-NEtFOSAA	100		70 - 130				12/22/21 07:45	12/28/21 16:41	1
13C2 PFDA	96		70 - 130				12/22/21 07:45	12/28/21 16:41	1
13C2 PFHxA	101		70 - 130				12/22/21 07:45	12/28/21 16:41	1

Client Sample ID: 7670061 302 DuPont Between Lead & Lag

Lab Sample ID: 410-67522-14

FB

Date Collected: 12/14/21 09:50

Matrix: Potable Water

Date Received: 12/17/21 15:20

Method: EPA 537 Ver 1.1 - EPA 537 Version 1.1

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorohexanoic acid	<1.7		1.7	0.43	ng/L		12/22/21 07:45	12/28/21 16:53	1
Perfluoroheptanoic acid	<1.7		1.7	0.43	ng/L		12/22/21 07:45	12/28/21 16:53	1
Perfluorooctanoic acid	<1.7		1.7	0.43	ng/L		12/22/21 07:45	12/28/21 16:53	1
Perfluorononanoic acid	<1.7		1.7	0.43	ng/L		12/22/21 07:45	12/28/21 16:53	1
Perfluorodecanoic acid	<1.7		1.7	0.43	ng/L		12/22/21 07:45	12/28/21 16:53	1
Perfluorotridecanoic acid	<1.7		1.7	0.43	ng/L		12/22/21 07:45	12/28/21 16:53	1
Perfluorotetradecanoic acid	<1.7		1.7	0.43	ng/L		12/22/21 07:45	12/28/21 16:53	1
Perfluorobutanesulfonic acid	<1.7		1.7	0.43	ng/L		12/22/21 07:45	12/28/21 16:53	1
Perfluorohexanesulfonic acid	<1.7		1.7	0.43	ng/L		12/22/21 07:45	12/28/21 16:53	1
Perfluorooctanesulfonic acid	<1.7		1.7	0.43	ng/L		12/22/21 07:45	12/28/21 16:53	1
NETFOSAA	<1.7		1.7	0.43	ng/L		12/22/21 07:45	12/28/21 16:53	1
NMeFOSAA	<1.7		1.7	0.43	ng/L		12/22/21 07:45	12/28/21 16:53	1
Perfluoroundecanoic acid	<1.7		1.7	0.43	ng/L		12/22/21 07:45	12/28/21 16:53	1
Perfluorododecanoic acid	<1.7		1.7	0.43	ng/L		12/22/21 07:45	12/28/21 16:53	1
Surrogate	%Recovery	Qualifier	Limits				Prepared	Analyzed	Dil Fac
d5-NEtFOSAA	95		70 - 130				12/22/21 07:45	12/28/21 16:53	1
13C2 PFDA	94		70 - 130				12/22/21 07:45	12/28/21 16:53	1
13C2 PFHxA	95		70 - 130				12/22/21 07:45	12/28/21 16:53	1

Client Sample Results

Client: SUEZ Water Environmental Services Inc
Project/Site: Newberry System

Job ID: 410-67522-1

Client Sample ID: 7670061 302 DuPont After Lag Vessel

Lab Sample ID: 410-67522-15

Date Collected: 12/14/21 09:45

Matrix: Drinking Water

Date Received: 12/17/21 15:20

Method: EPA 537 Ver 1.1 - EPA 537 Version 1.1

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorohexanoic acid	12		1.8	0.44	ng/L		12/22/21 07:45	12/28/21 17:04	1
Perfluoroheptanoic acid	<1.8		1.8	0.44	ng/L		12/22/21 07:45	12/28/21 17:04	1
Perfluorooctanoic acid	<1.8		1.8	0.44	ng/L		12/22/21 07:45	12/28/21 17:04	1
Perfluorononanoic acid	<1.8		1.8	0.44	ng/L		12/22/21 07:45	12/28/21 17:04	1
Perfluorodecanoic acid	<1.8		1.8	0.44	ng/L		12/22/21 07:45	12/28/21 17:04	1
Perfluorotridecanoic acid	<1.8		1.8	0.44	ng/L		12/22/21 07:45	12/28/21 17:04	1
Perfluorotetradecanoic acid	<1.8		1.8	0.44	ng/L		12/22/21 07:45	12/28/21 17:04	1
Perfluorobutanesulfonic acid	6.1		1.8	0.44	ng/L		12/22/21 07:45	12/28/21 17:04	1
Perfluorohexanesulfonic acid	1.9		1.8	0.44	ng/L		12/22/21 07:45	12/28/21 17:04	1
Perfluorooctanesulfonic acid	<1.8		1.8	0.44	ng/L		12/22/21 07:45	12/28/21 17:04	1
NEtFOSAA	<1.8		1.8	0.44	ng/L		12/22/21 07:45	12/28/21 17:04	1
NMeFOSAA	<1.8		1.8	0.44	ng/L		12/22/21 07:45	12/28/21 17:04	1
Perfluoroundecanoic acid	<1.8		1.8	0.44	ng/L		12/22/21 07:45	12/28/21 17:04	1
Perfluorododecanoic acid	<1.8		1.8	0.44	ng/L		12/22/21 07:45	12/28/21 17:04	1
Surrogate	%Recovery	Qualifier	Limits				Prepared	Analyzed	Dil Fac
d5-NEtFOSAA	90		70 - 130				12/22/21 07:45	12/28/21 17:04	1
13C2 PFDA	85		70 - 130				12/22/21 07:45	12/28/21 17:04	1
13C2 PFHxA	92		70 - 130				12/22/21 07:45	12/28/21 17:04	1

Client Sample ID: 7670061 302 DuPont After Lag Vessel FB

Lab Sample ID: 410-67522-16

Date Collected: 12/14/21 09:45

Matrix: Potable Water

Date Received: 12/17/21 15:20

Method: EPA 537 Ver 1.1 - EPA 537 Version 1.1

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorohexanoic acid	<1.7		1.7	0.43	ng/L		12/22/21 07:45	12/28/21 17:16	1
Perfluoroheptanoic acid	<1.7		1.7	0.43	ng/L		12/22/21 07:45	12/28/21 17:16	1
Perfluorooctanoic acid	<1.7		1.7	0.43	ng/L		12/22/21 07:45	12/28/21 17:16	1
Perfluorononanoic acid	<1.7		1.7	0.43	ng/L		12/22/21 07:45	12/28/21 17:16	1
Perfluorodecanoic acid	<1.7		1.7	0.43	ng/L		12/22/21 07:45	12/28/21 17:16	1
Perfluorotridecanoic acid	<1.7		1.7	0.43	ng/L		12/22/21 07:45	12/28/21 17:16	1
Perfluorotetradecanoic acid	<1.7		1.7	0.43	ng/L		12/22/21 07:45	12/28/21 17:16	1
Perfluorobutanesulfonic acid	<1.7		1.7	0.43	ng/L		12/22/21 07:45	12/28/21 17:16	1
Perfluorohexanesulfonic acid	<1.7		1.7	0.43	ng/L		12/22/21 07:45	12/28/21 17:16	1
Perfluorooctanesulfonic acid	<1.7		1.7	0.43	ng/L		12/22/21 07:45	12/28/21 17:16	1
NEtFOSAA	<1.7		1.7	0.43	ng/L		12/22/21 07:45	12/28/21 17:16	1
NMeFOSAA	<1.7		1.7	0.43	ng/L		12/22/21 07:45	12/28/21 17:16	1
Perfluoroundecanoic acid	<1.7		1.7	0.43	ng/L		12/22/21 07:45	12/28/21 17:16	1
Perfluorododecanoic acid	<1.7		1.7	0.43	ng/L		12/22/21 07:45	12/28/21 17:16	1
Surrogate	%Recovery	Qualifier	Limits				Prepared	Analyzed	Dil Fac
d5-NEtFOSAA	94		70 - 130				12/22/21 07:45	12/28/21 17:16	1
13C2 PFDA	87		70 - 130				12/22/21 07:45	12/28/21 17:16	1
13C2 PFHxA	92		70 - 130				12/22/21 07:45	12/28/21 17:16	1

Client Sample Results

Client: SUEZ Water Environmental Services Inc
Project/Site: Newberry System

Job ID: 410-67522-1

Client Sample ID: 7670061 102 DuPont EP

Lab Sample ID: 410-67522-17

Date Collected: 12/14/21 09:40

Matrix: Drinking Water

Date Received: 12/17/21 15:20

Method: EPA 537 Ver 1.1 - EPA 537 Version 1.1

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorohexanoic acid	13		1.8	0.45	ng/L		12/22/21 07:45	12/28/21 17:28	1
Perfluoroheptanoic acid	<1.8		1.8	0.45	ng/L		12/22/21 07:45	12/28/21 17:28	1
Perfluorooctanoic acid	<1.8		1.8	0.45	ng/L		12/22/21 07:45	12/28/21 17:28	1
Perfluorononanoic acid	<1.8		1.8	0.45	ng/L		12/22/21 07:45	12/28/21 17:28	1
Perfluorodecanoic acid	<1.8		1.8	0.45	ng/L		12/22/21 07:45	12/28/21 17:28	1
Perfluorotridecanoic acid	<1.8		1.8	0.45	ng/L		12/22/21 07:45	12/28/21 17:28	1
Perfluorotetradecanoic acid	<1.8		1.8	0.45	ng/L		12/22/21 07:45	12/28/21 17:28	1
Perfluorobutanesulfonic acid	6.5		1.8	0.45	ng/L		12/22/21 07:45	12/28/21 17:28	1
Perfluorohexanesulfonic acid	2.1		1.8	0.45	ng/L		12/22/21 07:45	12/28/21 17:28	1
Perfluorooctanesulfonic acid	<1.8		1.8	0.45	ng/L		12/22/21 07:45	12/28/21 17:28	1
NEtFOSAA	<1.8		1.8	0.45	ng/L		12/22/21 07:45	12/28/21 17:28	1
NMeFOSAA	<1.8		1.8	0.45	ng/L		12/22/21 07:45	12/28/21 17:28	1
Perfluoroundecanoic acid	<1.8		1.8	0.45	ng/L		12/22/21 07:45	12/28/21 17:28	1
Perfluorododecanoic acid	<1.8		1.8	0.45	ng/L		12/22/21 07:45	12/28/21 17:28	1
Surrogate	%Recovery	Qualifier	Limits				Prepared	Analyzed	Dil Fac
d5-NEtFOSAA	102		70 - 130				12/22/21 07:45	12/28/21 17:28	1
13C2 PFDA	93		70 - 130				12/22/21 07:45	12/28/21 17:28	1
13C2 PFHxA	97		70 - 130				12/22/21 07:45	12/28/21 17:28	1

Client Sample ID: 7670061 102 DuPont EP FB

Lab Sample ID: 410-67522-18

Date Collected: 12/14/21 09:40

Matrix: Potable Water

Date Received: 12/17/21 15:20

Method: EPA 537 Ver 1.1 - EPA 537 Version 1.1

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorohexanoic acid	<1.8		1.8	0.44	ng/L		12/22/21 07:45	12/28/21 17:39	1
Perfluoroheptanoic acid	<1.8		1.8	0.44	ng/L		12/22/21 07:45	12/28/21 17:39	1
Perfluorooctanoic acid	<1.8		1.8	0.44	ng/L		12/22/21 07:45	12/28/21 17:39	1
Perfluorononanoic acid	<1.8		1.8	0.44	ng/L		12/22/21 07:45	12/28/21 17:39	1
Perfluorodecanoic acid	<1.8		1.8	0.44	ng/L		12/22/21 07:45	12/28/21 17:39	1
Perfluorotridecanoic acid	<1.8		1.8	0.44	ng/L		12/22/21 07:45	12/28/21 17:39	1
Perfluorotetradecanoic acid	<1.8		1.8	0.44	ng/L		12/22/21 07:45	12/28/21 17:39	1
Perfluorobutanesulfonic acid	<1.8		1.8	0.44	ng/L		12/22/21 07:45	12/28/21 17:39	1
Perfluorohexanesulfonic acid	<1.8		1.8	0.44	ng/L		12/22/21 07:45	12/28/21 17:39	1
Perfluorooctanesulfonic acid	<1.8		1.8	0.44	ng/L		12/22/21 07:45	12/28/21 17:39	1
NEtFOSAA	<1.8		1.8	0.44	ng/L		12/22/21 07:45	12/28/21 17:39	1
NMeFOSAA	<1.8		1.8	0.44	ng/L		12/22/21 07:45	12/28/21 17:39	1
Perfluoroundecanoic acid	<1.8		1.8	0.44	ng/L		12/22/21 07:45	12/28/21 17:39	1
Perfluorododecanoic acid	<1.8		1.8	0.44	ng/L		12/22/21 07:45	12/28/21 17:39	1
Surrogate	%Recovery	Qualifier	Limits				Prepared	Analyzed	Dil Fac
d5-NEtFOSAA	82		70 - 130				12/22/21 07:45	12/28/21 17:39	1
13C2 PFDA	78		70 - 130				12/22/21 07:45	12/28/21 17:39	1
13C2 PFHxA	85		70 - 130				12/22/21 07:45	12/28/21 17:39	1

Surrogate Summary

Client: SUEZ Water Environmental Services Inc
 Project/Site: Newberry System

Job ID: 410-67522-1

Method: EPA 537 Ver 1.1 - EPA 537 Version 1.1

Matrix: Drinking Water

Prep Type: Total/NA

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)		
		d5NEFOS (70-130)	PFDA (70-130)	PFHxA (70-130)
410-67522-1	7670061 001 Playground Well	90	88	89
410-67522-3	7670061 005 Conley Well	91	87	90
410-67522-5	7670061 301 Conley Between Lead & Lag	91	88	89
410-67522-7	7670061 301 Conley After Lag Vessel	93	85	89
410-67522-9	7670061 101 Conley EP Grab Water	91	89	89
410-67522-11	7670061 003 DuPont Well	95	94	90
410-67522-11	7670061 003 DuPont Well	102	102	107
410-67522-13	7670061 302 DuPont Between Lead & Lag	100	96	101
410-67522-15	7670061 302 DuPont After Lag Vessel	90	85	92
410-67522-17	7670061 102 DuPont EP	102	93	97
LCS 410-208332/2-A	Lab Control Sample	94	92	96
LCS 410-208332/3-A	Lab Control Sample Dup	91	85	90
LLCS 410-208332/4-A	Lab Control Sample	89	86	89
MB 410-208332/1-A	Method Blank	97	91	94

Surrogate Legend

d5NEFOS = d5-NEtFOSAA
 PFDA = 13C2 PFDA
 PFHxA = 13C2 PFHxA

Method: EPA 537 Ver 1.1 - EPA 537 Version 1.1

Matrix: Potable Water

Prep Type: Total/NA

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)		
		d5NEFOS (70-130)	PFDA (70-130)	PFHxA (70-130)
410-67522-2	7670061 001 Playground Well FB	92	91	93
410-67522-4	7670061 005 Conley Well FB	94	91	96
410-67522-6	7670061 301 Conley Between Lead & Lag FB	93	89	96
410-67522-8	7670061 301 Conley After Lag Vessel FB	92	87	91
410-67522-10	7670061 101 Conley Field Blank Grab Water	90	84	89
410-67522-12	7670061 003 DuPont Well FB	105	97	103
410-67522-14	7670061 302 DuPont Between Lead & Lag FB	95	94	95
410-67522-16	7670061 302 DuPont After Lag Vessel FB	94	87	92
410-67522-18	7670061 102 DuPont EP FB	82	78	85

Surrogate Legend

d5NEFOS = d5-NEtFOSAA
 PFDA = 13C2 PFDA
 PFHxA = 13C2 PFHxA

QC Sample Results

Client: SUEZ Water Environmental Services Inc
Project/Site: Newberry System

Job ID: 410-67522-1

Method: EPA 537 Ver 1.1 - EPA 537 Version 1.1

Lab Sample ID: MB 410-208332/1-A
Matrix: Drinking Water
Analysis Batch: 209647

Client Sample ID: Method Blank
Prep Type: Total/NA
Prep Batch: 208332

Analyte	MB	MB	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
	Result	Qualifier							
Perfluorohexanoic acid	<2.0		2.0	0.50	ng/L		12/22/21 07:45	12/28/21 13:23	1
Perfluoroheptanoic acid	<2.0		2.0	0.50	ng/L		12/22/21 07:45	12/28/21 13:23	1
Perfluorooctanoic acid	<2.0		2.0	0.50	ng/L		12/22/21 07:45	12/28/21 13:23	1
Perfluorononanoic acid	<2.0		2.0	0.50	ng/L		12/22/21 07:45	12/28/21 13:23	1
Perfluorodecanoic acid	<2.0		2.0	0.50	ng/L		12/22/21 07:45	12/28/21 13:23	1
Perfluorotridecanoic acid	<2.0		2.0	0.50	ng/L		12/22/21 07:45	12/28/21 13:23	1
Perfluorotetradecanoic acid	<2.0		2.0	0.50	ng/L		12/22/21 07:45	12/28/21 13:23	1
Perfluorobutanesulfonic acid	<2.0		2.0	0.50	ng/L		12/22/21 07:45	12/28/21 13:23	1
Perfluorohexanesulfonic acid	<2.0		2.0	0.50	ng/L		12/22/21 07:45	12/28/21 13:23	1
Perfluorooctanesulfonic acid	<2.0		2.0	0.50	ng/L		12/22/21 07:45	12/28/21 13:23	1
NEtFOSAA	<2.0		2.0	0.50	ng/L		12/22/21 07:45	12/28/21 13:23	1
NMeFOSAA	<2.0		2.0	0.50	ng/L		12/22/21 07:45	12/28/21 13:23	1
Perfluoroundecanoic acid	<2.0		2.0	0.50	ng/L		12/22/21 07:45	12/28/21 13:23	1
Perfluorododecanoic acid	<2.0		2.0	0.50	ng/L		12/22/21 07:45	12/28/21 13:23	1

Surrogate	MB	MB	Limits	Prepared	Analyzed	Dil Fac
	%Recovery	Qualifier				
d5-NEtFOSAA	97		70 - 130	12/22/21 07:45	12/28/21 13:23	1
13C2 PFDA	91		70 - 130	12/22/21 07:45	12/28/21 13:23	1
13C2 PFHxA	94		70 - 130	12/22/21 07:45	12/28/21 13:23	1

Lab Sample ID: LCS 410-208332/2-A
Matrix: Drinking Water
Analysis Batch: 209647

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 208332

Analyte	Spike Added	LCS	LCS	Unit	D	%Rec	%Rec. Limits
		Result	Qualifier				
Perfluorohexanoic acid	20.5	19.8		ng/L		96	70 - 130
Perfluoroheptanoic acid	20.5	19.8		ng/L		97	70 - 130
Perfluorooctanoic acid	20.5	20.6		ng/L		100	70 - 130
Perfluorononanoic acid	20.5	18.9		ng/L		93	70 - 130
Perfluorodecanoic acid	20.5	19.0		ng/L		93	70 - 130
Perfluorotridecanoic acid	20.5	18.4		ng/L		90	70 - 130
Perfluorotetradecanoic acid	20.5	19.2		ng/L		94	70 - 130
Perfluorobutanesulfonic acid	18.1	20.0		ng/L		111	70 - 130
Perfluorohexanesulfonic acid	18.7	18.6		ng/L		99	70 - 130
Perfluorooctanesulfonic acid	19.0	18.6		ng/L		98	70 - 130
NEtFOSAA	20.5	19.0		ng/L		93	70 - 130
NMeFOSAA	20.5	18.9		ng/L		92	70 - 130
Perfluoroundecanoic acid	20.5	20.2		ng/L		98	70 - 130
Perfluorododecanoic acid	20.5	19.3		ng/L		94	70 - 130

Surrogate	LCS	LCS	Limits
	%Recovery	Qualifier	
d5-NEtFOSAA	94		70 - 130
13C2 PFDA	92		70 - 130
13C2 PFHxA	96		70 - 130

QC Sample Results

Client: SUEZ Water Environmental Services Inc
Project/Site: Newberry System

Job ID: 410-67522-1

Method: EPA 537 Ver 1.1 - EPA 537 Version 1.1 (Continued)

Lab Sample ID: LCSD 410-208332/3-A

Matrix: Drinking Water

Analysis Batch: 209647

Client Sample ID: Lab Control Sample Dup

Prep Type: Total/NA

Prep Batch: 208332

Analyte	Spike Added	LCSD Result	LCSD Qualifier	Unit	D	%Rec	%Rec. Limits	RPD	RPD Limit
Perfluorohexanoic acid	20.5	18.6		ng/L		91	70 - 130	6	30
Perfluoroheptanoic acid	20.5	18.3		ng/L		89	70 - 130	8	30
Perfluorooctanoic acid	20.5	19.5		ng/L		95	70 - 130	5	30
Perfluorononanoic acid	20.5	18.0		ng/L		88	70 - 130	5	30
Perfluorodecanoic acid	20.5	17.4		ng/L		85	70 - 130	8	30
Perfluorotridecanoic acid	20.5	17.8		ng/L		87	70 - 130	4	30
Perfluorotetradecanoic acid	20.5	17.6		ng/L		86	70 - 130	9	30
Perfluorobutanesulfonic acid	18.1	18.8		ng/L		104	70 - 130	6	30
Perfluorohexanesulfonic acid	18.7	17.4		ng/L		93	70 - 130	7	30
Perfluorooctanesulfonic acid	19.0	17.5		ng/L		92	70 - 130	6	30
NEtFOSAA	20.5	18.5		ng/L		90	70 - 130	3	30
NMeFOSAA	20.5	18.4		ng/L		90	70 - 130	2	30
Perfluoroundecanoic acid	20.5	19.0		ng/L		93	70 - 130	6	30
Perfluorododecanoic acid	20.5	18.2		ng/L		89	70 - 130	6	30

Surrogate	LCSD LCSD		Limits
	%Recovery	Qualifier	
d5-NEtFOSAA	91		70 - 130
13C2 PFDA	85		70 - 130
13C2 PFHxA	90		70 - 130

Lab Sample ID: LLCS 410-208332/4-A

Matrix: Drinking Water

Analysis Batch: 209647

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 208332

Analyte	Spike Added	LLCS Result	LLCS Qualifier	Unit	D	%Rec	%Rec. Limits
Perfluorohexanoic acid	1.92	1.67	J	ng/L		87	50 - 150
Perfluoroheptanoic acid	1.92	1.65	J	ng/L		86	50 - 150
Perfluorooctanoic acid	1.92	1.87	J	ng/L		97	50 - 150
Perfluorononanoic acid	1.92	1.60	J	ng/L		83	50 - 150
Perfluorodecanoic acid	1.92	1.54	J	ng/L		80	50 - 150
Perfluorotridecanoic acid	1.92	1.55	J	ng/L		81	50 - 150
Perfluorotetradecanoic acid	1.92	1.53	J	ng/L		80	50 - 150
Perfluorobutanesulfonic acid	1.70	1.67	J	ng/L		98	50 - 150
Perfluorohexanesulfonic acid	1.75	1.55	J	ng/L		88	50 - 150
Perfluorooctanesulfonic acid	1.78	1.56	J	ng/L		88	50 - 150
NEtFOSAA	1.92	1.68	J	ng/L		88	50 - 150
NMeFOSAA	1.92	1.63	J	ng/L		85	50 - 150
Perfluoroundecanoic acid	1.92	1.63	J	ng/L		85	50 - 150
Perfluorododecanoic acid	1.92	1.55	J	ng/L		81	50 - 150

Surrogate	LLCS LLCS		Limits
	%Recovery	Qualifier	
d5-NEtFOSAA	89		70 - 130
13C2 PFDA	86		70 - 130
13C2 PFHxA	89		70 - 130

QC Association Summary

Client: SUEZ Water Environmental Services Inc
 Project/Site: Newberry System

Job ID: 410-67522-1

LCMS

Prep Batch: 208332

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
410-67522-1	7670061 001 Playground Well	Total/NA	Drinking Water	EPA 537 Ver 1.1	
410-67522-2	7670061 001 Playground Well FB	Total/NA	Potable Water	EPA 537 Ver 1.1	
410-67522-3	7670061 005 Conley Well	Total/NA	Drinking Water	EPA 537 Ver 1.1	
410-67522-4	7670061 005 Conley Well FB	Total/NA	Potable Water	EPA 537 Ver 1.1	
410-67522-5	7670061 301 Conley Between Lead & Lag	Total/NA	Drinking Water	EPA 537 Ver 1.1	
410-67522-6	7670061 301 Conley Between Lead & Lag FB	Total/NA	Potable Water	EPA 537 Ver 1.1	
410-67522-7	7670061 301 Conley After Lag Vessel	Total/NA	Drinking Water	EPA 537 Ver 1.1	
410-67522-8	7670061 301 Conley After Lag Vessel FB	Total/NA	Potable Water	EPA 537 Ver 1.1	
410-67522-9	7670061 101 Conley EP Grab Water	Total/NA	Drinking Water	EPA 537 Ver 1.1	
410-67522-10	7670061 101 Conley Field Blank Grab Water	Total/NA	Potable Water	EPA 537 Ver 1.1	
410-67522-11	7670061 003 DuPont Well	Total/NA	Drinking Water	EPA 537 Ver 1.1	
410-67522-12	7670061 003 DuPont Well FB	Total/NA	Potable Water	EPA 537 Ver 1.1	
410-67522-13	7670061 302 DuPont Between Lead & Lag	Total/NA	Drinking Water	EPA 537 Ver 1.1	
410-67522-14	7670061 302 DuPont Between Lead & Lag FB	Total/NA	Potable Water	EPA 537 Ver 1.1	
410-67522-15	7670061 302 DuPont After Lag Vessel	Total/NA	Drinking Water	EPA 537 Ver 1.1	
410-67522-16	7670061 302 DuPont After Lag Vessel FB	Total/NA	Potable Water	EPA 537 Ver 1.1	
410-67522-17	7670061 102 DuPont EP	Total/NA	Drinking Water	EPA 537 Ver 1.1	
410-67522-18	7670061 102 DuPont EP FB	Total/NA	Potable Water	EPA 537 Ver 1.1	
MB 410-208332/1-A	Method Blank	Total/NA	Drinking Water	EPA 537 Ver 1.1	
LCS 410-208332/2-A	Lab Control Sample	Total/NA	Drinking Water	EPA 537 Ver 1.1	
LCSD 410-208332/3-A	Lab Control Sample Dup	Total/NA	Drinking Water	EPA 537 Ver 1.1	
LLCS 410-208332/4-A	Lab Control Sample	Total/NA	Drinking Water	EPA 537 Ver 1.1	

Analysis Batch: 209647

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
410-67522-1	7670061 001 Playground Well	Total/NA	Drinking Water	EPA 537 Ver 1.1	208332
410-67522-2	7670061 001 Playground Well FB	Total/NA	Potable Water	EPA 537 Ver 1.1	208332
410-67522-3	7670061 005 Conley Well	Total/NA	Drinking Water	EPA 537 Ver 1.1	208332
410-67522-4	7670061 005 Conley Well FB	Total/NA	Potable Water	EPA 537 Ver 1.1	208332
410-67522-5	7670061 301 Conley Between Lead & Lag	Total/NA	Drinking Water	EPA 537 Ver 1.1	208332
410-67522-6	7670061 301 Conley Between Lead & Lag FB	Total/NA	Potable Water	EPA 537 Ver 1.1	208332
410-67522-7	7670061 301 Conley After Lag Vessel	Total/NA	Drinking Water	EPA 537 Ver 1.1	208332
410-67522-8	7670061 301 Conley After Lag Vessel FB	Total/NA	Potable Water	EPA 537 Ver 1.1	208332
410-67522-9	7670061 101 Conley EP Grab Water	Total/NA	Drinking Water	EPA 537 Ver 1.1	208332
410-67522-10	7670061 101 Conley Field Blank Grab Water	Total/NA	Potable Water	EPA 537 Ver 1.1	208332
410-67522-11	7670061 003 DuPont Well	Total/NA	Drinking Water	EPA 537 Ver 1.1	208332
410-67522-12	7670061 003 DuPont Well FB	Total/NA	Potable Water	EPA 537 Ver 1.1	208332
410-67522-13	7670061 302 DuPont Between Lead & Lag	Total/NA	Drinking Water	EPA 537 Ver 1.1	208332
410-67522-14	7670061 302 DuPont Between Lead & Lag FB	Total/NA	Potable Water	EPA 537 Ver 1.1	208332
410-67522-15	7670061 302 DuPont After Lag Vessel	Total/NA	Drinking Water	EPA 537 Ver 1.1	208332
410-67522-16	7670061 302 DuPont After Lag Vessel FB	Total/NA	Potable Water	EPA 537 Ver 1.1	208332
410-67522-17	7670061 102 DuPont EP	Total/NA	Drinking Water	EPA 537 Ver 1.1	208332
410-67522-18	7670061 102 DuPont EP FB	Total/NA	Potable Water	EPA 537 Ver 1.1	208332
MB 410-208332/1-A	Method Blank	Total/NA	Drinking Water	EPA 537 Ver 1.1	208332
LCS 410-208332/2-A	Lab Control Sample	Total/NA	Drinking Water	EPA 537 Ver 1.1	208332
LCSD 410-208332/3-A	Lab Control Sample Dup	Total/NA	Drinking Water	EPA 537 Ver 1.1	208332
LLCS 410-208332/4-A	Lab Control Sample	Total/NA	Drinking Water	EPA 537 Ver 1.1	208332

Analysis Batch: 211118

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
410-67522-11	7670061 003 DuPont Well	Total/NA	Drinking Water	EPA 537 Ver 1.1	208332

Lab Chronicle

Client: SUEZ Water Environmental Services Inc
Project/Site: Newberry System

Job ID: 410-67522-1

Client Sample ID: 7670061 001 Playground Well

Lab Sample ID: 410-67522-1

Date Collected: 12/14/21 10:40

Matrix: Drinking Water

Date Received: 12/17/21 15:20

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	EPA 537 Ver 1.1			208332	12/22/21 07:45	RDL8	ELLE
Total/NA	Analysis	EPA 537 Ver 1.1		1	209647	12/28/21 14:10	DCS9	ELLE

Client Sample ID: 7670061 001 Playground Well FB

Lab Sample ID: 410-67522-2

Date Collected: 12/14/21 10:40

Matrix: Potable Water

Date Received: 12/17/21 15:20

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	EPA 537 Ver 1.1			208332	12/22/21 07:45	RDL8	ELLE
Total/NA	Analysis	EPA 537 Ver 1.1		1	209647	12/28/21 14:22	DCS9	ELLE

Client Sample ID: 7670061 005 Conley Well

Lab Sample ID: 410-67522-3

Date Collected: 12/14/21 10:35

Matrix: Drinking Water

Date Received: 12/17/21 15:20

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	EPA 537 Ver 1.1			208332	12/22/21 07:45	RDL8	ELLE
Total/NA	Analysis	EPA 537 Ver 1.1		1	209647	12/28/21 14:34	DCS9	ELLE

Client Sample ID: 7670061 005 Conley Well FB

Lab Sample ID: 410-67522-4

Date Collected: 12/14/21 10:35

Matrix: Potable Water

Date Received: 12/17/21 15:20

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	EPA 537 Ver 1.1			208332	12/22/21 07:45	RDL8	ELLE
Total/NA	Analysis	EPA 537 Ver 1.1		1	209647	12/28/21 14:46	DCS9	ELLE

Client Sample ID: 7670061 301 Conley Between Lead & Lag

Lab Sample ID: 410-67522-5

Date Collected: 12/14/21 10:30

Matrix: Drinking Water

Date Received: 12/17/21 15:20

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	EPA 537 Ver 1.1			208332	12/22/21 07:45	RDL8	ELLE
Total/NA	Analysis	EPA 537 Ver 1.1		1	209647	12/28/21 14:57	DCS9	ELLE

Client Sample ID: 7670061 301 Conley Between Lead & Lag FB

Lab Sample ID: 410-67522-6

Date Collected: 12/14/21 10:30

Matrix: Potable Water

Date Received: 12/17/21 15:20

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	EPA 537 Ver 1.1			208332	12/22/21 07:45	RDL8	ELLE
Total/NA	Analysis	EPA 537 Ver 1.1		1	209647	12/28/21 15:09	DCS9	ELLE

Lab Chronicle

Client: SUEZ Water Environmental Services Inc
Project/Site: Newberry System

Job ID: 410-67522-1

Client Sample ID: 7670061 301 Conley After Lag Vessel

Lab Sample ID: 410-67522-7

Date Collected: 12/14/21 10:25

Matrix: Drinking Water

Date Received: 12/17/21 15:20

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	EPA 537 Ver 1.1			208332	12/22/21 07:45	RDL8	ELLE
Total/NA	Analysis	EPA 537 Ver 1.1		1	209647	12/28/21 15:20	DCS9	ELLE

Client Sample ID: 7670061 301 Conley After Lag Vessel FB

Lab Sample ID: 410-67522-8

Date Collected: 12/14/21 10:25

Matrix: Potable Water

Date Received: 12/17/21 15:20

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	EPA 537 Ver 1.1			208332	12/22/21 07:45	RDL8	ELLE
Total/NA	Analysis	EPA 537 Ver 1.1		1	209647	12/28/21 15:32	DCS9	ELLE

Client Sample ID: 7670061 101 Conley EP Grab Water

Lab Sample ID: 410-67522-9

Date Collected: 12/14/21 10:20

Matrix: Drinking Water

Date Received: 12/17/21 15:20

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	EPA 537 Ver 1.1			208332	12/22/21 07:45	RDL8	ELLE
Total/NA	Analysis	EPA 537 Ver 1.1		1	209647	12/28/21 15:43	DCS9	ELLE

Client Sample ID: 7670061 101 Conley Field Blank Grab Water

Lab Sample ID: 410-67522-10

Date Collected: 12/14/21 10:20

Matrix: Potable Water

Date Received: 12/17/21 15:20

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	EPA 537 Ver 1.1			208332	12/22/21 07:45	RDL8	ELLE
Total/NA	Analysis	EPA 537 Ver 1.1		1	209647	12/28/21 15:55	DCS9	ELLE

Client Sample ID: 7670061 003 DuPont Well

Lab Sample ID: 410-67522-11

Date Collected: 12/14/21 09:55

Matrix: Drinking Water

Date Received: 12/17/21 15:20

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	EPA 537 Ver 1.1			208332	12/22/21 07:45	RDL8	ELLE
Total/NA	Analysis	EPA 537 Ver 1.1		1	209647	12/28/21 16:18	DCS9	ELLE
Total/NA	Prep	EPA 537 Ver 1.1			208332	12/22/21 07:45	RDL8	ELLE
Total/NA	Analysis	EPA 537 Ver 1.1		10	211118	01/03/22 14:10	VK3G	ELLE

Client Sample ID: 7670061 003 DuPont Well FB

Lab Sample ID: 410-67522-12

Date Collected: 12/14/21 09:55

Matrix: Potable Water

Date Received: 12/17/21 15:20

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	EPA 537 Ver 1.1			208332	12/22/21 07:45	RDL8	ELLE
Total/NA	Analysis	EPA 537 Ver 1.1		1	209647	12/28/21 16:30	DCS9	ELLE

Lab Chronicle

Client: SUEZ Water Environmental Services Inc
Project/Site: Newberry System

Job ID: 410-67522-1

Client Sample ID: 7670061 302 DuPont Between Lead & Lag

Lab Sample ID: 410-67522-13

Date Collected: 12/14/21 09:50

Matrix: Drinking Water

Date Received: 12/17/21 15:20

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	EPA 537 Ver 1.1			208332	12/22/21 07:45	RDL8	ELLE
Total/NA	Analysis	EPA 537 Ver 1.1		1	209647	12/28/21 16:41	DCS9	ELLE

Client Sample ID: 7670061 302 DuPont Between Lead & Lag FB

Lab Sample ID: 410-67522-14

Date Collected: 12/14/21 09:50

Matrix: Potable Water

Date Received: 12/17/21 15:20

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	EPA 537 Ver 1.1			208332	12/22/21 07:45	RDL8	ELLE
Total/NA	Analysis	EPA 537 Ver 1.1		1	209647	12/28/21 16:53	DCS9	ELLE

Client Sample ID: 7670061 302 DuPont After Lag Vessel

Lab Sample ID: 410-67522-15

Date Collected: 12/14/21 09:45

Matrix: Drinking Water

Date Received: 12/17/21 15:20

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	EPA 537 Ver 1.1			208332	12/22/21 07:45	RDL8	ELLE
Total/NA	Analysis	EPA 537 Ver 1.1		1	209647	12/28/21 17:04	DCS9	ELLE

Client Sample ID: 7670061 302 DuPont After Lag Vessel FB

Lab Sample ID: 410-67522-16

Date Collected: 12/14/21 09:45

Matrix: Potable Water

Date Received: 12/17/21 15:20

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	EPA 537 Ver 1.1			208332	12/22/21 07:45	RDL8	ELLE
Total/NA	Analysis	EPA 537 Ver 1.1		1	209647	12/28/21 17:16	DCS9	ELLE

Client Sample ID: 7670061 102 DuPont EP

Lab Sample ID: 410-67522-17

Date Collected: 12/14/21 09:40

Matrix: Drinking Water

Date Received: 12/17/21 15:20

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	EPA 537 Ver 1.1			208332	12/22/21 07:45	RDL8	ELLE
Total/NA	Analysis	EPA 537 Ver 1.1		1	209647	12/28/21 17:28	DCS9	ELLE

Client Sample ID: 7670061 102 DuPont EP FB

Lab Sample ID: 410-67522-18

Date Collected: 12/14/21 09:40

Matrix: Potable Water

Date Received: 12/17/21 15:20

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Prep	EPA 537 Ver 1.1			208332	12/22/21 07:45	RDL8	ELLE
Total/NA	Analysis	EPA 537 Ver 1.1		1	209647	12/28/21 17:39	DCS9	ELLE

Laboratory References:

ELLE = Eurofins Lancaster Laboratories Env, LLC, 2425 New Holland Pike, Lancaster, PA 17601, TEL (717)656-2300

Eurofins Lancaster Laboratories Env, LLC

Accreditation/Certification Summary

Client: SUEZ Water Environmental Services Inc
Project/Site: Newberry System

Job ID: 410-67522-1

Laboratory: Eurofins Lancaster Laboratories Env, LLC

The accreditations/certifications listed below are applicable to this report.

Authority	Program	Identification Number	Expiration Date
Pennsylvania	NELAP	36-00037	01-31-22

1

2

3

4

5

6

7

8

9

10

11

12

13

14

15

Method Summary

Client: SUEZ Water Environmental Services Inc
Project/Site: Newberry System

Job ID: 410-67522-1

Method	Method Description	Protocol	Laboratory
EPA 537 Ver 1.1	EPA 537 Version 1.1	EPA	ELLE
EPA 537 Ver 1.1	EPA 537 Version 1.1	EPA	ELLE

Protocol References:

EPA = US Environmental Protection Agency

Laboratory References:

ELLE = Eurofins Lancaster Laboratories Env, LLC, 2425 New Holland Pike, Lancaster, PA 17601, TEL (717)656-2300



Sample Summary

Client: SUEZ Water Environmental Services Inc
Project/Site: Newberry System

Job ID: 410-67522-1

Lab Sample ID	Client Sample ID	Matrix	Collected	Received
410-67522-1	7670061 001 Playground Well	Drinking Water	12/14/21 10:40	12/17/21 15:20
410-67522-2	7670061 001 Playground Well FB	Potable Water	12/14/21 10:40	12/17/21 15:20
410-67522-3	7670061 005 Conley Well	Drinking Water	12/14/21 10:35	12/17/21 15:20
410-67522-4	7670061 005 Conley Well FB	Potable Water	12/14/21 10:35	12/17/21 15:20
410-67522-5	7670061 301 Conley Between Lead & Lag	Drinking Water	12/14/21 10:30	12/17/21 15:20
410-67522-6	7670061 301 Conley Between Lead & Lag FB	Potable Water	12/14/21 10:30	12/17/21 15:20
410-67522-7	7670061 301 Conley After Lag Vessel	Drinking Water	12/14/21 10:25	12/17/21 15:20
410-67522-8	7670061 301 Conley After Lag Vessel FB	Potable Water	12/14/21 10:25	12/17/21 15:20
410-67522-9	7670061 101 Conley EP Grab Water	Drinking Water	12/14/21 10:20	12/17/21 15:20
410-67522-10	7670061 101 Conley Field Blank Grab Water	Potable Water	12/14/21 10:20	12/17/21 15:20
410-67522-11	7670061 003 DuPont Well	Drinking Water	12/14/21 09:55	12/17/21 15:20
410-67522-12	7670061 003 DuPont Well FB	Potable Water	12/14/21 09:55	12/17/21 15:20
410-67522-13	7670061 302 DuPont Between Lead & Lag	Drinking Water	12/14/21 09:50	12/17/21 15:20
410-67522-14	7670061 302 DuPont Between Lead & Lag FB	Potable Water	12/14/21 09:50	12/17/21 15:20
410-67522-15	7670061 302 DuPont After Lag Vessel	Drinking Water	12/14/21 09:45	12/17/21 15:20
410-67522-16	7670061 302 DuPont After Lag Vessel FB	Potable Water	12/14/21 09:45	12/17/21 15:20
410-67522-17	7670061 102 DuPont EP	Drinking Water	12/14/21 09:40	12/17/21 15:20
410-67522-18	7670061 102 DuPont EP FB	Potable Water	12/14/21 09:40	12/17/21 15:20

1

2

3

4

5

6

7

8

9

10

11

12

13

14

15



410-67522 Chain of Custody



Lancaster Laboratories
Environmental

Acct. # 44297 Group #

Client: SUEZ WATER PA				Matrix				Analyses Requested				For Lab Use Only				
Project Name: Newberry System				Site ID #:				Preservation and Filtration Codes				SF #:				
Project Manager: Elizabeth Zanar				P.O. #:				O				SCR #:				
Sampler: Penny Bumbarger				PWSID #: 7670061				Total # of Containers PFAS (14) 537 v 1.1				Preservation Codes H = HCl T = Thiosulfate N = HNO ₃ B = NaOH S = H ₂ SO ₄ P = H ₃ PO ₄ F = Field Filtered O = Other				
Phone #: 717-773-0185				Quote #: 219948A												
State where samples were collected: PA				For Compliance: Yes <input checked="" type="checkbox"/> No <input type="checkbox"/>												
Sample Identification		Collection		Grab	Composite	Soil	Sediment	Tissue	Potable Water	Ground	Surface	Other: GAC Filtered Water			Remarks	
		Date	Time			<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>				
001 Playground Well		12-14-21	1040	X					X				2	X		Monthly Compliance
FB - Playground Well		12-14-21	1040										2	X		
005 Conley Well		12-14-21	1035	X					X				2	X		
FB - Conley Well		12-14-21	1035										2	X		
301s Conley Between Lead and Lag		12-14-21	1030	X							X		2	X		
FB - Conley Between Lead and Lag		12-14-21	1030										2	X		
301s Conley After Lag		12-14-21	1025	X							X		2	X		
FB - Conley After Lag		12-14-21	1025										2	X		
EP 101 Conley		12-14-21	1020	X							X		2	X		
FB - EP 101 Conley		12-14-21	1020										2	X		
Turnaround Time Requested (TAT) (please check):				Standard <input checked="" type="checkbox"/> Rush <input type="checkbox"/>				Relinquished by:		Date	Time	Received by:		Date	Time	
(Rush TAT is subject to laboratory approval and surcharges.)								Penny Bumbarger		12-17-21	11:50	[Signature]		12-17-21	11:50	
Date results are needed:				Rush results requested by (please check):				Relinquished by:		Date	Time	Received by:		Date	Time	
				E-Mail <input checked="" type="checkbox"/> Phone <input type="checkbox"/>				[Signature]		12-17-21	12:58					
E-mail Address: penny.bumbarger@suez.com								Relinquished by:		Date	Time	Received by:		Date	Time	
Phone: 717-773-0185																
Data Package Options (please check if required)								Relinquished by:		Date	Time	Received by:		Date	Time	
Type I (Validation/non-CLP)	<input type="checkbox"/>	MA MCP	<input type="checkbox"/>													
Type III (Reduced non-CLP)	<input type="checkbox"/>	CT RCP	<input type="checkbox"/>													
Type VI (Raw Data Only)	<input type="checkbox"/>	TX TRRP-13	<input type="checkbox"/>													
NJ DKQP	<input type="checkbox"/>	NYSDEC Category	<input type="checkbox"/>	A or	<input type="checkbox"/>	B										
EDD Required? Yes <input type="checkbox"/> No <input type="checkbox"/>				If yes, format: _____				Relinquished by Commercial Carrier:				Temperature upon receipt		1.3	°C	
								UPS _____ FedEx _____ Other _____								

Arrive at receipt, 12:50

APR 16 12/17/21 Not signed 045 0717 at

Environmental Analysis Request/Chain of Custody



Lancaster Laboratories
Environmental

Acct. # 44297 Group # _____

Sample # _____

Client: SUEZ WATER PA				Matrix			Analyses Requested						For Lab Use Only						
Project Name: Newberry System			Site ID #:	<input type="checkbox"/> Tissue	<input checked="" type="checkbox"/> Ground	<input type="checkbox"/> Surface	Preservation and Filtration Codes						SF #: _____						
Project Manager: Elizabeth Zanar			P.O. #:	<input type="checkbox"/> Sediment	<input type="checkbox"/> Potable	<input type="checkbox"/> NPDES	Total # of Containers	PFAS (14) 537 v 1.1								SCR #: _____			
Sampler: Penny Bumbarger			PWSID #: 7670061	<input type="checkbox"/> Soil	<input type="checkbox"/> Water	<input type="checkbox"/> Other: GAC Filtered Water													
Phone #: 717-773-0185			Quote #: 219948A																
State where samples were collected: PA			For Compliance: Yes <input checked="" type="checkbox"/> No <input type="checkbox"/>																
Sample Identification			Collection		Grab	Composite	Soil	Water	Other: GAC Filtered Water	Total # of Containers	PFAS (14) 537 v 1.1							Remarks	
			Date	Time															
003 DuPont Well			12-14-21	0955	X				2	X								Monthly Compliance	
FB - DuPont Well			12-14-21	0955					2	X									
302s DuPont Between Lead and Lag			12-14-21	0950	X				2	X									
FB - DuPont Between Lead and Lag			12-14-21	0950					2	X									
302s DuPont After Lag			12-14-21	0945	X			X	2	X									
FB - DuPont After Lag			12-14-21	0945					2	X									
EP 102 DuPont			12-14-21	0940	X			X	2	X									
FB - EP 102 DuPont			12-14-21	0940					2	X									
Turnaround Time Requested (TAT) (please check): Standard <input checked="" type="checkbox"/> Rush <input type="checkbox"/>																			
(Rush TAT is subject to laboratory approval and surcharges.)																			
Date results are needed:																			
Rush results requested by (please check): E-Mail <input checked="" type="checkbox"/> Phone <input type="checkbox"/>																			
E-mail Address: penny.bumbarger@suez.com																			
Phone: 717-773-0185																			
Data Package Options (please check if required)																			
Type I (Validation/non-CLP)	<input type="checkbox"/>	MA MCP	<input type="checkbox"/>																
Type III (Reduced non-CLP)	<input type="checkbox"/>	CT RCP	<input type="checkbox"/>																
Type VI (Raw Data Only)	<input type="checkbox"/>	TX TRRP-13	<input type="checkbox"/>																
NJ DKQP	<input type="checkbox"/>	NYSDEC Category	<input type="checkbox"/>	A or	<input type="checkbox"/>	B													
EDD Required? Yes <input type="checkbox"/> No <input type="checkbox"/>				If yes, format: _____															
UPS _____ FedEx _____ Other _____																			
Temperature upon receipt: 1.3 °C																			

APB180 12/17/21 @ Not signed date of receipt 1/4/2022
12:50

Login Sample Receipt Checklist

Client: SUEZ Water Environmental Services Inc

Job Number: 410-67522-1

Login Number: 67522

List Source: Eurofins Lancaster Laboratories Env, LLC

List Number: 1

Creator: Leakway, Christian

Question	Answer	Comment
The cooler's custody seal is intact.	True	
The cooler or samples do not appear to have been compromised or tampered with.	True	
Samples were received on ice.	True	
Cooler Temperature is acceptable ($\leq 6^{\circ}\text{C}$, not frozen).	True	
Cooler Temperature is recorded.	True	
WV: Container Temperature is acceptable ($\leq 6^{\circ}\text{C}$, not frozen).	N/A	
WV: Container Temperature is recorded.	N/A	
COC is present.	True	
COC is filled out in ink and legible.	True	
COC is filled out with all pertinent information.	True	
There are no discrepancies between the containers received and the COC.	True	
Sample containers have legible labels.	True	
Containers are not broken or leaking.	True	
Sample collection date/times are provided.	True	
Appropriate sample containers are used.	True	
Sample bottles are completely filled.	True	
There is sufficient vol. for all requested analyses.	True	
Is the Field Sampler's name present on COC?	True	
Sample custody seals are intact.	N/A	